

We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,900

Open access books available

186,000

International authors and editors

200M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities



WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com



Automated Segmentation of Real-Time 3D Echocardiography Using an Incompressibility Constraint

Yun Zhu, Xenophon Papademetris,
Albert Sinusas and James Duncan
Yale University
USA

1. Introduction

Coronary Heart Disease (CHD) is characterized by reduced blood flow and oxygen supply to the heart muscle, resulting from the occlusion of one or more major coronary arteries. CHD remains the most prevalent cause of mortality in developed countries and represents one of the major burdens on the healthcare systems today. Approximately every 25 seconds, an American will suffer from a coronary event, and about every minute, someone will die from one. According to the 2009 American Heart Association (AHS) report (Lloyd-Jones et al., 2009), an estimated of 16.8 million American adults have CHD (extrapolated to US population in 2009 from National Health and Nutrition Examination Survey (NHANES) 2005-2006) and about 20% of total deaths in the United States are caused by CHD.

The advancements in noninvasive imaging techniques, such as real-time three-dimensional (RT3D) echocardiography, have enabled physicians to detect CHD in its earliest and most treatable stage. With cardiac imaging, physicians are able to evaluate essential global and local functional parameters, such as ejection fraction (EF), wall thickening, strain/strain rate, and etc. The rapid progress in cardiac imaging, however, has led to new challenges in handling of huge amount of image data involved in comprehensive functional patient studies. Manually analyzing these data sets becomes a formidable task for cardiologists, radiologists, and technicians in order to interpret the data and derive clinically useful information for diagnosis or decision support for surgical and pharmacological interventions. Also, manual analysis is subjective and therefore compromises the accuracy and reproducibility of quantitative measurements.

The abovementioned reasons have triggered a great demand for computerized techniques to automate the analysis of cardiac images. Various image-processing tasks need to be performed in order to recover diagnostically useful information, among which myocardial segmentation is one of the most important tasks. Myocardial segmentation aims to delineate the endocardial (ENDO) and epicardial (EPI) boundaries from cardiac images. Accurate segmentation of myocardial boundaries is essential for deriving cardiac global functional parameters such as ventricular mass/volume, ejection fraction, and wall thickening. It is also a prerequisite step for accurate myocardial deformation analysis.

However, myocardial segmentation is challenging, particularly for the EPI surface. This is due to the misleading, physically corrupted, and sometimes incomplete visual information of the EPI contours.

- Myocardium and liver are anatomically close to each other and often share similar intensity values. Thus, there is usually no clear and visible edge between them;
- The myocardium/lung air contrast is lower than the blood pool/myocardium contrast. This implies that myocardium and lung air have similar intensity properties, and might be misclassified as belonging in the same class;
- The right ventricle (RV) myocardium, which has similar intensity values as the left ventricular (LV) myocardium, often merges into the LV myocardium at the LV/RV juncture.

In this chapter, we present an automated coupled deformable model for the segmentation of LV myocardial borders from RT3D echocardiographic images. This approach was originally proposed in (Zhu et al., 2010). It incorporates the incompressibility property of myocardial, and therefore is able to handle fuzzy EPI borders. This model is formulated in a Bayesian framework, which maximizes the regional homogeneities of a cardiac image, while maintaining the myocardial volume during a cycle. By simultaneously evolving both ENDO and EPI surfaces, an automatic segmentation of the full myocardium is achieved from RT3D echocardiographic images.

The remaining sections are organized as follows. Section 2 presents background of echocardiographic image segmentation. Section 3 discusses the incompressibility property of the myocardium. Section 4 is the general framework and details of the method. Section 5 presents experimental results and finally section 6 is the conclusion.

2. Background

The heart is composed of four chambers: left atrium (LA), left ventricle, right atrium (RA), and right ventricle. Fig. 1(a) shows a long-axis cross-section of the heart. As shown in Fig. 1(a), the atria and ventricles are surrounded by muscle tissues called myocardium. Contraction and relaxation of the muscle fibers in the myocardium cause the pumping of the heart.




Fig. 1. Heart anatomy. (a) long-axis cross-section, (b) short-axis cross-section. The grey region represents the myocardium.

Fig. 1(b) shows a short-axis cross-section of the heart. The grey region represents the myocardium. The inner surface of the myocardium is called endocardium, and outer surface is called epicardium. As the LV is a powerful blood pump for the systemic circulation, the delineation of the LV endocardium and epicardium is often the object of interest for cardiologists, particularly because it is an important step for analyzing LV anatomical structure, quantifying cardiac function, and estimating LV motion.

Automatic LV segmentation algorithms can be broadly categorized into region-based and boundary-based methods. Region-based methods exploit the homogeneity of spatially dense information, e.g. pixel-wise grey level values, to produce segmented images. Examples of region-based segmentation methods include thresholding, region-growing, Markov random field-based approaches, and etc. In contrast to the region-based methods, boundary-based methods rely on the pixel-wise difference to guide the process of segmentation. They try to locate points of abrupt changes in the grey tone images. Examples of boundary-based methods include edge-detection, Hough transform, boundary-based snakes, and etc. However, in numerous medical imaging modalities, the LV boundaries cannot be reliably and accurately detected using the algorithms which rely only on boundary or edge information. Reasons for this include significant signal loss, noise, complicated geometric shapes. These problems are ever present for acquired in echocardiography, where the boundary detection problem is further complicated by the presence of attenuation, speckle, shadows, confusing anatomical structures.

In an effort to overcome these difficulties, various sources of image information have been used in the segmentation process. Most commonly used ones include grey level distribution, gradient, texture, and phase information. The choice of which information to use depends on imaging modality, image quality, and specific applications.

2.1 Grey level distribution

Incorporation of imaging physics as prior knowledge has been proven to be useful for cardiac segmentation (Paragios, 2002; Pluempitiwiriya et al., 2005, Chen et al., 2008). The random intensity distribution in ultrasound images, also known as speckle, is caused by the interference of energy from randomly distributed scatters, too small to be resolved by the imaging system. When an acoustic pulse travels through tissue or any medium, backscattering from the scatters in the range cell contributes to the returned echo. This contribution to the echo from the scatters in the range cell has been treated as a random walk because the locations of the scatters are considered to be random. The backscattered echo, therefore, is complex valued with a real part X and imaginary part Y . By virtue of central limit theorem, X and Y have identical Gaussian distribution $N(0, \sigma^2)$, where σ^2 is its variance.

The envelop of the backscattered echo, I , is given by

$$I = \sqrt{X^2 + Y^2}$$

It has been shown that the envelop I has a Rayleigh distribution

$$P(I; \sigma) = \frac{I}{\sigma^2} \exp\left(-\frac{I^2}{2\sigma^2}\right)$$

The Rayleigh distribution has been proven to be one of the most popular models in ultrasound image analysis. For example, a Rayleigh distribution was incorporated into a region-based level-set approach for the segmentation of ultrasound images (Sarti et al., 2005). Also Cohen *et al.* proposed to use Rayleigh distribution as an image formation model in a maximum likelihood motion estimation scheme for noisy ultrasound images (Cohen and Dinstein 2002). Steen *et al.* proposed to use Rayleigh distribution for anisotropic diffusion-based edge detection (Steen and Olstad 1994).

While Rayleigh distribution is extensively used in ultrasound image processing, it is not always not norm because it is valid only in the special case of a large number of randomly distributed scatters. In this condition, the speckle is called fully developed. It has been shown that the echo envelop has a Rayleigh distribution when signal noise ratio (SNR) is approximately 1.91. Deviations from such special scattering conditions are called pre-Rayleigh condition when $\text{SNR} < 1.91$, and post-Rayleigh condition when $\text{SNR} > 1.91$ (Tuthill et al., 1988). The K-distribution has been proposed to model the pre-Rayleigh condition by accounting for low effective scatter density (Shankar et al., 1993; Shankar, 1995), and Rice distribution has been proposed to model the post-Rayleigh condition by accounting for a coherence component due to the presence of a regular structure of scatters within the tissue (Insana et al., 1986; Tuthill et al., 1988). However, the Rician family fails to model the pre-Rayleigh condition, and K-distribution model does not take into account the post-Rayleigh condition. Generalized K-distribution (Jakeman, 1999), homodyned K-distribution (Dutt and Greenleaf, 1994), and Rician Inverse of Gaussian distribution (Eltoft, 2003) have been proposed as general models to encompass pre-Rayleigh, Rayleigh, and post-Rayleigh conditions. Unfortunately, the complex nature of these models limited their practical applications.

Recently, Nakagami distribution has been proposed as a simple generalized model to collectively represent pre-Rayleigh, Rayleigh, and post-Rayleigh distributions (Shankar, 2000). It is a two-parameter distribution with analytical simplicity, which makes it relatively easy to estimate its parameters. The probability density function (pdf) of the Nakagami distribution is given by

$$P(I; \mu, \Omega) = \frac{2\mu^\mu I^{2\mu-1}}{\Gamma(\mu)\Omega^\mu} \exp\left(-\frac{\mu}{\Omega} I^2\right)$$

where $\Gamma(\cdot)$ is the Gamma function, μ is the shape parameter, and Ω is the scaling parameter. For $\mu = 1$ the pdf reduces to Rayleigh condition. For $\mu > 1$, the pdf can be described as post-Rayleigh, which is similar to Rician distribution, while for $\mu < 1$, the pdf can be described as pre-Rayleigh.

As shown in (Shankar, 2000), the shape parameter and scaling parameters can be obtained from the moments of the envelop as follows

$$\mu = \frac{\left[E(I^2)\right]^2}{E\left[I^2 - E(I^2)\right]^2}$$

and

$$\Omega = E(I^2)$$

where $E(\cdot)$ is the expectation.

Davignon *et al.* used a Nakagmi distribution to segment ultrasound images (Davignon *et al.*, 2005).

Apart from the theoretical models mentioned above, empirical models have also been used in ultrasound image segmentation. For example, Tao *et al.* modeled the ultrasound speckle with a Gamma distribution and incorporated it into a tunneling descent optimization framework to overcome local minima (Tao and Tagare, 2007). Xiao *et al.* used a log-normal distribution for modeling speckle in breast images (Xiao *et al.*, 2002). Qian *et al.* incorporated a log-normal distribution into a level-set framework to segment rat echocardiographic images with large dropout (Qian *et al.*, 2006).

2.2 Gradient

Intensity gradient looks for intensity discontinuities between subregions corresponding to different tissue types. Intensity gradient can be computed from an image with standard differential operators (e.g. Sobel operator). A voxel is considered to be a boundary voxel if its intensity gradient is above a threshold. Gradient-based segmentation, such as gradient-based snake (Kass *et al.*, 1987), Active Shape Model (ASM) (Cootes *et al.*, 1995), and level-set approaches (Malladi *et al.*, 1995), has been extensively used in cardiac applications (Lynch *et al.*, 2006; Lynch *et al.*, 2008; Assen *et al.*, 2008; Chen *et al.*, 2002; Corsi *et al.*, 2002).

One limitation of intensity gradient is that it may suffer from spurious, missing, and discontinuous edges. This is especially evident in EPI segmentation because there are usually no clear and visible edges between the myocardium and background. In addition, intensity gradient is suboptimal for ultrasound images because a boundary response in ultrasound images is anisotropic and depends on the relative orientation of the transducer to the boundary. Therefore, intensity gradient information is often used in conjunction with other image features (Pluempitiwiriyaew *et al.*, 2005; Chen *et al.*, 2003).

2.3 Texture

Although no formal definition of texture exists, it can be loosely defined as one or more local patterns that are repeated in a periodic manner. Texture analysis has been attempted in numerous ways in medical image analysis, particularly for ultrasound images. The three principal approaches to describe texture are image pyramids, random fields, and statistical models.

The idea behind image pyramids is to generate a number of homogeneous parameters that represent the response of a bank of filters at different scales and possibly different orientations. Different filters have been proposed including Gabor filters (Xie *et al.*, 2005; Zhan and Shen, 2006), Gaussian derivatives (Chen *et al.*, 1998) and wavelet transforms (Mojsilovic *et al.*, 1998; Yoshida *et al.*, 2003). The idea behind random fields is that the value

at each pixel/voxel is chosen by two-dimensional (2-D)/ three dimensional (3-D) stochastic process. Given a type of pdf of this stochastic process, one can estimate the value at a particular pixel/voxel given the values of other pixels/voxels in its neighborhood. The most commonly used random field is Markov Random Field (MRF) (Bouhlef et al., 2004). Statistical models analyze the spatial distribution of grey level values, by computing local features at each point in the image, and deriving a set of statistics from the distributions of the local features. Statistical approaches yield characterizations of textures as smooth, coarse, grainy, etc. Major statistical texture descriptors include co-occurrence matrix (Nicholas et al., 1986, Sahiner et al., 2004, Kuo et al., 2001), auto-correlation (Chen et al., 2000), edge frequency, run length, Law's texture energy, and fractal texture description. Several researchers have proposed to extract fractal texture features (Wu et al., 1992; Lee et al., 2003; Chen et al., 2005). In addition, several attempts have been made to combine multiple texture measures to improve discrimination abilities (Hao et al., 2001; Christodoulou et al., 2003; Stoitsis et al., 2006).

2.4 Phase

The local phase provides an alternative to intensity gradient to characterize structures in an image (Kovesi, 1996; Boukerroui et al., 2001; Mulet-Parada and Noble, 2000; Hacıhaliloglu et al., 2008). Phase-based methods postulate that feature information is encoded at points where phase congruency is maximized, i.e. all the Fourier components are in phase. Generally, phase is estimated by quadrature filter bank. Thus, there is a link between phase-based methods and other wavelet methods. Phase-based methods are suggested to be more robust than intensity gradient for ultrasound images because intensity gradients in ultrasound images depend on the relative orientation of the transducer to the boundary. In addition, the presence of speckles and imaging artifacts might cause the variation of intensity gradient of equally significant features in the data. One limitation of phase-based methods, however, is that speckle also has its own phase signatures, and therefore an appropriate spatial scale has to be selected.

3. Incompressibility of myocardium

The heart is a remarkably efficient and during mechanical pump composed of complex biological materials. The main structural elements of the myocardium are inter-connected networks of muscle fibers and collagen fibers, as well as matrix that embeds them. The fibers are generally tangential to the ENDO and EPI surfaces, following a path of a right-handed helical geometry. The interstitial fluid carries only hydrostatic pressure, which in turn, is affected by the length and configuration changes of the fibers. These cause pressure gradients, which may result in the flow of the matrix. However, the fluid within the tissue is negligible for the duration of a cardiac cycle because the myocardium has low permeability. Consequently, the myocardium can be assumed to be nearly incompressible (Glass et al., 1990).

A few independent studies have been performed to quantitatively analyze the change of myocardial volume (MV) over an entire cardiac cycle. For example, Hamilton *et al.* performed experiments on frogs, turtles, and dogs, and discovered a relatively consistency of the MV during a cardiac cycle (Hamilton et al., 1932). Hoffman *et al.* used

canine Dynamic Spatial Reconstructor data to study the changes in MV, obtaining a relatively constant volume that was consistent with Hamilton's findings (Hoffman and Ritman, 1987; Hoffman and Ritman, 1985). More recent work done by King *et al.*, who used freehand three dimensional (3-D) echocardiography to measure MV and mass, showed a 1.1% difference of volume between end-diastole(ED) and end-systole (ES) (King *et al.*, 2002). Bowman *et al.* found a variation of around 5% from ED to ES by using high-resolution magnetic resonance (MR) imaging (Bowman and Kovacs, 2003). O'Donnell also analyzed the myocardial volume using MR imaging, and found the difference of ED and ES to be around 2.5% (O'Donnell and Funka-Lea, 1998). The common conclusion of these studies is that the myocardial is nearly incompressible and the variation is less than 5% during a cardiac cycle. This incompressibility property of the myocardium is used as an important cardiac structure constraint which is taken into consideration in our approach, as will be detailed in Section 4.3.

4. Method

4.1 General framework

In this section, we formulate our segmentation problem in a maximum a posterior (MAP) framework combining image-derived information and the incompressibility property of the myocardium. Let I be a 3-D cardiac image, S_{in} be the ENDO surface, and S_{out} be the EPI surface. A MAP framework that combines image information and incompressibility constraint can be expressed as follows

$$(\hat{S}_{in}, \hat{S}_{out}) = \arg \max_{S_{in}, S_{out}} P(S_{in}, S_{out} | I) = \arg \max_{S_{in}, S_{out}} \underbrace{P(I | S_{in}, S_{out})}_{\text{data adherence}} \underbrace{P(S_{in}, S_{out})}_{\text{incompressibility constraint}} \quad (1)$$

Equation (1) is a probability function which adheres to image data, modulated by the prior knowledge of incompressibility property of the myocardium. In particular, $P(I | S_{in}, S_{out})$ is the probability of producing an image I given S_{in} and S_{out} by assuming the piecewise homogeneities of each region enclosed by S_{in} and S_{out} . $P(S_{in}, S_{out})$ is the incompressibility constraint which keeps the MV enclosed by S_{in} and S_{out} nearly constant during a cardiac cycle. Since the myocardium is only nearly incompressible, it is more reasonable to define it within a probabilistic framework rather than imposing a deterministic constraint.

4.2 Data adherence

Region-based deformable models have been successfully applied in the segmentation of images with weak boundaries, due to their robustness (Chan and Vese, 2001). In this work, we evolve a three-phase region-based deformable model based on the statistical intensity distribution from RT3D echocardiographic images.

To evolve a region-based model, we first need to determine the intensity distribution of each region within a cardiac image. It is obvious that the entire cardiac image is partitioned by S_{in} and S_{out} into three regions, namely, LV blood pool, LV myocardium, and background, as shown in Fig. 2.




Fig. 2. The short-axis view of a heart. Two dotted circles are ENDO- and EPI contours, which partitioned the entire image into three regions: LV blood pool, LV myocardium, and background. The background is inhomogeneous because it contains more than one tissues, such as RV blood pool, RV myocardium, lung air, and liver.

The LV blood pool and myocardium are homogeneous, and therefore can modeled with a single pdf. In this work, we use Nakagami distribution (Shankar, 1995) (see Section 2.1) to model the intensity distributions of LV blood pool and myocardium as follows

$$P(I; \mu_l, \omega_l) = \frac{2\mu_l^{\mu_l}}{\Gamma(\mu_l)\omega_l^{\mu_l}} I^{2\mu_l-1} \exp\left(-\frac{\mu_l}{\omega_l} I^2\right) \quad (2)$$

where μ_l is the shape parameter of the Nakagami distribution, and ω_l is the scaling parameter. Equation (2) describes the intensity distribution for LV blood pool when $l=1$, and intensity distribution for LV myocardium when $l=2$.

Unlike LV blood pool and myocardium, the background (see Fig. 2) because it contains more than one tissues (RV blood pool, RV myocardium, and liver), and therefore modeling it with a single distribution function would be insufficient because it contains a wide range of intensities. To handle this problem, we use a mixture model (McLachlan and Peel, 2000) to describe the intensity distribution of the background.

$$P(I; \mu_3, \omega_3) = \sum_{k=1}^M \alpha_k P_{3,k}(I; \mu_{3,k}, \omega_{3,k})$$

where μ_3 and ω_3 are the shape and scaling parameters of the component distributions. For ultrasound images, we choose $M = 2$ because the background histogram has two peaks, as shown in Fig. 3 (c). The first peak corresponds to RV blood pool and lung air, while the second peak corresponds to RV myocardium and liver.




Fig. 3. The histograms of echocardiographic images with their fitted distributions. (a) LV blood pool, (b) LV myocardium, (c) background.

Fig. 3 shows the histogram of each region for echocardiographic images.

Let Ω be a bounded open set of R^3 , and be partitioned by S_{in} and S_{out} into three regions, namely, LV blood pool, LV myocardium, and background, which are denoted as Ω_1 , Ω_2 , and Ω_3 respectively. Thus, the data adherence term can be defined by a three-phase deformable model

$$\log P(I|S_{in},S_{out})=\sum_{l=1}^3\int_{\Omega_l}\log P(I;\mu_l,\omega_l)d\mathbf{x} \tag{3}$$

4.3 Incompressibility constraint

It is mentioned in Section 3 that the MV is nearly constant during cardiac cycle. Therefore, we assume that the MV has a Gaussian distribution $N(V_0,\sigma_V^2)$

$$P(S_{in},S_{out})=\frac{1}{\sqrt{2\pi}\sigma_V}\exp\left(-\frac{(V-V_0)^2}{2\sigma_V^2}\right) \tag{4}$$

where $V=\int_{\Omega_2}d\mathbf{x}$ is the MV enclosed by the deforming ENDO and EPI surfaces. Parameter V_0 is the average volume, which can be calculated from the manual segmentation of the first frame. As mentioned in Section 3, the MV changes by less than 5% during a cardiac cycle.

Assuming that $0.025V_0$ is the maximum deviation and using the three σ rule, we have the following relationship:

$$3\sigma_V = 0.025V_0$$

$$\sigma_V = \frac{1}{120}V_0$$

Therefore, the incompressibility constraint is encoded into a probability function which favors small variance of the MV while penalizing large deviations.

Combining Equations (1), (3), and (4), we arrive at the following optimization problem

$$(\hat{S}_{in}, \hat{S}_{out}) = \arg \max_{S_{in}, S_{out}} P(S_{in}, S_{out} | I) = \arg \max_{S_{in}, S_{out}} \left\{ \sum_{l=1}^3 \int_{\Omega_l} \log P(I; \mu_l, \omega_l) d\mathbf{x} - \frac{(V - V_0)^2}{2\sigma_V^2} \right\} \quad (5)$$

The maximization of Equation (5) can be identified by the coupled surface evolution equations

$$\frac{\partial S_{in}}{\partial \tau} = \left(\log \left(\frac{P(I; \mu_2, \omega_2)}{P(I; \mu_1, \omega_1)} \right) - \frac{V - V_0}{\sigma_V^2} \right) \mathbf{n}_{in} \quad (6)$$

$$\frac{\partial S_{out}}{\partial \tau} = \left(\log \left(\frac{P(I; \mu_3, \omega_3)}{P(I; \mu_2, \omega_2)} \right) + \frac{V - V_0}{\sigma_V^2} \right) \mathbf{n}_{out} \quad (7)$$

where $\mathbf{n}_{in}$ and $\mathbf{n}_{out}$ are the normals of S_{in} and S_{out} respectively, and τ is the propagation time step.

To implement the evolution of Equations (6) and (7), we embed the surface S_{in} and S_{out} in two higher dimensional functions $\varphi_{in}(\mathbf{x})$ and $\varphi_{out}(\mathbf{x})$, which implicitly represents S_{in} and S_{out} as zero level sets, i.e. $S_{in} = \{\mathbf{x} : \varphi_{in}(\mathbf{x}) = 0\}$ and $S_{out} = \{\mathbf{x} : \varphi_{out}(\mathbf{x}) = 0\}$.

To establish the relationship between surface evolution and level set function, we use the relation $\mathbf{n} = \frac{\nabla \varphi}{|\nabla \varphi|}$. Given the curve/surface evolution form

$$\frac{\partial S}{\partial \tau} = F\mathbf{n}$$

the correspondent level set evolution equation reads (Osher and Paragios, 2003)

$$\frac{\partial \varphi}{\partial \tau} = F|\nabla \varphi|$$

where ∇ is the gradient operator. Hence, the level set evolution equation for Equations (8) and (9) are

$$\frac{\partial \phi_{\text{in}}}{\partial \tau} = \left(\log \left(\frac{P(I; \mu_2, \omega_2)}{P(I; \mu_1, \omega_1)} \right) - \frac{V - V_0}{\sigma_V^2} \right) |\nabla \phi_{\text{in}}| \quad (8)$$

$$\frac{\partial \phi_{\text{out}}}{\partial \tau} = \left(\log \left(\frac{P(I; \mu_3, \omega_3)}{P(I; \mu_2, \omega_2)} \right) + \frac{V - V_0}{\sigma_V^2} \right) |\nabla \phi_{\text{out}}| \quad (9)$$

5. Experiments

5.1 Experimental setup

The ultrasound data were acquired using Philips echocardiographic system with a 4 MHz X4 xMatrix transducer. The transducer consists of 3000 miniaturized piezoelectric elements and offers steering in both azimuth and elevation of the beam, permitting real-time volumetric image acquisition and rendering (Philips, 2005). In this work, we acquired 11 sequences of canine images, and each sequence consists of 20-30 frames per cardiac cycle depending on the cardiac rate. Hence, we ran our program with a total of 286 sets of volumetric data.

5.2 Quantitative measures

To quantify the segmentation results, we used two distance error metrics and two area error metrics, namely, mean absolute distance (MAD), Hausdorff distance (HD), percentage of true positives (PTP), and percentage of false positives (PFP).

Let A and B be two surfaces from automatic and manual segmentation, respectively. Suppose they are represented by point sets, i.e. $A = \{\mathbf{a}_1, \mathbf{a}_2, \dots, \mathbf{a}_N\}$ and $B = \{\mathbf{b}_1, \mathbf{b}_2, \dots, \mathbf{b}_M\}$, we define

$$\text{MAD}(A, B) = \frac{1}{2} \left\{ \frac{1}{N} \sum_{i=1}^N d(\mathbf{a}_i, B) + \frac{1}{M} \sum_{j=1}^M d(\mathbf{b}_j, A) \right\}$$

$$\text{HD}(A, B) = \max \left\{ \max_i d(\mathbf{a}_i, B), \max_j d(\mathbf{b}_j, A) \right\}$$

where $d(\mathbf{a}_i, B) = \min_j \|\mathbf{a}_i - \mathbf{b}_j\|$ is the distance from point $\mathbf{a}_i$ to the closest point on surface B . While MAD represents the global agreement between two contours, HD compares their local similarities.

Let Ω_A and Ω_B be the region enclosed by surface A and B , we define

$$\text{PTP} = \frac{\text{Volume}(\Omega_A \cap \Omega_B)}{\text{Volume}(\Omega_A)}$$

$$\text{PFP} = \frac{\text{Volume}(\Omega_A) - \text{Volume}(\Omega_A \cap \Omega_B)}{\text{Volume}(\Omega_B)}$$

5.3 Experimental results

In Fig. 4, we show the long axis view of the automatically segmented ENDO- and EPI surfaces at frames 2, 4, 6, and 8 during ventricular systole.




Fig. 4. Long axis view of segmented ENDO- and EPI contours at frames 2, 4, 6, and 8 during cardiac systole.

In Fig. 5, we compare the segmentation results of the EPI surface with and without incompressibility constraint. For a fair comparison, we used the same data adherence term for both cases. We observed that while the ENDO border was correctly detected even without the incompressibility constraint, the EPI contour leaked into other tissues, such as liver, that look like myocardium. This is because the LV myocardium/background contrast is lower than the LV blood pool/myocardium contrast, especially for the lateral and inferior sectors of myocardium. This low contrast obscures the exact location of the EPI boundary, making EPI segmentation more challenging than ENDO segmentation. When the incompressibility constraint was applied, however, the coupling of ENDO and EPI contours forced the evolution of the EPI contour to be consistent with that of the ENDO contour, thus preventing the leakage of the EPI contour.




Fig. 5. Comparison of EPI segmentation with and without incompressibility constraint. (a) with incompressibility constraint; (b) without incompressibility constraint. Arrow: myocardium juncture.

As explained in Section 1, another challenge in the segmentation of the EPI boundary is the presence of the LV/RV myocardium junctures. The intensity similarity between the LV and RV myocardium makes the EPI boundary ambiguous at the juncture. When the incompressibility constraint was not applied, the EPI contour leaked out to segment RV myocardium. When the incompressibility constraint was applied, however, the LV myocardium was successfully separated from the RV myocardium at the juncture.

Fig. 6 compares the MV for an entire sequence from manual segmentation, automatic segmentation with incompressibility constraint, and automatic segmentation without incompressibility constraint. We took the mean value from three experts as the MV from manual segmentation. We observed that the variation of MV for manual segmentation is

$\frac{92.1 - 88.2}{88.2} = 4.42\%$, the variation of MV for automatic segmentation with incompressibility constraint is $\frac{92.2 - 88.1}{88.1} = 4.65\%$. The variation of MV obtained without incompressibility

constraint, however, is $\frac{143.2 - 106.1}{106.1} = 34.97\%$. Also, we noticed that the MV obtained without incompressibility constraint is much larger than that from manual segmentation. This is because the EPI contour leaked into the background, leading to the over-estimation of the MV.




Fig. 6. Comparison of MV for an example sequence from manual segmentation, automatic segmentation with incompressibility constraint, and automatic segmentation without incompressibility constraint.

In addition, we performed Bland-Altman analysis (Bland and Altman, 1986) to assess the agreement of the MV measurements from manual segmentation and automatic segmentation. The Bland-Altman analysis revealed small bias and good coherence between the MV measurements from manual segmentation and automatic segmentation with incompressibility constraint. As shown in Fig. 7, the bias is -0.2%, and 95% of the computer measurements of MV can be expected to differ from expert measurements by less than 6.4% below and 5.9% above the mean. In comparison, when the incompressibility constraint was not applied, the MV largely deviated from the MV from manual segmentation (bias = 18.8%, 95% confidence interval = [2.4%,35.2%]).




Fig. 7. Bland-Altman analysis showing the agreement between the MV measurements from manual segmentation and automatic segmentation.

Table 1 shows that the ENDO boundaries were detected with sufficient accuracy even if the incompressibility constraint was not applied. This is because that the ENDO boundaries are relatively clear compared to the EPI boundaries. However, we observed from Table 2 that for EPI boundaries, when the incompressibility constraint was not applied, the automatic-manual MAD was 2.78mm larger than the manual-manual MAD, the automatic-manual HD was 3.82mm larger than the manual-manual HD, and the automatic-manual PTP&PFP were 18%-20% worse than the manual-manual PTP&PFP. This is because of the leakage problem of EPI boundaries without incompressibility constraint. When the incompressibility constraint was applied, however, the automatic-manual MAD decreased by 3.38mm, the automatic-manual HD decreased by 4.11mm, and the automatic-manual PTP&PFP decreased by 16%-19%. We can see from Tables 1 and 2 that the automatic algorithm produced results with comparable accuracy to a manual segmentation. Furthermore, we observed from Tables 1 and 2 that the variability of manual-manual segmentation was smaller for the ENDO surface than for the EPI surface. This is probably because the EPI boundaries are more ambiguous for observers to detect, which was also the reason why we have multiple observers instead of a single one.

	MAD (mm)	HD (mm)	PTP (%)	PFP (%)
Automatic-manual (with constraint)	1.43 ± 0.29	2.40 ± 0.48	96.22 ± 1.97	3.99 ± 1.99
Automatic-manual (without constraint)	1.50 ± 0.29	2.42 ± 0.60	95.70 ± 1.76	4.29 ± 1.80
Manual-manual	1.31 ± 0.27	2.12 ± 0.50	96.21 ± 1.90	3.79 ± 1.84

Table 1. Comparison of automatic outline to three observers’ outline of ENDO boundaries.

	MAD (mm)	HD (mm)	PTP (%)	PFP (%)
Automatic-manual (with constraint)	1.74 ± 0.49	2.91 ± 0.80	94.34 ± 1.21	5.66 ± 1.20
Automatic-manual (without constraint)	5.12 ± 2.12	7.02 ± 2.42	75.23 ± 7.17	22.45 ± 7.99
Manual-manual	2.34 ± 0.89	3.20 ± 0.99	95.90 ± 2.31	4.10 ± 2.30

Table 2. Comparison of automatic outline to three observers’ outline of EPI boundaries.

6. Conclusion

In this chapter, we have presented a novel approach to segmenting the full myocardial volume from cardiac MR and ultrasound images. Motivated by the incompressibility property of myocardium during a cardiac cycle, we coupled the propagation of the ENDO and EPI surfaces according to image-derived information that maximized the piecewise homogeneities of a cardiac image, as well as the incompressibility constraint which constrains the variation of myocardial volume within 5%.

To validate our approach, we computed the MAD, HD, PTP, and PFP between the contours from automatic algorithm and manual segmentation. We observed that when the incompressibility constraint was not applied, the EPI boundaries leaked into other tissues. When incompressibility constraint was applied, however, the MAD, HD, PTP, and PFP were significantly improved.

7. References

Assen, H; Danilouchkine, M.; Dirksen, M.; Reiber, J. & Lelieveldt, B. (2008). A 3-D Active Shape Model Driven by Fuzzy Inference: Application to Cardiac CT and MR. *IEEE Trans. on Information Technology in Biomedicine*, vol.12, no.5 (2008), pp.595-605

Bland, J. & Altman, D. (1986). Statistical Methods for Assessing Agreement Between Two Methods of Clinical Measurement. *Lancet*, vol.1 (1986)

Bouhlef, N.; Sevestre-Ghalila, S. & Rajhi, H. & Hamza, R. (2004). New Markov Random Field Model Based on K-Distribution for Textured Ultrasound Image. In *Proceedings of Medical Imaging: Ultrasonic Imaging and Signal Processing*, ser. SPIE, W.F. Walker and S.Y. Emelianov, pp.363-372

Boukerroui, D.; Noble, A.; Robini, M. & Brady, M. (2001). Enhancement of Contrast Region in Suboptimal Ultrasound Images with Application to Echocardiography. *Ultrasound in Medicine & Biology*, vol.27, no.12 (2001), pp.1583-1594

Bowman, A. & Kovacs, S. (2003). Assessment and Consequences of the Constant-Volume Attribute of the Four-Chamber Heart. *American Journal of Physiology - Heart Circulation Physiology*, vol.285 (2003), pp.H2027-2033

Chan, T. & Vese, L. (2001). Active Contours Without Edges. *IEEE Transactions on Image Processing*, vol.10, no.2 (2001), pp.266-277


Chen, C.; Pau, L. & Wang, P. (1998). Handbook of Pattern Recognition and Computer Vision (2nd edition). World Scientific

Chen, D.; Chang, R. & Huang, Y. (2000). Breast cancer diagnosis using self-organizing map for sonography. *Ultrasound in Medicine and Biology*, vol.26, no.3 (2000), pp.405-411

- Chen, D.; Chang, R.; Chen, C.; Ho, M.; Kuo, S.; Chen, S.; Hung, S. & Moon, W. (2005). Classification of breast ultrasound images using fractal feature. *Clinical Imaging*, vol.29 (2005), pp.235-245
- Chen, T.; Babb, J.; Kellman, P.; Axel, L. & Kim, D. (2008). Semiautomated Segmentation of Myocardial Contours for Fast Strain Analysis in Cine Displacement-Encoded MRI. *IEEE Transactions on Medical Imaging*, vol.27, no.8 (2008), pp.1084-1094
- Chen, Y.; Tagare, H.; Thiruvenkadam, S.; Huang, F.; Wilson, D.; Gopinath, K.; Briggs, R & Geiser, E (2002). Using Prior Shapes in Geometric Active Contours in a Variational Framework. *International Journal of Computer Vision*, vol.50, no.3 (2002), pp.315-328
- Christodoulou, C.; Pattichis, C.; Pantziaris, M. & Nicolaides, A. (2003). Texture-Based Classification of Atherosclerotic Carotid Plaques. *IEEE Transactions on Medical Imaging* vol.22, no.7 (2003), pp.902-912
- Chen, Y.; Huang, F.; Tagare, H.; Rao, M.; Wilson, D. & Geiser, E. (2003). Using Prior Shape and Intensity Prior Profile in Medical Image Segmentation. In *Proceedings of International Conference on Computer Vision and Pattern Recognition*, vol.2 (2003), pp.1117-1125
- Cohen, B. & Dinstein, I. (2002). New Maximum Likelihood Motion Estimation Schemes for Noisy Ultrasound Images. *Pattern Recognition*, vol.48, no.2 (2002), pp.455-463
- Cootes, T. & Taylor, C.; Cooper, D. & Graham, J. (1995). Active Shape Models - Their Training and Application. *Computer Vision and Image Understanding*, vol.61, no.1 (1995), pp.38-59
- Corsi, C; Saracino, G.; Sarti, A. & Lamberti, C. (2002). Left Ventricular Volume Estimation for Real-Time Three-Dimensional Echocardiography. *IEEE Transactions on Medical Imaging*, vol.21, no.9 (2002), pp.1202-1208
- Davignon, F.; Deprez, J. & Basset, O. (2005). A Parametric Imaging Approach for the Segmentation of Ultrasound Data. *Ultrasonics*, vol.43 (2005), pp.789-801
- Dutt, V. & Greenleaf, J. (1994). Ultrasound Echo Envelop Analysis Using A Homodyned K-distribution Signal Model. *Ultrasonic Imaging*, vol.16, no.4 (1994), pp.265-287
- Eltoft, T. (2003). Speckle: Modeling and Filtering. In *Norwegian Signal Processing Symposium* (2003)
- Glass, L.; Hunter, P. & McCulloch, A. (1990). Theory of Heart: Biomechanics, Biophysics, and Nonlinear Dynamics of Cardiac Function. Springer-Verlag
- Hacihaliloglu, I.; Abugharbieh, R.; Hodgson, A. & Rohling, R. (2008). Bone Segmentation and Fracture Detection in Ultrasound Using 3D Local Phase Features. In *Proceedings of International Conference on Medical Imaging and Computer Assisted Intervention* (2008), pp.287-295
- Hamilton, W. & Rompf, J. (1932). Movement of the Base of the Ventricle and the Relative Constancy of the Cardiac Volume. *American Journal of Physiology*, vol.132 (1932), pp.559-565
- Hao, X.; Bruce, C.; Pislaru, C. & Greenleaf, J. (2001). Segmenting High-Frequency Intracardiac Ultrasound Images of Myocardium Into Infarcted, Ischemic, and Normal Regions. *IEEE Transactions on Medical Imaging*, vol.20, no.12 (2001), pp.1373-1383
- Hoffman, E. & Ritman, E. (1985). Invariant Total Heart Volume in the Intact Thorax. *American Journal of Physiology - Heart Circulation Physiology*, vol.249 (1985), pp.H883-890
- Hoffman, E. & Ritman, E. (1987). Heart-Lung Interaction: Effect of Regional Lung Air Content and Total Heart Volume. *Annual Biomedical Engineering*, vol.15 (1987), pp.241-257
- Insana, M.; Wagner, R.; Garra, B.; Brown D. & Shawker, T. (1986). Analysis of Ultrasound Image Texture via Generalized Rician Statistics. *Optical Engineering*, vol.25, no.6 (1986), pp.743-748

- Jakeman, E. (1999). K-distributed Noise. *Journal of Optics A: Pure and Applied Optics*, vol.1 (1999), pp.784-789
- Kass, M.; Witkin, A. & Terzopoulos, D. (1987). Snakes: Active Contour Models. *International Journal of Computer Vision*, vol.1 (1987), pp.321-331
- King, D.; Coffin, L. & Maurer, M. (2002). Noncompressibility of Myocardium During Systole With Freehand Three-Dimensional Echocardiography. *Journal of the American Society of Echocardiography*, vol.15, no.12 (2002), pp.1503-1506
- Kovesi, P. (1996). Invariant Measures of Image Features from Phase Information. PhD Thesis, University of Western Australia
- Kuo, W.; Chang, R.; Chen, D. & Lee, C. (2001). Data mining with decision trees for diagnosis of breast tumor in medical ultrasonic images. *Breast Cancer Research and Treatment*, vol.66, no.1 (2001), pp.51-57
- Lee, W.; Chen, Y. & Hsieh, K. (2003). Ultrasonic liver tissues classification by fractal feature vector based on M-band wavelet transform. *IEEE Transactions on Medical Imaging*, vol.22, no.3 (2003), pp.382-392
- Lloyd-Jones, D.; et al. (2009). Heart Disease and Stroke Statistics 2009 Update: A Report From the American Heart Association Statistics Committee and Stroke Statistics Subcommittee. *Circulation*, vol.119, (2009), pp.e21-e181
- Lynch, M.; Ghita, O. & Whelan, P. (2006). Left-Ventricle Myocardium Segmentation Using a Coupled Level-Set With a Prior Knowledge. *Computerized Medical Imaging and Graphics*, vol.30, no.4 (2006), pp.255-262
- Lynch, M.; Ghita, H & Whelan, P (2008). Segmentation of the Left Ventricle of the Heart in 3-D+t MRI Data Using an Optimized Nonrigid Temporal Model. *IEEE Transactions on Medical Imaging*, vol.27, no.2 (2008), pp.195-203
- Malladi, R.; Sethian, J. & Vemuri, B. (1995). Shape Modeling with Front Propagation: A Level Set Approach. *IEEE Transactions on Pattern Analysis and Machine Intelligence*, vol.17, no.2 (1995), pp.158-175
- Mojsilovic, A.; Popovic, M.; Markovic, S. & Krstic, M. (1998). Characterization of visually similar diffuse diseases from B-scan liver images using nonseparable wavelet transform. *IEEE Transactions on Medical Imaging*, vol.17, no.4 (1998), pp.541-549
- Mulet-Parada, M. & Noble, A. (2000). 2D+T Acoustic Boundary Detection in Echocardiography. *Medical Image Analysis*, vol.4 (2000), pp.21-30
- Nicholas, D.; Nassiri, D.; Garbutt, P.; Hill, C. (1986). Tissue characterization from ultrasound B-scan data. *Ultrasound in Medicine & Biology*, vol.12, no.2 (1986), pp.135-145
- O'Donnell, T. & Funka-Lea, G. (1998). 3-D Cardiac Volume Analysis Using Magnetic Resonance Imaging. In *Proceedings of the Fourth IEEE Workshop on Applications of Computer Vision (WACV)* (1998), pp.240-241
- Osher, S. & Paragios, N. (2003). Geometric Level Set Methods in Imaging, Vision, and Graphics. Springer
- Paragios, N (2002). A Variational Approach for the Segmentation of the Left Ventricle in Cardiac Image Analysis. *International Journal of Computer Vision*, vol.50, no.3 (2002), pp.345-362
- Philips (2005). <http://www.medical.philips.com/main/products/ultrasound/cardiology/>. Philips Medical Systems Corporate Website
- Pluempitiwiriyaew, C.; Moura, J.; Wu, Y. & Ho, C. (2005). STACS: New Active Contour Scheme for Cardiac MR Image Segmentation. *IEEE Transactions on Medical Imaging*, vol.24, no.5 (2005), pp.593-603

- Qian, X.; Tagare H. & Tao, Z. (2006). Segmentation of Rat Cardiac Ultrasound Images with Large Dropout Regions. In *Proceedings of Mathematical Methods in Biomedical Image Analysis* (2006)
- Sahiner, B.; Chan, H.; Roubidoux, M.; Helvie, M.; Hadjiiski, L.; Ramachandran, A.; Paramagul, C.; LeCarpentier, G.; Nees, A. & Blane, C. (2004). Computerized characterization of breast masses on three dimensional ultrasound volumes. *Medical Physics*, vol.31, no.4 (2004), pp.744-754
- Sarti, A.; Corsi, C.; Mazzini, E. & Lamberti C. (2005). Maximum Likelihood Segmentation of Ultrasound Images with Rayleigh Distribution. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, vol.52, no.6 (2005), pp.947-960
- Shankar, P.; Reid, J.; Ortega H.; Piccoli C. & Goldberg, B. (1993). Use of Non-Rayleigh Statistics for the Identification of Tumors in Ultrasonic B-Scans of the Breast. *IEEE Transactions on Medical Imaging*, vol.12, no.4 (1993), pp.687-692
- Shankar, P. (1995). A Model for Ultrasonic Scattering from Tissues Based on the K Distribution. *Physics in Medicine & Biology*, vol.40 (1995), pp.1633-1649
- Shankar, P. (2000). A General Statistical Model for Ultrasonic Backscattering from Tissues. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, vol.47, no.3 (2000), pp.727-736
- Steen, E. & Olstad, B. (1994). Scale-Space and Boundary Detection in Ultrasound Imaging, Using Signal-Adaptive Anisotropic Diffusion. In *Proceedings of SPIE Medical Imaging: Image Processing*
- Stoitsis, J.; Golemati, S. & Nikita, K. (2006). A Modular Software System to Assist Interpretation of Medical Images – Application to Vascular Ultrasound Images. *IEEE Transactions on Instrumentation and Measurement*, vol.55, no.6 (2006), pp.1944-1952
- Tao, Z. & Tagare, H. (2007). Tunneling Descent for M.A.P. Active Contours in Ultrasound Segmentation. *Medical Image Analysis*, vol.11, no.3 (2007), pp.266-281
- Tuthill, T.; Sperry, R. & Parker, K. (1988). Deviations From Rayleigh Statistics in Ultrasonic Speckle. *Ultrasonic Imaging*, vol.10 (1988), pp.81-89
- Wu, C.; Chen, Y. & Hsieh, K. (1992). Texture features for classification of ultrasonic liver images. *IEEE Transactions on Medical Imaging*, vol.11, no.2 (1992), pp.141-152
- Xiao, G.; Brady, M.; Noble J. & Zhang, Y. (2002). Segmentation of Ultrasound B-mode Images with Intensity inhomogeneity Correction. *IEEE Transactions on Medical Imaging*, vol.21, no.1 (2002), pp.48-57
- Xie, J.; Jiang, Y & Tsui, H. (2005). Segmentation of Kidney from Ultrasound Images Based on Texture and Shape Priors. *IEEE Transactions on Medical Imaging*, vol.24, no.1 (2005), pp.45-57
- Yoshida, H.; Casalino, D.; Keserci, B.; Coskun, A.; Ozturk, O. & Savranlar, A. (2003). Wavelet-packet-based texture analysis for differentiation between benign and malignant liver tumours in ultrasound images. *Physics in Medicine and Biology*, vol.48 (2003), pp.3735-3753
- Zhan, Y. & Shen, D. (2006). Deformable Segmentation of 3-D Ultrasound Prostate Images Using Statistical Texture Matching Method. *IEEE Transactions on Medical Imaging*, vol.25, no.3 (2006), pp.256-27
- Zhu, Y.; Papademetris, X.; Sinusas, A. & Duncan, J. (2010). A Coupled Deformable Model for Tracking Myocardial Borders from Real-time Echocardiography Using an Incompressibility Constraint. *Medical Image Analysis*, vol.14, no.3 (2010), pp.428-449



Echocardiography - New Techniques

Edited by Prof. Gani Bajraktari

ISBN 978-953-307-762-8

Hard cover, 218 pages

Publisher InTech

Published online 18, January, 2012

Published in print edition January, 2012

The book "Echocardiography - New Techniques" brings worldwide contributions from highly acclaimed clinical and imaging science investigators, and representatives from academic medical centers. Each chapter is designed and written to be accessible to those with a basic knowledge of echocardiography. Additionally, the chapters are meant to be stimulating and educational to the experts and investigators in the field of echocardiography. This book is aimed primarily at cardiology fellows on their basic echocardiography rotation, fellows in general internal medicine, radiology and emergency medicine, and experts in the arena of echocardiography. Over the last few decades, the rate of technological advancements has developed dramatically, resulting in new techniques and improved echocardiographic imaging. The authors of this book focused on presenting the most advanced techniques useful in today's research and in daily clinical practice. These advanced techniques are utilized in the detection of different cardiac pathologies in patients, in contributing to their clinical decision, as well as follow-up and outcome predictions. In addition to the advanced techniques covered, this book expounds upon several special pathologies with respect to the functions of echocardiography.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Yun Zhu, Xenophon Papademetris, Albert Sinusas and James Duncan (2012). Automated Segmentation of Real-Time 3D Echocardiography Using an Incompressibility Constraint, Echocardiography - New Techniques, Prof. Gani Bajraktari (Ed.), ISBN: 978-953-307-762-8, InTech, Available from:
<http://www.intechopen.com/books/echocardiography-new-techniques/automated-segmentation-of-real-time-3d-echocardiography-using-an-incompressibility-constraint>

INTECH
open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia
Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元
Phone: +86-21-62489820
Fax: +86-21-62489821

© 2012 The Author(s). Licensee IntechOpen. This is an open access article distributed under the terms of the [Creative Commons Attribution 3.0 License](https://creativecommons.org/licenses/by/3.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

IntechOpen

IntechOpen