We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,300

Open access books available

172,000

International authors and editors

190M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities

WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com

Critical Aspects for Mechanical Simulation in Dental Implantology

Erika O. Almeida^{1,2} et al.*

¹*Department of Biomaterials and Biomimetics,
New York University College of Dentistry, New York, NY,*

²*Department of Dental Materials and Prosthodontics,
Sao Paulo State University College of Dentistry, Araçatuba, SP,*

¹USA

²Brazil

1. Introduction

Dental implants have been widely used for the rehabilitation of completely and partially edentulous patients. (Branemark et al. 1969; Branemark et al. 1977; Adell et al. 1990; van Steenberghe et al. 1990) Despite the high success rates reported by a vast number of clinical studies, early or late implant failures are still unavoidable. (Esposito et al. 1998) Mechanical complications and failures have frequently been reported during prosthetic treatment. (Roberts 1970; Randow et al. 1986; Walton et al. 1986; Levine et al. 1999)

From a biomechanical point of view, forces occurring either from functional or parafunctional occlusal contact may result in a physiologic adaptation of the supporting tissues since implants are rigidly anchored to the bone. However, if the stress generated is beyond the adaptive capacity of the host, the response of the supporting tissues and prosthetic components may result in failures. Therefore, the load magnitude and duration employed to implant-restoration systems play a significant role in biomechanical stress dissipation on the implant-prosthesis system and surrounding tissues. (Menicucci et al. 2002) Other factors that are known to affect the stress/strain distribution on bone surrounding implants are the implant position and angulations, implant-abutment connection and the magnitude and direction of the occlusal load. (Stanford and Brand 1999; Kozlovsky et al. 2007; Lin et al. 2009)

*Amilcar C. Freitas Júnior³, Eduardo P. Rocha², Roberto S. Pessoa⁴, Nikhil Gupta⁵, Nick Tovar¹ and Paulo G. Coelho¹

¹*Department of Biomaterials and Biomimetics, New York University College of Dentistry, New York, USA*

²*Department of Dental Materials and Prosthodontics, São Paulo State University College of Dentistry, Araçatuba, SP, Brazil*

³*Postgraduate Program in Dentistry, Potiguar University, College of Dentistry – UnP, Natal, RN, Brazil*

⁴*Department of Mechanical Engineering - Federal University at Uberlândia, FEMEC - Uberlândia, Uberlândia, MG, Brazil*

⁵*Department of Mechanical and Aerospace Engineering, Polytechnic Institute of New York University, New York, USA*

While randomized controlled clinical studies are strongly suggested (Bozini et al. 2011; Pieri et al. 2011; Turkyilmaz 2011) as the optimal approach to evaluate the performance of biomaterials and biomechanical aspects of dental implants and prosthetic components, often times these studies are not economically viable (especially in cases where multiple variables are to be included). Thus, well designed *in vitro* studies utilizing virtual models via finite element analysis (FEA) should be considered, as this method allows researchers to address a range of questions that are otherwise intractable due to the number of variations within clinical trials or the difficulty in solving analytically (Ross 2005).

Although there are several implant dentistry studies using the FEA method, (Rayfield 2004; Yokoyama et al. 2005; Galantucci et al. 2006; Huang et al. 2007; Assuncao et al. 2008; Lim et al. 2008; de Almeida et al. 2011) each step of the method deserves discussion in order to facilitate its understanding when applied to implant/restoration/bone system. Thus, the aim of this chapter is to conduct a critical review of simulated biomechanical scenarios describing essential basic aspects and approaches currently used in implantology.

2. Model creation

The first step in creating a finite element (FE) model is to determine the appropriate number of dimension to be utilized for evaluation (1, 2, or 3D). While three dimensions are more realistic for the complex anatomy/implant/restoration biomechanical interaction, it is proportionally more challenging when CAD modeling, solving, and software output interpretation is considered. (Richmond et al. 2005) For selected situations, 2D analyses are often adequate for the questions at hand. (Rayfield 2004; Assuncao et al. 2008; Freitas et al. 2010; Freitas Junior et al. 2010)

The choice between 2D and 3D FEA for investigating the biomechanical behavior of complex structures depends on several factors, including the complexity of the geometry, the type of analyses required, expectations in terms of accuracy as well as the general applicability of the results. (Romeed et al. 2006)

Two-dimensional FEA has previously been used in different areas of dental research. (Burak Ozcelik et al. ; Freitas, Rocha et al. 2010; Freitas Junior, Rocha et al. 2010) However, its main limitation is known to be the lower accuracy and reliability relative to 3D and its utilization has drastically fallen from favor when dental treatment biomechanical simulations are considered. (Yang et al. 1999; Yang et al. 2001) In contrast, 3D FEA has acceptable accuracy/reliability while properly capturing the geometry of complex structures. However, the higher the complexity of 3D FE models the higher the difficulty in generating appropriate mesh refinement for simulation, which is more easily achievable for 2D models. (Romeed, Fok et al. 2006)

It is general consensus that 3D reconstruction is essential to study the interaction between anatomy and mechanical behavior of restorative components and dental implants. (Galantucci, Percoco et al. 2006) For that purpose, models may be manually constructed or may be generated from imaging methods such as computed tomography (Figures 1 A and B). The choice for building a model using either a manual or automatic technique depends on the purpose of the study and the structure of interest. Considering the specific morphology of biological structures, the non-uniform rational basis spline (NURBS) has

been commonly used for 3D modeling as it allows the achievement of reliable analytic or freeform parts based on an efficient management of curves and surfaces.

Fig. 1. Models of maxilla (A) and mandible (B) implant supported prosthesis based on (A) model reconstruction from imaging techniques and (B) manual construction on CAD software (Solid Works Corp).

The manual input technique generate structures in appropriate aided design software such as AutoCAD (Autodesk Inc, San Rafael, CA, USA), SolidWorks (SolidWorks Corp., Concord, MA, USA), Pro/Engineer (Wildfire, PTC, Needham, MA, USA), Rhino 3D (McNeel North America, Seattle, WA, USA) (Figure 1B).

Fig. 2. CT scan data as seen in Mimics 13.0 (Materialise, Leuven, Belgium). (A-C) The maxilla is presented in three different cross-sectional views. Masks have been applied according to voxel density thresholding to determine the regions of interest. (D) 3D representation of maxilla as a result of segmentation in Mimics.

The imaging approach involves transforming available medical imaging files from computed tomography (CT) scans, magnetic resonance images (MRI), ultrasound, and laser digitizers into wireframe models that are then converted into FE models. (Romeed, Fok et al. 2006) While laser scans offer a high-resolution representation of the outer surface, these lack information about internal geometry. (Kappelman 1998) Creating models from CT or MRI is often time-consuming but can provide accurate models with fine structural details based on image density thresholding. (Cohen et al. 1999; Ryan and van Rietbergen 2005) The 3D object is automatically created in the form of masks by thresholding the region of interest on the entire stack of scans (Figure 2A-D). The degree of automation and high resolution make this model creation method attractive, but determining the appropriate thresholding algorithms to the bone-air boundary reliably throughout a structure with varying bone thicknesses and density can be challenging. (Fajardo et al. 2002)

When advanced imaging data is used to generate solid models, surface smoothing is advised in order to decrease the number of nodes and elements in the discretized FE model as such an approach generally decreases computation time. (Wang et al. 2005; Magne 2007) Fig. 3A shows an excessive number of elements in a dental implant that was subsequently reduced (Fig. 3B) by computer software (Materialise, Leuven, Belgium). However, it is also advisable that when surface smoothing is performed it does not over simplify the geometries, causing a decrease in solution accuracy.

Fig. 3. (A) 3D CAD of the Nobel Speed TM RP Implant (Nobel Biocare, CA, USA) based on the micro-CT (μ CT; CT40, Scanco Medical, Bassersdorf, Switzerland). (B) The Mimics Remesh (Materialise) function “quality preserving reduce triangle” and “reduce triangles” were used to reduce the number of elements at the implant CAD.

3. Material properties and software input limitation

Material properties such as heat conductivity, linear and nonlinear elastic properties, and temperature-dependent elastic properties may be utilized in FEA. (Richmond, Wright et al. 2005) In dentistry, the majority of previous FE scientific communications in implant prosthodontics has considered material properties to be isotropic, homogeneous, and linear elastic. (Huang et al. 2008; Canay and Akca 2009; Eser et al. 2009; Li et al. 2009; Chang et al. 2010; Chou et al. 2010; Okumura et al. 2010; Sagat et al. 2010; Wu et al. 2010; Burak Ozcelik et al. 2011) An isotropic material indicates that the mechanical response is similar regardless of the stress field direction, requiring Young’s modulus (E) and Poisson’s ratio (v) values for the FE calculation. (Richmond, Wright et al. 2005)

The elastic, or Young’s modulus (E), is defined as stress/strain (σ/ϵ) and is measured in simple extension or compression. It is a measure of material deformation under a given axial load. In other words, a numerical description of its stiffness. Poisson’s ratio (v) is the lateral strain divided by axial strain, thus representing how much the sides of a material as it is tensile tested. (Richmond, Wright et al. 2005)

Since bone is one of the structures to be simulated in FEA, (Peterson and Dechow 2003; Bozkaya et al. 2004; Danza et al. 2010; de Almeida et al. 2010) it is often treated as an anisotropic structure and three elastic modulus (E), three Poisson’s ratio (v) and three shear modulus (G) are required (Table 1). (Chen et al. 2010; Sotto-Maior et al. 2010) Anisotropic materials are characterized by different stress responses under forces applied in varied directions within the structure. (O’Mahony et al. 2001; Natali et al. 2009; Eraslan and Inan 2010) The elastic behavior in cortical bone approximates to orthotropic, which is a type of anisotropy in which the internal structure of the material results in unique elastic behavior along each of the three orthogonal axes of the material. In this case, three elastic (E) and shear modulus (G) and six Poisson’s ratios (v) are necessary for model input. (Richmond, Wright et al. 2005; Natali et al. 2010)

MATERIAL PROPERTIES	TRABECULAR BONE	CORTICAL BONE
E _x (MPa)	1,148	12,600
E _y (MPa)	210	12,600
E _z (MPa)	1,148	19,400
G _{xy} (MPa)	68	4,850
G _{yz} (MPa)	68	5,700
G _{xz} (MPa)	434	5,700
v _{yx}	0,010	0,300
v _{zy}	0,055	0,390
v _{zx}	0,322	0,390
v _{xy}	0,055	0,300
v _{yz}	0,010	0,253
v _{xz}	0,322	0,253

Table 1. Material properties used in an anisotropic model. The material axes correspond to the global coordinate system. E = Young’s modulus. G = shear modulus. v_{yx} = Poisson’s ratio for strain in the y-direction when loaded in the x-direction.

Since in dental implantology bone quality has been related to the structural efficiency of the cortical and trabecular bone architecture and ratio (lower bone quality results in biomechanically challenged treatments), mechanical simulations of poor bone quality critical as clinical studies have shown that dental implants placed in regions of the jaw bones with lower density have a higher chance to fail than implants placed at regions with higher bone density. (Genna 2003)

To date, no consensus regarding the mechanical properties that are appropriate for simulating the different bone density scenarios clinically encountered in implant dentistry has been reached. For instance, the value of trabecular bone elastic modulus observed in the literature range from 0.3 to 9.5 GPa. (Zarone et al. 2003; Eskitascioglu et al. 2004; Sevimay et al. 2005; Yokoyama, Wakabayashi et al. 2005) A different approach has been employed by Tada and coworkers (Tada et al. 2003) who assigned different elastic moduli to bone depending on its density from most dense (Type 1) to least dense (Type IV). The moduli utilized were 9.5 GPa, 5.5 GPa, 1.6 GPa and 0.69 GPa for bone types I, II, III, and IV, respectively. Recent work has also been carried out using these bone property values for different simulations (Table 2).

MATERIAL	YOUNG’S MODULUS (GPa)	POISSON’S RATIO	REFERENCES
Bone type I	9.5	0.3	Tada et al. (2003)
Bone type II	5.5	0.3	Tada et al. (2003)
Bone type III	1.6	0.3	Tada et al. (2003)
Bone type III	0.69	0.3	Tada et al. (2003)
Cortical bone	13.70	0.3	Shunmugasamy et al. (2011)
Titanium	110	0.35	Huang et al. (2008)

Table 2. Mechanical properties of the most used structures in implant prosthodontics research through finite element analysis.

Considering the fact that the most common bone quality types in mandible are type I or II in the anterior region and type III in the posterior region, the elastic modulus that would best represent the mandible would be 9.5 GPa or 5.5 GPa for the anterior region and 1.6 GPa for the posterior region. For the maxillary bone, the most prevalent types of bone are type II or III in the anterior region and type IV in the posterior region. Therefore, the Young’s modulus that would best represent these sites would be 5.5 GPa or 1.6 GPa for the anterior region and 0.69 GPa for the posterior region, respectively. For cortical bone, studies typically use the elastic modulus (E) of 13.7 GPa and the Poisson ratio (v) similar to 0.3 for the trabecular and cortical bone (Table 2). (Sevimay, Turhan et al. 2005; Huang, Hsu et al. 2008)

4. Bone-implant interface

In implant therapy, the fact that bone quantity and related biomechanical behavior differs for each patient implies a challenge to model the percentage of osseointegration. A critical issue when evaluating a study is to analyze the conditions specified for the interface. What is to be specified is the ability of the interface to resist three different types of stresses: compressive stresses at a right angle to the interface; tensile stresses, also at a right angle to

the interface; and shear stresses in parallel with the interface (Figure 4). (Hansson and Halldin 2009)

Fig. 4. The different types of stresses occurring at the implant-bone interface.

Linear static models have been employed extensively in previous FEA studies. (Kohal et al. 2002; Menicucci, Mossolov et al. 2002; Romanos 2004; Van Staden et al. 2006; Lim, Chew et al. 2008; Eser, Akca et al. 2009; Hsu et al. 2009; Li, Kong et al. 2009; Faegh and Muftu 2010; Freitas Junior et al. 2010; Hasan et al. 2011) These analyses usually assumed that all modeled volumes were bonded as one unit, which indicates that the trabecular and cortical bone are perfectly bonded to the implant interface. (Winter et al. 2004; Maeda et al. 2007; Fazel et al. 2009; Chang, Chen et al. 2010; Okumura, Stegaroiu et al. 2010). However, the validity of a linear static analysis may be questionable when the investigation aims to explore more realistic situations that are generally encountered in the dental implant field. Some actual implant clinical situations will give rise to nonlinearities, mainly related to the change of interrelations between the simulated structures of a FE model. (Wakabayashi et al. 2008) Moreover, frictional contact mode potentially provides an improved accuracy with respect to the relative component's micromotion within the implant system, and, therefore, a more reasonable representation of the real implant clinical condition. (Merz et al. 2000) This configuration allows minor displacements between all components of the model without interpenetration. Under these conditions, the contact zones transfer pressure and tangential forces (i.e. friction), but not tension. Some FE analyses have shown remarkable differences in the values and even in the distribution of stresses between "fixed bond" and "non-linear contact" interface conditions. (Brunski 1992; Van Oosterwyck et al. 1998; Huang, Hsu et al. 2008) Not only the stress and strain levels but also the stress and strain highly affected by the interface state (Figure 5). Van Oosterwyck et al. (Van Oosterwyck, Duyck et al. 1998) argued that through the bonded interface the force was dissipated evenly in both the compressive and the tension site. However, on the contact interfaces, tensions are not transferred and force is only passed on through the compressive site, which results in excessive stresses. Additionally, the condition of the bone to implant interface also influences the strain distribution and level inside of the implant system. (Pessoa et al. 2010)

Fig. 5. Equivalent strain ($\mu\epsilon$) distribution for 100N loaded implant in a superior central incisor region, in a median buccopalatal plane. The bone to implant interfaces were assumed as fixed bond ("glued") (A) and frictional contact (B). The arrows indicate the loading direction for clarity.

Conventionally, the osseointegrated bone to implant interface is treated as fully bonded. (Kohal, Papavasiliou et al. 2002; Menicucci, Mossolov et al. 2002; Romanos 2004; Van Staden, Guan et al. 2006; Lim, Chew et al. 2008; Eser, Akca et al. 2009; Hsu, Fuh et al. 2009; Li, Kong et al. 2009; Faegh and Muftu 2010; Freitas Junior, Rocha et al. 2010; Hasan, Rahimi et al. 2011)

This assumption is supported by experimental investigations in which removal of rough implants frequently resulted in fractures in bone distant from the implant surface, (Gottfredsen et al. 2000) suggesting the existence of an implant-bone "bond". On the other hand, frictional contact elements are used to simulate a nonintegrated bone to implant interface (i.e. in immediately loaded protocols), which allows minor displacements between the implant and the bone. (Pessoa, Muraru et al. 2010) The occurrence of relative motion between implant and bone introduces a source of non-linearity in FEA, since the contact conditions will change during load application.

The friction coefficient (μ) to be used in such simulations depends on many factors including mechanical properties and the roughness of the contact interface, exposure to interfacial contaminants (Williams 2000) and in some cases the normal load. A $\mu = 0.3$ was measured for interfaces between a smooth metal surface and bone, while a $\mu = 0.45$, for interfaces between a rough metal surface and bone. (Rancourt et al. 1990) Frequently, the friction coefficient between bone and implant is assumed as being $\mu = 0.3$. Huang et al. (Huang, Hsu et al. 2008) investigated the effects of different frictional coefficients ($\mu = 0.3, 0.45$ and 1) on the stress and displacement of an immediately loaded implant simulation. The authors

demonstrated that the value of μ shows no significant influence for increasing or decreasing the tensile and compressive stresses of bone. Nevertheless, increasing μ from 0.3 to 1, the interfacial sliding between implant and bone was mainly reduced from 20% to 30–60%, depending on the implant design. (Natali, Carniel et al. 2009)

Moreover, when an implant is surgically placed into the jawbone, the implant is mechanically screwed into a drilled hole of a smaller diameter. Large stresses will occur due to the torque applied in the process of implant insertion. As the implant stability and stress state around an immediately loaded implant may be influenced by such conditions, this should be also considered in FE simulation of immediately loaded implants. However, this phenomenon has not been thoroughly investigated to date. The implementation of such implant insertion stresses in FEA is still unclear and should therefore be a matter of further investigations.

Another commonly observed assumption in dental implantology FE models, perfect bonding between implant, abutment and abutment screw also is not the most realistic scenario. Non-linear contact analysis was proven to be the most effective interface condition for realistically simulating the relative micromotions occurring between different components within the implant system. (Williams 2000; Pessoa, Muraru et al. 2010; Pessoa et al. 2010) Therefore, for correct simulation of an implant-abutment connection, frictional contact should be defined between the implant components. Accordingly, between implant, abutment and abutment screw regions in contact, a frictional coefficient of 0.5 was generally assumed in non-linear simulations of implant-abutment connection. (Merz, Hunenbart et al. 2000; Lin et al. 2007; Pessoa, Muraru et al. 2010) When using a contact interface between implant components, for lateral or oblique loading conditions, specific parts can separate, or new parts that were initially not in contact can come into contact. Consequently, higher stress levels may be expected to occur in an implant-abutment connection simulated with contact interfaces, compared to a glued connection. In this regard, the pattern and magnitude of deformation in both periimplant bone and implant components will be influenced by the implant connection design. (Pessoa, Vaz et al. 2010)

5. Mesh and convergence analysis

Since the components in a dental implant-bone system are complex from a geometric standpoint, FEA has been viewed as the most suitable tool for mechanically analyzing them. A mesh is needed in FEA to divide the whole domain into elements. The process of creating the mesh, elements, their respective nodes, and defining boundary conditions is called “discretization” of the problem domain. (Geng et al. 2001; Richmond, Wright et al. 2005)

The 2D structures are typically meshed with triangular or quadrilateral elements (Figure 6A). These elements may possess two or more nodes per side, with one node at each vertex. If nodes are only placed at the vertices, the element is called linear because a line function describes the element geometry and how the displacement field will vary along an element edge. (Richmond, Wright et al. 2005)

Typical 3D elements include eight-node bricks, six-node wedges, five-node pyramids, and four-node tetrahedral (Figure 6B). With increasing numbers of elements and nodes, the model becomes more complex and computationally more difficult and lengthy. (Assuncao et al. 2009)

Fig. 6. (A) 2D Finite element model meshed with triangular elements. (B) 3D Finite element model meshed with tetrahedral elements.

To address this problem, researchers often devise strategies to minimize computational expense, such as taking advantage of symmetry when possible by modeling only half the structure, using a generally coarse mesh with finer elements only near regions of geometric complexity or high stress and strain, and/or using a 2D model when it suffices. (Richmond, Wright et al. 2005)

A convergence study of the model is always important to verify the mesh quality. A measurement of convergence is the degree of difference in the total strain energy between two successive mesh refinements. (Hasan, Rahimi et al. 2011) When the difference in energy is less than some tolerable limit specified by the user, the solution is considered converged

Fig. 7. Maximum principal stress ($\sigma_{\max}$) in hypothetical convergence testing of mandible FE models meshed by 0.6 mm elements with no refinement (mesh 1) and with refinement levels 2, 3, 4 and 5 at a posterior implant region (meshes 2, 3, 4 and 5, respectively).

(Figure 7). Some authors considered that the convergence criterion between meshes refinement was a change of less than 5 (Li, Kong et al. 2009; Lin et al. 2010) or 6% (Huang, Hsu et al. 2008; Huang et al. 2009) in the maximum simulated stress of the bone-implant edge. Alternatively, a mesh can be considered converged when the rate of density change was less than 10^{-5} between subsequent iterations. (Hasan, Rahimi et al. 2011)

6. Loading and boundary conditions

Attempting to successfully replicate the clinical situation that an implant might encounter in the oral environment, it is also important to understand and correctly reproduce the natural forces that are exerted throughout the system. These forces are mainly the result of the masticatory muscles action, and are related to the amount, frequency and duration of the masticatory function.

Forces acting on dental implants possess both magnitude and direction, and are referred to as vector quantities. For accurate predictions on the implant-bone behavior, it is essential to determine realistic in vivo loading magnitudes and directions. However, at each specific bite point, bite forces can be generated in a wide range of directions. Also, although bite forces are generally act downward, toward the apex of the implant and thereby tending to compress the implant into the alveolar bone, tensile forces and bending moments may also be present depending on where the bite force is applied relative to the implant-supported prosthesis. This fact is even more important when the investigation aims to simulate multiple-implant treatment modalities, because of the geometric factors involving the restorations that are linking the implants, such as the existence of distal cantilevers. (Mericske-Stern et al. 1996; Fontijn-Tekamp et al. 1998) Nevertheless, although for implants used in single-tooth replacement simulations the in vivo forces ought to replicate the forces exerted on natural teeth, factors such as the width of the crown occlusal table, the height of the abutment above the bone level, and the angulation of the implant with respect to the occlusal plane will affect the value of the moment on the single-tooth implant.

A significant amount of investigations have assumed the direction of the load applied to the implant to be horizontal, vertical and oblique. However, the rationale for use of an oblique loading condition is based on finding the vertical (axial) forces directed to the implant system that are relatively low and well tolerated in comparison to oblique forces, which generate bending moments (Figure 8A). The combination of axial and transverse loading, termed as mixed loading, simulates practical conditions where the actual applied force may be inclined with respect to the implant axis and components can be solved in the longitudinal and transverse directions. (Shunmugasamy et al. 2011) These oblique forces have been considered more clinically realistic in FEA than vertical ones. (Chun et al. 2002; Pierrisnard et al. 2002)

Several investigators have tried to gain insight into implant loading magnitudes by performing tests using experimental, analytical, and computer-based simulations of various implant-supported prosthesis types. (Mericske-Stern, Piotti et al. 1996; Fontijn-Tekamp, Slagter et al. 1998; Duyck et al. 2000; Morneburg and Proschel 2002) Bite forces ranging from 50 to 400 N in the molar regions and 25 to 170 N in the incisor areas have been reported. These variations are influenced by patient's gender, muscle mass, exercise, diet, bite location, parafunction, number of teeth and implants, type of implant-supported prosthesis, physical status and age. (Duyck, Van Oosterwyck et al. 2000; Morneburg and Proschel 2002; Eser, Akca et al. 2009)

Fig. 8. (A) Transversal loading being applied on the lingual surface of the anterior crown. (B) Boundary conditions of the anterior maxilla at the upper and lateral sides. The green region represent the fixed support constrained of x , y and z directions (displacement = 0).

Ideally, the entire jawbone structure should be evaluated for its contribution to the force exerted onto the dental implant. However, since the simulation of the whole mandibular and maxillary bone is very elaborate, smaller models have been proposed. (Lin, Chang et al. 2007; Lin et al. 2008; Natali, Carniel et al. 2010; Pessoa et al. 2010). (Pierrisnard et al. 2003; Tada, Stegaroiu et al. 2003) It can be explained by the Saint Venant principles, the actual force system may be replaced by a equivalent load system and the distribution of stress and strains are only affected near the region of loading. (Ugural 2003) Thus, if the interest in the study is on the biomechanical peri-implant environment, the modeling of no more than the relevant segment of the bone is required (Figure 9). This procedure allows saving computing and modeling time. In addition, Teixeira et al. (Teixeira et al. 1998) demonstrated by a 3D FEA that modeling the mandible at a distance greater than 4.2 mm medially or distally from the implant did not result in any significant improvement in accuracy. Hence, besides the application of a proper implant loading, the determination of restrictions to the model displacement compatible with the anatomic segment to be simulated is advised. Obviously, if necessary, expanding the domain of the model could reduce the effect of inaccurate modeling of the boundary condition. (Zhou et al. 1999)

It has been common to apply fixed constraints to the upper region of the maxilla and to its lateral sides during the construction of the FE model to simulate the continuity of the bone (Figure 8B). (Van Staden, Guan et al. 2006) These fixed supports represent the constrained of x , y and z directions (displacement = 0). (Chaichanasiri et al. 2009; Hsu, Fuh et al. 2009; Faegh and Muftu 2010; Wu, Liao et al. 2010)

Ishigaki and coworkers (Ishigaki et al. 2003) determined the directions of displacement constrains, which were applied to the jawbone according to the angles of the closing pathways of chopping type and grinding type chewing patterns, where the models were constrained at the base of the maxillary first molar to avoid sliding of the entire model.

Shunmugasamy and coworkers (Shunmugasamy, Gupta et al. 2011) simulated different macro-geometries of unit implants and fixed the base of the outer sides of the cortical bone.

Fig. 9. (A) 3D CT-based model of an upper central incisor extraction socket and an implant model positioned inside of the alveolus. (B) Proximal view. (C) Final model: only the relevant segment were included.

7. Validation and interpretation of FE modeling

The validation of the results is the final and most important step in FEA, depending on the degree to which the system reflects its biologic influence reality. (Richmond, Wright et al. 2005) For instance, a model would be precise but inaccurate if the mesh is exceedingly dense but the loading and boundary conditions are unrealistic. (Richmond, Wright et al. 2005)

Validation of FEA entails comparing the behavior of the model with *in vivo* or *in vitro* data gathered from parts of the modeled structures. A combination of *in vitro* and *in vivo* experimentation potentially offers the best validation. *In vitro* validation allows one to carefully control the loads and boundary conditions in order to assess the validity of the model's geometry and elastic properties. (Richmond, Wright et al. 2005) Studies of the consistency of numeric models and their agreement with biologic data are scarce (Mellal et al. 2004; Ozan et al. 2010) and the level of agreement and consistency between different engineering methods, especially those regarding quantified stress/strain, remains a concern. (Iplikcioglu et al. 2003)

8. Analysis criteria involved in steps of the FEA

The analysis criteria widely used in implant prosthodontics are: (1) von Mises stress (σ_{VM}) (Yokoyama, Wakabayashi et al. 2005; Assuncao, Tabata et al. 2008; Huang, Hsu et al. 2008; Lim, Chew et al. 2008; Rubo and Souza 2008; Schrotenboer et al. 2008; Assuncao et al. 2009; Assuncao et al. 2009; Chaichanasiri, Nanakorn et al. 2009; Ding et al. 2009; Fazel, Aalai et al. 2009; Gomes et al. 2009; Kong et al. 2009; Li, Kong et al. 2009; Qian et al. 2009; Abreu et al. 2010; Ao et al. 2010; Caglar et al. 2010; Eraslan and Inan 2010; Eser et al. 2010; Faegh and Muftu 2010; Gomes et al. 2010; Miyamoto et al. 2010; Okumura, Stegaroiu et al. 2010; Pessoa, Muraru et al. 2010; Sagat, Yalcin et al. 2010; Takahashi et al. 2010; Teixeira et al. 2010; Winter

et al. 2010; Bevilacqua et al. 2011; Burak Ozcelik, Ersoy et al. 2011; de Almeida, Rocha et al. 2011; Okumura et al. 2011), (2) maximum and minimum principal elastic strain ($\epsilon_{\max}$ and $\epsilon_{\min}$) (Saab et al. 2007; Qian, Todo et al. 2009; Chou, Muftu et al. 2010; Danza, Quaranta et al. 2010; Eser, Tonuk et al. 2010; Limbert et al. 2010; Okumura, Stegaroiu et al. 2010; Pessoa, Muraru et al. 2010; de Almeida, Rocha et al. 2011) and (3) maximum and minimum principal stress ($\sigma_{\max}$ and $\sigma_{\min}$). (de Almeida, Rocha et al. 2010; Degerliyurt et al. 2010; Hudieb et al. 2010; Jofre et al. 2010; Wu, Liao et al. 2010; de Almeida, Rocha et al. 2011)

Although the majority of the studies have used the σ_{VM} for evaluation of bone interface stress, several studies suggest that the magnitudes of the concentrations should be presented in the $\sigma_{\max}$ for evaluation of stress distribution in a brittle structure as bone, (de Almeida, Rocha et al. 2010; Degerliyurt, Simsek et al. 2010; Wu, Liao et al. 2010) as this criterion offers the possibility of making a distinction between tensile stress and compressive stress. (Degerliyurt, Simsek et al. 2010) In addition, displacement components of specific points may provide information about the deformation of the model and facilitate interpretation of the results. Principal stress values of fragile compact bone can be compared with its ultimate compressive strength and ultimate tensile strength values. (Ciftci and Canay 2000; Furmanski et al. 2009)

In fact, interfacial failure and bone resorption under different stress types are attributed to different mechanisms. Accordingly, it may be erroneous to emphasize the peak compressive or σ_{VM} without considering the risks of the tensile and the shear stresses at the interface. (Hudieb, Wakabayashi et al. 2010) However, when the titanium component is available as abutment, screw and implant, which are ductile structures, σ_{VM} is a recommended analysis criterion. (Cattaneo et al. 2005; Degerliyurt, Simsek et al. 2010; Wu, Liao et al. 2010)

Considering the FEA characteristics described above, the main advantage is the virtual simulation of real structures that are difficult to be clinically evaluated, i.e. stress and strain distribution on periimplant bone. Moreover, it is a low cost alternative in comparison to other *in vitro* methods since only a virtual model is used. Limitations of the FEA models include mainly the patient specific anatomy, and parameters such as the wet environment and damage accumulation under repetitive loading.

9. Summary and final remarks

The success of a dental implant depends on a variety of biomechanical factors including the design and position of the implant, implant-abutment connection, cantilever length, surface roughness, bone quality and type, depth of insertion, arch configuration, the nature of bone-implant interface, and occlusal conditions. (Randow, Glantz et al. 1986; Adell, Eriksson et al. 1990; van Steenberghe, Lekholm et al. 1990; Bozkaya and Muftu 2004; Bozkaya, Muftu et al. 2004; Misch et al. 2005; Yokoyama, Wakabayashi et al. 2005; De Smet et al. 2007; Abreu, Spazzin et al. 2010; Turkyilmaz 2011) All these biomechanical factors have been simulated by FEA in previous studies. While an increased number and quality of investigations has been published over the last decade, the results are often contradictory due to differences in model construction and meshing.

The biomechanical behavior of all components used in implant prosthodontics has been regarded as an important factor in determining the life expectancy of the restoration. Although substantial improvement has been made concerning implant/restorative

component design, further improvement is achievable through better understanding the different dental implant treatment modalities' biomechanical behavior. While the field has gained key knowledge in model fabrication (experimental designing) and improvements in all steps here described concerning the basic FEA experimental design, it is expected that all steps will be further refined based on future gains in computer power and correlations made between modeling results and clinical observation, ultimately providing improved care for patients in need of oral rehabilitation.

10. References

- Abreu, R. T., A. O. Spazzin, P. Y. Noritomi, et al. (2010). Influence of material of overdenture-retaining bar with vertical misfit on three-dimensional stress distribution, *Journal of prosthodontics: official journal of the American College of Prosthodontists*, Vol.19, No.6, pp. 425-431, 1532-849X (Electronic) 1059-941X (Linking).
- Adell, R., B. Eriksson, U. Lekholm, et al. (1990). Long-term follow-up study of osseointegrated implants in the treatment of totally edentulous jaws, *The International journal of oral & maxillofacial implants*, Vol.5, No.4, pp. 347-359, 0882-2786 (Print) 0882-2786 (Linking).
- Ao, J., T. Li, Y. Liu, et al. (2010). Optimal design of thread height and width on an immediately loaded cylinder implant: a finite element analysis, *Computers in biology and medicine*, Vol.40, No.8, pp. 681-686, 1879-0534 (Electronic) 0010-4825 (Linking).
- Assuncao, W. G., V. A. Barao, L. F. Tabata, et al. (2009). Comparison between complete denture and implant-retained overdenture: effect of different mucosa thickness and resiliency on stress distribution, *Gerodontology*, Vol.26, No.4, pp. 273-281, 1741-2358 (Electronic) 0734-0664 (Linking).
- Assuncao, W. G., V. A. Barao, L. F. Tabata, et al. (2009). Biomechanics studies in dentistry: bioengineering applied in oral implantology, *The Journal of craniofacial surgery*, Vol.20, No.4, pp. 1173-1177, 1536-3732 (Electronic) 1049-2275 (Linking).
- Assuncao, W. G., E. A. Gomes, V. A. Barao, et al. (2009). Stress analysis in simulation models with or without implant threads representation, *The International journal of oral & maxillofacial implants*, Vol.24, No.6, pp. 1040-1044, 0882-2786 (Print) 0882-2786 (Linking).
- Assuncao, W. G., L. F. Tabata, V. A. Barao, et al. (2008). Comparison of stress distribution between complete denture and implant-retained overdenture-2D FEA, *J Oral Rehabil*, Vol.35, No.10, pp. 766-774, 1365-2842 (Electronic) 0305-182X (Linking).
- Assuncao, W. G., L. F. Tabata, V. A. Barao, et al. (2008). Comparison of stress distribution between complete denture and implant-retained overdenture-2D FEA, *Journal of oral rehabilitation*, Vol.35, No.10, pp. 766-774, 1365-2842 (Electronic) 0305-182X (Linking).
- Bevilacqua, M., T. Tealdo, M. Menini, et al. (2011). The influence of cantilever length and implant inclination on stress distribution in maxillary implant-supported fixed dentures, *The Journal of prosthetic dentistry*, Vol.105, No.1, pp. 5-13, 1097-6841 (Electronic) 0022-3913 (Linking).
- Bozini, T., H. Petridis & K. Garefis (2011). A meta-analysis of prosthodontic complication rates of implant-supported fixed dental prostheses in edentulous patients after an

- observation period of at least 5 years, *The International journal of oral & maxillofacial implants*, Vol.26, No.2, pp. 304-318, 0882-2786 (Print) 0882-2786 (Linking).
- Bozkaya, D.&S. Muftu (2004). Efficiency considerations for the purely tapered interference fit (TIF) abutments used in dental implants, *Journal of biomechanical engineering*, Vol.126, No.4, pp. 393-401, 0148-0731 (Print) 0148-0731 (Linking).
- Bozkaya, D., S. Muftu&A. Muftu (2004). Evaluation of load transfer characteristics of five different implants in compact bone at different load levels by finite elements analysis, *The Journal of prosthetic dentistry*, Vol.92, No.6, pp. 523-530, 0022-3913 (Print) 0022-3913 (Linking).
- Branemark, P. I., R. Adell, U. Breine, et al. (1969). Intra-osseous anchorage of dental prostheses. I. Experimental studies, *Scand J Plast Reconstr Surg*, Vol.3, No.2, pp. 81-100, 0036-5556 (Print) 0036-5556 (Linking).
- Branemark, P. I., B. O. Hansson, R. Adell, et al. (1977). Osseointegrated implants in the treatment of the edentulous jaw. Experience from a 10-year period, *Scand J Plast Reconstr Surg Suppl*, Vol.16, pp. 1-132, 0581-9474 (Print) 0581-9474 (Linking).
- Brunski, J. B. (1992). Biomechanical factors affecting the bone-dental implant interface, *Clinical materials*, Vol.10, No.3, pp. 153-201, 0267-6605 (Print) 0267-6605 (Linking).
- Burak Ozcelik, T., E. Ersoy&B. Yilmaz Biomechanical evaluation of tooth- and implant-supported fixed dental prostheses with various nonrigid connector positions: a finite element analysis, *J Prosthodont*, Vol.20, No.1, pp. 16-28, 1532-849X (Electronic) 1059-941X (Linking).
- Burak Ozcelik, T., E. Ersoy&B. Yilmaz (2011). Biomechanical evaluation of tooth- and implant-supported fixed dental prostheses with various nonrigid connector positions: a finite element analysis, *Journal of prosthodontics : official journal of the American College of Prosthodontists*, Vol.20, No.1, pp. 16-28, 1532-849X (Electronic) 1059-941X (Linking).
- Caglar, A., B. T. Bal, C. Aydin, et al. (2010). Evaluation of stresses occurring on three different zirconia dental implants: three-dimensional finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.25, No.1, pp. 95-103, 0882-2786 (Print) 0882-2786 (Linking).
- Canay, S.&K. Akca (2009). Biomechanical aspects of bone-level diameter shifting at implant-abutment interface, *Implant dentistry*, Vol.18, No.3, pp. 239-248, 1538-2982 (Electronic) 1056-6163 (Linking).
- Cattaneo, P. M., M. Dalstra&B. Melsen (2005). The finite element method: a tool to study orthodontic tooth movement, *J Dent Res*, Vol.84, No.5, pp. 428-433, 0022-0345 (Print) 0022-0345 (Linking).
- Chaichanasiri, E., P. Nanakorn, W. Tharanon, et al. (2009). Finite element analysis of bone around a dental implant supporting a crown with a premature contact, *Journal of the Medical Association of Thailand = Chotmaihet thangphaet*, Vol.92, No.10, pp. 1336-1344, 0125-2208 (Print) 0125-2208 (Linking).
- Chang, C. L., C. S. Chen&M. L. Hsu (2010). Biomechanical effect of platform switching in implant dentistry: a three-dimensional finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.25, No.2, pp. 295-304, 0882-2786 (Print) 0882-2786 (Linking).

- Chen, J. H., C. Liu, L. You, et al. (2010). Boning up on Wolff's Law: mechanical regulation of the cells that make and maintain bone, *Journal of biomechanics*, Vol.43, No.1, pp. 108-118, 1873-2380 (Electronic) 0021-9290 (Linking).
- Chou, H. Y., S. Muftu&D. Bozkaya (2010). Combined effects of implant insertion depth and alveolar bone quality on periimplant bone strain induced by a wide-diameter, short implant and a narrow-diameter, long implant, *The Journal of prosthetic dentistry*, Vol.104, No.5, pp. 293-300, 1097-6841 (Electronic) 0022-3913 (Linking).
- Chun, H. J., S. Y. Cheong, J. H. Han, et al. (2002). Evaluation of design parameters of osseointegrated dental implants using finite element analysis, *Journal of oral rehabilitation*, Vol.29, No.6, pp. 565-574, 0305-182X (Print) 0305-182X (Linking).
- Ciftci, Y.&S. Canay (2000). The effect of veneering materials on stress distribution in implant-supported fixed prosthetic restorations, *The International journal of oral & maxillofacial implants*, Vol.15, No.4, pp. 571-582, 0882-2786 (Print) 0882-2786 (Linking).
- Cohen, Z. A., D. M. McCarthy, S. D. Kwak, et al. (1999). Knee cartilage topography, thickness, and contact areas from MRI: in-vitro calibration and in-vivo measurements, *Osteoarthritis and cartilage / OARS, Osteoarthritis Research Society*, Vol.7, No.1, pp. 95-109, 1063-4584 (Print) 1063-4584 (Linking).
- Danza, M., A. Quaranta, F. Carinci, et al. (2010). Biomechanical evaluation of dental implants in D1 and D4 bone by Finite Element Analysis, *Minerva stomatologica*, Vol.59, No.6, pp. 305-313, 0026-4970 (Print) 0026-4970 (Linking).
- de Almeida, E. O., E. P. Rocha, W. G. Assuncao, et al. (2011). Cortical bone stress distribution in mandibles with different configurations restored with prefabricated bar-prosthesis protocol: a three-dimensional finite-element analysis, *Journal of prosthodontics : official journal of the American College of Prosthodontists*, Vol.20, No.1, pp. 29-34, 1532-849X (Electronic) 1059-941X (Linking).
- de Almeida, E. O., E. P. Rocha, A. C. Freitas, Jr., et al. (2010). Finite element stress analysis of edentulous mandibles with different bone types supporting multiple-implant superstructures, *The International journal of oral & maxillofacial implants*, Vol.25, No.6, pp. 1108-1114, 0882-2786 (Print) 0882-2786 (Linking).
- De Smet, E., S. V. Jaecques, J. J. Jansen, et al. (2007). Effect of constant strain rate, composed of varying amplitude and frequency, of early loading on peri-implant bone (re)modelling, *Journal of clinical periodontology*, Vol.34, No.7, pp. 618-624, 0303-6979 (Print) 0303-6979 (Linking).
- Degerliyurt, K., B. Simsek, E. Erkmen, et al. (2010). Effects of different fixture geometries on the stress distribution in mandibular peri-implant structures: a 3-dimensional finite element analysis, *Oral surgery, oral medicine, oral pathology, oral radiology, and endodontics*, Vol.110, No.2, pp. e1-11, 1528-395X (Electronic) 1079-2104 (Linking).
- Ding, X., X. H. Zhu, S. H. Liao, et al. (2009). Implant-bone interface stress distribution in immediately loaded implants of different diameters: a three-dimensional finite element analysis, *Journal of prosthodontics : official journal of the American College of Prosthodontists*, Vol.18, No.5, pp. 393-402, 1532-849X (Electronic) 1059-941X (Linking).
- Duyck, J., H. Van Oosterwyck, J. Vander Sloten, et al. (2000). Magnitude and distribution of occlusal forces on oral implants supporting fixed prostheses: an in vivo study,

- Clinical oral implants research*, Vol.11, No.5, pp. 465-475, 0905-7161 (Print) 0905-7161 (Linking).
- Eraslan, O.&O. Inan (2010). The effect of thread design on stress distribution in a solid screw implant: a 3D finite element analysis, *Clinical oral investigations*, Vol.14, No.4, pp. 411-416, 1436-3771 (Electronic) 1432-6981 (Linking).
- Eser, A., K. Akca, S. Eckert, et al. (2009). Nonlinear finite element analysis versus ex vivo strain gauge measurements on immediately loaded implants, *The International journal of oral & maxillofacial implants*, Vol.24, No.3, pp. 439-446, 0882-2786 (Print) 0882-2786 (Linking).
- Eser, A., E. Tonuk, K. Akca, et al. (2010). Predicting time-dependent remodeling of bone around immediately loaded dental implants with different designs, *Medical engineering & physics*, Vol.32, No.1, pp. 22-31, 1873-4030 (Electronic) 1350-4533 (Linking).
- Eskitascioglu, G., A. Usumez, M. Sevimay, et al. (2004). The influence of occlusal loading location on stresses transferred to implant-supported prostheses and supporting bone: A three-dimensional finite element study, *The Journal of prosthetic dentistry*, Vol.91, No.2, pp. 144-150, 0022-3913 (Print) 0022-3913 (Linking).
- Esposito, M., J. M. Hirsch, U. Lekholm, et al. (1998). Biological factors contributing to failures of osseointegrated oral implants. (II). Etiopathogenesis, *Eur J Oral Sci*, Vol.106, No.3, pp. 721-764, 0909-8836 (Print) 0909-8836 (Linking).
- Faegh, S.&S. Muftu (2010). Load transfer along the bone-dental implant interface, *Journal of biomechanics*, Vol.43, No.9, pp. 1761-1770, 1873-2380 (Electronic) 0021-9290 (Linking).
- Fajardo, R. J., T. M. Ryan&J. Kappelman (2002). Assessing the accuracy of high-resolution X-ray computed tomography of primate trabecular bone by comparisons with histological sections, *Am J Phys Anthropol*, Vol.118, No.1, pp. 1-10, 0002-9483 (Print) 0002-9483 (Linking).
- Fazel, A., S. Aalai&M. Rismanchian (2009). Effect of macro-design of immediately loaded implants on micromotion and stress distribution in surrounding bone using finite element analysis, *Implant dentistry*, Vol.18, No.4, pp. 345-352, 1538-2982 (Electronic) 1056-6163 (Linking).
- Fontijn-Tekamp, F. A., A. P. Slagter, M. A. van't Hof, et al. (1998). Bite forces with mandibular implant-retained overdentures, *Journal of dental research*, Vol.77, No.10, pp. 1832-1839, 0022-0345 (Print) 0022-0345 (Linking).
- Freitas, A. C., Jr., E. P. Rocha, P. H. dos Santos, et al. (2010). All-ceramic crowns over single implant zircon abutment. Influence of young's modulus on mechanics, *Implant Dent*, Vol.19, No.6, pp. 539-548, 1538-2982 (Electronic) 1056-6163 (Linking).
- Freitas Junior, A. C., E. P. Rocha, P. H. Santos, et al. (2010). Mechanics of the maxillary central incisor. Influence of the periodontal ligament represented by beam elements, *Comput Methods Biomech Biomed Engin*, Vol.13, No.5, pp. 515-521, 1476-8259 (Electronic) 1025-5842 (Linking).
- Freitas Junior, A. C., E. P. Rocha, P. H. Santos, et al. (2010). Mechanics of the maxillary central incisor. Influence of the periodontal ligament represented by beam elements, *Computer methods in biomechanics and biomedical engineering*, Vol.13, No.5, pp. 515-521, 1476-8259 (Electronic) 1025-5842 (Linking).

- Furmanski, J., M. Anderson, S. Bal, et al. (2009). Clinical fracture of cross-linked UHMWPE acetabular liners, *Biomaterials*, Vol.30, No.29, pp. 5572-5582, 1878-5905 (Electronic) 0142-9612 (Linking).
- Galantucci, L. M., G. Percoco, G. Angelelli, et al. (2006). Reverse engineering techniques applied to a human skull, for CAD 3D reconstruction and physical replication by rapid prototyping, *J Med Eng Technol*, Vol.30, No.2, pp. 102-111, 0309-1902 (Print) 0309-1902 (Linking).
- Geng, J. P., K. B. Tan & G. R. Liu (2001). Application of finite element analysis in implant dentistry: a review of the literature, *The Journal of prosthetic dentistry*, Vol.85, No.6, pp. 585-598, 0022-3913 (Print) 0022-3913 (Linking).
- Genna, F. (2003). On the effects of cyclic transversal forces on osseointegrated dental implants: experimental and finite element shakedown analyses, *Computer methods in biomechanics and biomedical engineering*, Vol.6, No.2, pp. 141-152, 1025-5842 (Print) 1025-5842 (Linking).
- Gomes, E. A., W. G. Assuncao, V. A. Barao, et al. (2010). Passivity versus unilateral angular misfit: evaluation of stress distribution on implant-supported single crowns: three-dimensional finite element analysis, *The Journal of craniofacial surgery*, Vol.21, No.6, pp. 1683-1687, 1536-3732 (Electronic) 1049-2275 (Linking).
- Gomes, E. A., W. G. Assuncao, L. F. Tabata, et al. (2009). Effect of passive fit absence in the prosthesis/implant/retaining screw system: a two-dimensional finite element analysis, *The Journal of craniofacial surgery*, Vol.20, No.6, pp. 2000-2005, 1536-3732 (Electronic) 1049-2275 (Linking).
- Gottfredsen, K., T. Berglundh & J. Lindhe (2000). Anchorage of titanium implants with different surface characteristics: an experimental study in rabbits, *Clinical implant dentistry and related research*, Vol.2, No.3, pp. 120-128, 1523-0899 (Print) 1523-0899 (Linking).
- Greco, G. D., W. C. Jansen, J. Landre Junior, et al. (2009). Stress analysis on the free-end distal extension of an implant-supported mandibular complete denture, *Brazilian oral research*, Vol.23, No.2, pp. 182-189, 1807-3107 (Electronic) 1806-8324 (Linking).
- Guan, H., R. van Staden, Y. C. Loo, et al. (2009). Influence of bone and dental implant parameters on stress distribution in the mandible: a finite element study, *The International journal of oral & maxillofacial implants*, Vol.24, No.5, pp. 866-876, 0882-2786 (Print) 0882-2786 (Linking).
- Hansson, S. & A. Halldin (2009). Re: effect of microthreads and platform switching on crestal bone stress levels: a finite element analysis, *Journal of periodontology*, Vol.80, No.7, pp. 1033-1035; authors response 1035-1036, 0022-3492 (Print) 0022-3492 (Linking).
- Hasan, I., A. Rahimi, L. Keilig, et al. (2011). Computational simulation of internal bone remodelling around dental implants: a sensitivity analysis, *Computer methods in biomechanics and biomedical engineering*, pp. 1, 1476-8259 (Electronic) 1025-5842 (Linking).
- Hsu, J. T., L. J. Fuh, D. J. Lin, et al. (2009). Bone strain and interfacial sliding analyses of platform switching and implant diameter on an immediately loaded implant: experimental and three-dimensional finite element analyses, *Journal of periodontology*, Vol.80, No.7, pp. 1125-1132, 0022-3492 (Print) 0022-3492 (Linking).
- Huang, H. L., C. H. Chang, J. T. Hsu, et al. (2007). Comparison of implant body designs and threaded designs of dental implants: a 3-dimensional finite element analysis, *The*

- International journal of oral & maxillofacial implants*, Vol.22, No.4, pp. 551-562, 0882-2786 (Print) 0882-2786 (Linking).
- Huang, H. L., L. J. Fuh, C. C. Ko, et al. (2009). Biomechanical effects of a maxillary implant in the augmented sinus: a three-dimensional finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.24, No.3, pp. 455-462, 0882-2786 (Print) 0882-2786 (Linking).
- Huang, H. L., J. T. Hsu, L. J. Fuh, et al. (2008). Bone stress and interfacial sliding analysis of implant designs on an immediately loaded maxillary implant: a non-linear finite element study, *Journal of dentistry*, Vol.36, No.6, pp. 409-417, 0300-5712 (Print) 0300-5712 (Linking).
- Hudieb, M. I., N. Wakabayashi&S. Kasugai (2010). Magnitude and Direction of Mechanical Stress at the Osseointegrated Interface of the Microthread Implant, *Journal of periodontology*, pp. 1943-3670 (Electronic) 0022-3492 (Linking).
- Hudieb, M. I., N. Wakabayashi&S. Kasugai (2011). Magnitude and direction of mechanical stress at the osseointegrated interface of the microthread implant, *Journal of periodontology*, Vol.82, No.7, pp. 1061-1070, 1943-3670 (Electronic) 0022-3492 (Linking).
- Ipplikcioglu, H., K. Akca, M. C. Cehreli, et al. (2003). Comparison of non-linear finite element stress analysis with in vitro strain gauge measurements on a Morse taper implant, *The International journal of oral & maxillofacial implants*, Vol.18, No.2, pp. 258-265, 0882-2786 (Print) 0882-2786 (Linking).
- Ishigaki, S., T. Nakano, S. Yamada, et al. (2003). Biomechanical stress in bone surrounding an implant under simulated chewing, *Clinical oral implants research*, Vol.14, No.1, pp. 97-102, 0905-7161 (Print) 0905-7161 (Linking).
- Jofre, J., P. Cendoya&P. Munoz (2010). Effect of splinting mini-implants on marginal bone loss: a biomechanical model and clinical randomized study with mandibular overdentures, *The International journal of oral & maxillofacial implants*, Vol.25, No.6, pp. 1137-1144, 0882-2786 (Print) 0882-2786 (Linking).
- Kappelman, J. (1998). Advances in three-dimensional data acquisition and analysis. Primate locomotion: recent advances. E. F. Strasser, J.G.; Rosenberger, A.; McHenry, H.M. New York, Plenum Press: 205-222.
- Kohal, R. J., G. Papavasiliou, P. Kamposiora, et al. (2002). Three-dimensional computerized stress analysis of commercially pure titanium and yttrium-partially stabilized zirconia implants, *The International journal of prosthodontics*, Vol.15, No.2, pp. 189-194, 0893-2174 (Print) 0893-2174 (Linking).
- Kong, L., Z. Gu, T. Li, et al. (2009). Biomechanical optimization of implant diameter and length for immediate loading: a nonlinear finite element analysis, *The International journal of prosthodontics*, Vol.22, No.6, pp. 607-615, 0893-2174 (Print) 0893-2174 (Linking).
- Kozlovsky, A., H. Tal, B. Z. Laufer, et al. (2007). Impact of implant overloading on the peri-implant bone in inflamed and non-inflamed peri-implant mucosa, *Clinical oral implants research*, Vol.18, No.5, pp. 601-610, 0905-7161 (Print) 0905-7161 (Linking).
- Levine, R. A., D. S. Clem, 3rd, T. G. Wilson, Jr., et al. (1999). Multicenter retrospective analysis of the ITI implant system used for single-tooth replacements: results of loading for 2 or more years, *The International journal of oral & maxillofacial implants*, Vol.14, No.4, pp. 516-520, 0882-2786 (Print) 0882-2786 (Linking).

- Li, T., L. Kong, Y. Wang, et al. (2009). Selection of optimal dental implant diameter and length in type IV bone: a three-dimensional finite element analysis, *International journal of oral and maxillofacial surgery*, Vol.38, No.10, pp. 1077-1083, 1399-0020 (Electronic) 0901-5027 (Linking).
- Lim, K. H., C. M. Chew, P. C. Chen, et al. (2008). New extensometer to measure in vivo uniaxial mechanical properties of human skin, *Journal of biomechanics*, Vol.41, No.5, pp. 931-936, 0021-9290 (Print) 0021-9290 (Linking).
- Limbert, G., C. van Lierde, O. L. Muraru, et al. (2010). Trabecular bone strains around a dental implant and associated micromotions--a micro-CT-based three-dimensional finite element study, *Journal of biomechanics*, Vol.43, No.7, pp. 1251-1261, 1873-2380 (Electronic) 0021-9290 (Linking).
- Lin, C. L., S. H. Chang, W. J. Chang, et al. (2007). Factorial analysis of variables influencing mechanical characteristics of a single tooth implant placed in the maxilla using finite element analysis and the statistics-based Taguchi method, *European journal of oral sciences*, Vol.115, No.5, pp. 408-416, 0909-8836 (Print) 0909-8836 (Linking).
- Lin, C. L., S. H. Chang, J. C. Wang, et al. (2006). Mechanical interactions of an implant/tooth-supported system under different periodontal supports and number of splinted teeth with rigid and non-rigid connections, *Journal of dentistry*, Vol.34, No.9, pp. 682-691, 0300-5712 (Print) 0300-5712 (Linking).
- Lin, C. L., Y. H. Lin & S. H. Chang (2010). Multi-factorial analysis of variables influencing the bone loss of an implant placed in the maxilla: prediction using FEA and SED bone remodeling algorithm, *Journal of biomechanics*, Vol.43, No.4, pp. 644-651, 1873-2380 (Electronic) 0021-9290 (Linking).
- Lin, C. L., J. C. Wang & W. J. Chang (2008). Biomechanical interactions in tooth-implant-supported fixed partial dentures with variations in the number of splinted teeth and connector type: a finite element analysis, *Clinical oral implants research*, Vol.19, No.1, pp. 107-117, 0905-7161 (Print) 0905-7161 (Linking).
- Lin, D., Q. Li, W. Li, et al. (2010). Mandibular bone remodeling induced by dental implant, *Journal of biomechanics*, Vol.43, No.2, pp. 287-293, 1873-2380 (Electronic) 0021-9290 (Linking).
- Lin, D., Q. Li, W. Li, et al. (2009). Dental implant induced bone remodeling and associated algorithms, *J Mech Behav Biomed Mater*, Vol.2, No.5, pp. 410-432, 1878-0180 (Electronic) 1878-0180 (Linking).
- Maeda, Y., J. Miura, I. Taki, et al. (2007). Biomechanical analysis on platform switching: is there any biomechanical rationale?, *Clinical oral implants research*, Vol.18, No.5, pp. 581-584, 0905-7161 (Print) 0905-7161 (Linking).
- Magne, P. (2007). Efficient 3D finite element analysis of dental restorative procedures using micro-CT data, *Dental materials : official publication of the Academy of Dental Materials*, Vol.23, No.5, pp. 539-548, 0109-5641 (Print) 0109-5641 (Linking).
- Mellal, A., H. W. Wiskott, J. Botsis, et al. (2004). Stimulating effect of implant loading on surrounding bone. Comparison of three numerical models and validation by in vivo data, *Clinical oral implants research*, Vol.15, No.2, pp. 239-248, 0905-7161 (Print) 0905-7161 (Linking).
- Menicucci, G., A. Mossolov, M. Mozzati, et al. (2002). Tooth-implant connection: some biomechanical aspects based on finite element analyses, *Clinical oral implants research*, Vol.13, No.3, pp. 334-341, 0905-7161 (Print) 0905-7161 (Linking).

- Mericske-Stern, R., M. Piotti&G. Sirtes (1996). 3-D in vivo force measurements on mandibular implants supporting overdentures. A comparative study, *Clinical oral implants research*, Vol.7, No.4, pp. 387-396, 0905-7161 (Print) 0905-7161 (Linking).
- Merz, B. R., S. Hunenbart&U. C. Belser (2000). Mechanics of the implant-abutment connection: an 8-degree taper compared to a butt joint connection, *The International journal of oral & maxillofacial implants*, Vol.15, No.4, pp. 519-526, 0882-2786 (Print) 0882-2786 (Linking).
- Misch, C. E., J. B. Suzuki, F. M. Misch-Dietsh, et al. (2005). A positive correlation between occlusal trauma and peri-implant bone loss: literature support, *Implant dentistry*, Vol.14, No.2, pp. 108-116, 1056-6163 (Print) 1056-6163 (Linking).
- Miyamoto, S., K. Ujigawa, Y. Kizu, et al. (2010). Biomechanical three-dimensional finite-element analysis of maxillary prostheses with implants. Design of number and position of implants for maxillary prostheses after hemimaxillectomy, *International journal of oral and maxillofacial surgery*, Vol.39, No.11, pp. 1120-1126, 1399-0020 (Electronic) 0901-5027 (Linking).
- Morneburg, T. R.&P. A. Proschel (2002). Measurement of masticatory forces and implant loads: a methodologic clinical study, *The International journal of prosthodontics*, Vol.15, No.1, pp. 20-27, 0893-2174 (Print) 0893-2174 (Linking).
- Natali, A. N., E. L. Carniel&P. G. Pavan (2009). Investigation of viscoelastoplastic response of bone tissue in oral implants press fit process, *J Biomed Mater Res B Appl Biomater*, Vol.91, No.2, pp. 868-875, 1552-4981 (Electronic) 1552-4973 (Linking).
- Natali, A. N., E. L. Carniel&P. G. Pavan (2010). Modelling of mandible bone properties in the numerical analysis of oral implant biomechanics, *Computer methods and programs in biomedicine*, Vol.100, No.2, pp. 158-165, 1872-7565 (Electronic) 0169-2607 (Linking).
- O'Mahony, A. M., J. L. Williams&P. Spencer (2001). Anisotropic elasticity of cortical and cancellous bone in the posterior mandible increases peri-implant stress and strain under oblique loading, *Clinical oral implants research*, Vol.12, No.6, pp. 648-657, 0905-7161 (Print) 0905-7161 (Linking).
- Okumura, N., R. Stegaroiu, E. Kitamura, et al. (2010). Influence of maxillary cortical bone thickness, implant design and implant diameter on stress around implants: a three-dimensional finite element analysis, *Journal of prosthodontic research*, Vol.54, No.3, pp. 133-142, 1883-1958 (Print) 1883-1958 (Linking).
- Okumura, N., R. Stegaroiu, H. Nishiyama, et al. (2011). Finite element analysis of implant-embedded maxilla model from CT data: comparison with the conventional model, *Journal of prosthodontic research*, Vol.55, No.1, pp. 24-31, 1883-1958 (Print) 1883-1958 (Linking).
- Ozan, F., H. Yildiz, O. A. Bora, et al. (2010). The effect of head trauma on fracture healing: biomechanical testing and finite element analysis, *Acta Orthop Traumatol Turc*, Vol.44, No.4, pp. 313-321, 1017-995X (Print) 1017-995X (Linking).
- Pessoa, R. S., L. Muraru, E. M. Junior, et al. (2010). Influence of implant connection type on the biomechanical environment of immediately placed implants - CT-based nonlinear, three-dimensional finite element analysis, *Clin Implant Dent Relat Res*, Vol.12, No.3, pp. 219-234, 1708-8208 (Electronic) 1523-0899 (Linking).
- Pessoa, R. S., L. Muraru, E. M. Junior, et al. (2010). Influence of implant connection type on the biomechanical environment of immediately placed implants - CT-based

- nonlinear, three-dimensional finite element analysis, *Clinical implant dentistry and related research*, Vol.12, No.3, pp. 219-234, 1708-8208 (Electronic) 1523-0899 (Linking).
- Pessoa, R. S., L. G. Vaz, E. Marcantonio, Jr., et al. (2010). Biomechanical evaluation of platform switching in different implant protocols: computed tomography-based three-dimensional finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.25, No.5, pp. 911-919, 0882-2786 (Print) 0882-2786 (Linking).
- Peterson, J.&P. C. Dechow (2003). Material properties of the human cranial vault and zygoma, *Anat Rec A Discov Mol Cell Evol Biol*, Vol.274, No.1, pp. 785-797, 1552-4884 (Print) 1552-4884 (Linking).
- Pieri, F., N. N. Aldini, C. Marchetti, et al. (2011). Influence of implant-abutment interface design on bone and soft tissue levels around immediately placed and restored single-tooth implants: a randomized controlled clinical trial, *The International journal of oral & maxillofacial implants*, Vol.26, No.1, pp. 169-178, 0882-2786 (Print) 0882-2786 (Linking).
- Pierrisnard, L., G. Hure, M. Barquins, et al. (2002). Two dental implants designed for immediate loading: a finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.17, No.3, pp. 353-362, 0882-2786 (Print) 0882-2786 (Linking).
- Pierrisnard, L., F. Renouard, P. Renault, et al. (2003). Influence of implant length and bicortical anchorage on implant stress distribution, *Clin Implant Dent Relat Res*, Vol.5, No.4, pp. 254-262, 1523-0899 (Print) 1523-0899 (Linking).
- Qian, L., M. Todo, Y. Matsushita, et al. (2009). Effects of implant diameter, insertion depth, and loading angle on stress/strain fields in implant/jawbone systems: finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.24, No.5, pp. 877-886, 0882-2786 (Print) 0882-2786 (Linking).
- Rancourt, D., A. Shirazi-Adl, G. Drouin, et al. (1990). Friction properties of the interface between porous-surfaced metals and tibial cancellous bone, *Journal of biomedical materials research*, Vol.24, No.11, pp. 1503-1519, 0021-9304 (Print) 0021-9304 (Linking).
- Random, K., P. O. Glantz&B. Zoger (1986). Technical failures and some related clinical complications in extensive fixed prosthodontics. An epidemiological study of long-term clinical quality, *Acta Odontol Scand*, Vol.44, No.4, pp. 241-255, 0001-6357 (Print) 0001-6357 (Linking).
- Rayfield, E. J. (2004). Cranial mechanics and feeding in *Tyrannosaurus rex*, *Proc Biol Sci*, Vol.271, No.1547, pp. 1451-1459, 0962-8452 (Print) 0962-8452 (Linking).
- Richmond, B. G., B. W. Wright, I. Grosse, et al. (2005). Finite element analysis in functional morphology, *Anat Rec A Discov Mol Cell Evol Biol*, Vol.283, No.2, pp. 259-274, 1552-4884 (Print) 1552-4884 (Linking).
- Roberts, D. H. (1970). The failure of retainers in bridge prostheses. An analysis of 2,000 retainers, *Br Dent J*, Vol.128, No.3, pp. 117-124, 0007-0610 (Print) 0007-0610 (Linking).
- Rodriguez-Ciurana, X., X. Vela-Nebot, M. Segala-Torres, et al. (2009). Biomechanical repercussions of bone resorption related to biologic width: a finite element analysis

- of three implant-abutment configurations, *The International journal of periodontics & restorative dentistry*, Vol.29, No.5, pp. 479-487, 0198-7569 (Print) 0198-7569 (Linking).
- Romanos, G. E. (2004). Present status of immediate loading of oral implants, *The Journal of oral implantology*, Vol.30, No.3, pp. 189-197, 0160-6972 (Print) 0160-6972 (Linking).
- Romeed, S. A., S. L. Fok & N. H. Wilson (2006). A comparison of 2D and 3D finite element analysis of a restored tooth, *J Oral Rehabil*, Vol.33, No.3, pp. 209-215, 0305-182X (Print) 0305-182X (Linking).
- Ross, C. F. (2005). Finite element analysis in vertebrate biomechanics, *Anat Rec A Discov Mol Cell Evol Biol*, Vol.283, No.2, pp. 253-258, 1552-4884 (Print) 1552-4884 (Linking).
- Rubo, J. H. & E. A. Souza (2008). Finite element analysis of stress in bone adjacent to dental implants, *The Journal of oral implantology*, Vol.34, No.5, pp. 248-255, 0160-6972 (Print) 0160-6972 (Linking).
- Ryan, T. M. & B. van Rietbergen (2005). Mechanical significance of femoral head trabecular bone structure in Loris and Galago evaluated using micromechanical finite element models, *Am J Phys Anthropol*, Vol.126, No.1, pp. 82-96, 0002-9483 (Print) 0002-9483 (Linking).
- Saab, X. E., J. A. Griggs, J. M. Powers, et al. (2007). Effect of abutment angulation on the strain on the bone around an implant in the anterior maxilla: a finite element study, *The Journal of prosthetic dentistry*, Vol.97, No.2, pp. 85-92, 0022-3913 (Print) 0022-3913 (Linking).
- Sagat, G., S. Yalcin, B. A. Gultekin, et al. (2010). Influence of arch shape and implant position on stress distribution around implants supporting fixed full-arch prosthesis in edentulous maxilla, *Implant dentistry*, Vol.19, No.6, pp. 498-508, 1538-2982 (Electronic) 1056-6163 (Linking).
- Schrotenboer, J., Y. P. Tsao, V. Kinariwala, et al. (2008). Effect of microthreads and platform switching on crestal bone stress levels: a finite element analysis, *Journal of periodontology*, Vol.79, No.11, pp. 2166-2172, 0022-3492 (Print) 0022-3492 (Linking).
- Sevimay, M., F. Turhan, M. A. Kilicarslan, et al. (2005). Three-dimensional finite element analysis of the effect of different bone quality on stress distribution in an implant-supported crown, *The Journal of prosthetic dentistry*, Vol.93, No.3, pp. 227-234, 0022-3913 (Print) 0022-3913 (Linking).
- Shunmugasamy, V. C., N. Gupta, R. S. Pessoa, et al. (2011). Influence of clinically relevant factors on the immediate biomechanical surrounding for a series of dental implant designs, *J Biomech Eng*, Vol.133, No.3, pp. 031005, 1528-8951 (Electronic) 0148-0731 (Linking).
- Silva, G. C., J. A. Mendonca, L. R. Lopes, et al. (2010). Stress patterns on implants in prostheses supported by four or six implants: a three-dimensional finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.25, No.2, pp. 239-246, 0882-2786 (Print) 0882-2786 (Linking).
- Sotto-Maior, B. S., E. P. Rocha, E. O. Almeida, et al. (2010). Influence of high insertion torque on implant placement: an anisotropic bone stress analysis, *Braz Dent J*, Vol.21, No.6, pp. 508-514, 1806-4760 (Electronic) 0103-6440 (Linking).
- Stanford, C. M. & R. A. Brand (1999). Toward an understanding of implant occlusion and strain adaptive bone modeling and remodeling, *The Journal of prosthetic dentistry*, Vol.81, No.5, pp. 553-561, 0022-3913 (Print) 0022-3913 (Linking).

- Tada, S., R. Stegaroiu, E. Kitamura, et al. (2003). Influence of implant design and bone quality on stress/strain distribution in bone around implants: a 3-dimensional finite element analysis, *The International journal of oral & maxillofacial implants*, Vol.18, No.3, pp. 357-368, 0882-2786 (Print) 0882-2786 (Linking).
- Takahashi, T., I. Shimamura&K. Sakurai (2010). Influence of number and inclination angle of implants on stress distribution in mandibular cortical bone with All-on-4 Concept, *Journal of prosthodontic research*, Vol.54, No.4, pp. 179-184, 1883-1958 (Print) 1883-1958 (Linking).
- Teixeira, E. R., Y. Sato, Y. Akagawa, et al. (1998). A comparative evaluation of mandibular finite element models with different lengths and elements for implant biomechanics, *Journal of oral rehabilitation*, Vol.25, No.4, pp. 299-303, 0305-182X (Print) 0305-182X (Linking).
- Teixeira, M. F., S. A. Ramalho, I. A. de Mattias Sartori, et al. (2010). Finite element analysis of 2 immediate loading systems in edentulous mandible: rigid and semirigid splinting of implants, *Implant dentistry*, Vol.19, No.1, pp. 39-49, 1538-2982 (Electronic) 1056-6163 (Linking).
- Turkyilmaz, I. (2011). 26-year follow-up of screw-retained fixed dental prostheses supported by machined-surface Branemark implants: a case report, *Texas dental journal*, Vol.128, No.1, pp. 15-19, 0040-4284 (Print) 0040-4284 (Linking).
- Ugural, A. C. F. S. (2003). *Advanced strength and applied elasticity*. New Jersey, Prentice-Hall.
- Van Oosterwyck, H., J. Duyck, J. Vander Sloten, et al. (1998). The influence of bone mechanical properties and implant fixation upon bone loading around oral implants, *Clinical oral implants research*, Vol.9, No.6, pp. 407-418, 0905-7161 (Print) 0905-7161 (Linking).
- Van Staden, R. C., H. Guan&Y. C. Loo (2006). Application of the finite element method in dental implant research, *Computer methods in biomechanics and biomedical engineering*, Vol.9, No.4, pp. 257-270, 1025-5842 (Print) 1025-5842 (Linking).
- van Steenberghe, D., U. Lekholm, C. Bolender, et al. (1990). Applicability of osseointegrated oral implants in the rehabilitation of partial edentulism: a prospective multicenter study on 558 fixtures, *The International journal of oral & maxillofacial implants*, Vol.5, No.3, pp. 272-281, 0882-2786 (Print) 0882-2786 (Linking).
- Wakabayashi, N., M. Ona, T. Suzuki, et al. (2008). Nonlinear finite element analyses: advances and challenges in dental applications, *Journal of dentistry*, Vol.36, No.7, pp. 463-471, 0300-5712 (Print) 0300-5712 (Linking).
- Walton, J. N., F. M. Gardner&J. R. Agar (1986). A survey of crown and fixed partial denture failures: length of service and reasons for replacement, *The Journal of prosthetic dentistry*, Vol.56, No.4, pp. 416-421, 0022-3913 (Print) 0022-3913 (Linking).
- Wang, Z. L., J. C. Teo, C. K. Chui, et al. (2005). Computational biomechanical modelling of the lumbar spine using marching-cubes surface smoothed finite element voxel meshing, *Computer methods and programs in biomedicine*, Vol.80, No.1, pp. 25-35, 0169-2607 (Print) 0169-2607 (Linking).
- Williams, J. A. (2000). *Engineering Tribology*. Oxford, Oxford University Press.
- Winter, W., S. M. Heckmann&H. P. Weber (2004). A time-dependent healing function for immediate loaded implants, *Journal of biomechanics*, Vol.37, No.12, pp. 1861-1867, 0021-9290 (Print) 0021-9290 (Linking).

- Winter, W., S. Mohrle, S. Holst, et al. (2010). Bone loading caused by different types of misfits of implant-supported fixed dental prostheses: a three-dimensional finite element analysis based on experimental results, *The International journal of oral & maxillofacial implants*, Vol.25, No.5, pp. 947-952, 0882-2786 (Print) 0882-2786 (Linking).
- Wu, T., W. Liao, N. Dai, et al. (2010). Design of a custom angled abutment for dental implants using computer-aided design and nonlinear finite element analysis, *Journal of biomechanics*, Vol.43, No.10, pp. 1941-1946, 1873-2380 (Electronic) 0021-9290 (Linking).
- Yang, H. S., L. A. Lang & D. A. Felton (1999). Finite element stress analysis on the effect of splinting in fixed partial dentures, *J Prosthet Dent*, Vol.81, No.6, pp. 721-728, 0022-3913 (Print) 0022-3913 (Linking).
- Yang, H. S., L. A. Lang, A. Molina, et al. (2001). The effects of dowel design and load direction on dowel-and-core restorations, *J Prosthet Dent*, Vol.85, No.6, pp. 558-567, 0022-3913 (Print) 0022-3913 (Linking).
- Yokoyama, S., N. Wakabayashi, M. Shiota, et al. (2005). Stress analysis in edentulous mandibular bone supporting implant-retained 1-piece or multiple superstructures, *The International journal of oral & maxillofacial implants*, Vol.20, No.4, pp. 578-583, 0882-2786 (Print) 0882-2786 (Linking).
- Zarone, F., A. Apicella, L. Nicolais, et al. (2003). Mandibular flexure and stress build-up in mandibular full-arch fixed prostheses supported by osseointegrated implants, *Clinical oral implants research*, Vol.14, No.1, pp. 103-114, 0905-7161 (Print) 0905-7161 (Linking).
- Zhou, X., Z. Zhao, M. Zhao, et al. (1999). [The boundary design of mandibular model by means of the three-dimensional finite element method], *Hua xi kou qiang yi xue za zhi* = *Huaxi kouqiang yixue zazhi* = *West China journal of stomatology*, Vol.17, No.1, pp. 29-32, 1000-1182 (Print) 1000-1182 (Linking).

IntechOpen

Finite Element Analysis - From Biomedical Applications to Industrial Developments

Edited by Dr. David Moratal

ISBN 978-953-51-0474-2

Hard cover, 496 pages

Publisher InTech

Published online 30, March, 2012

Published in print edition March, 2012

Finite Element Analysis represents a numerical technique for finding approximate solutions to partial differential equations as well as integral equations, permitting the numerical analysis of complex structures based on their material properties. This book presents 20 different chapters in the application of Finite Elements, ranging from Biomedical Engineering to Manufacturing Industry and Industrial Developments. It has been written at a level suitable for use in a graduate course on applications of finite element modelling and analysis (mechanical, civil and biomedical engineering studies, for instance), without excluding its use by researchers or professional engineers interested in the field, seeking to gain a deeper understanding concerning Finite Element Analysis.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Erika O. Almeida, Amilcar C. Freitas Júnior, Eduardo P. Rocha, Roberto S. Pessoa, Nikhil Gupta, Nick Tovar and Paulo G. Coelho (2012). Critical Aspects for Mechanical Simulation in Dental Implantology, Finite Element Analysis - From Biomedical Applications to Industrial Developments, Dr. David Moratal (Ed.), ISBN: 978-953-51-0474-2, InTech, Available from: <http://www.intechopen.com/books/finite-element-analysis-from-biomedical-applications-to-industrial-developments/critical-aspects-for-mechanical-simulation-in-dental-implantology>

INTECH
open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia
Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元
Phone: +86-21-62489820
Fax: +86-21-62489821

© 2012 The Author(s). Licensee IntechOpen. This is an open access article distributed under the terms of the [Creative Commons Attribution 3.0 License](https://creativecommons.org/licenses/by/3.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

IntechOpen

IntechOpen