We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,300

Open access books available

172,000

International authors and editors

190M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities

WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com

Simple Signals for System Identification

Paul Annus, Raul Land, Mart Min and Jaan Ojarand
Tallinn University of Technology / Eliko
Estonia

1. Introduction

Event accessible for observation can be investigated by examining timeline of certain measureable values. Usually the process is referred to as time domain analysis of signals. Such a signal can be of many different origins. It could be comprised of values of electrical current or voltage, mechanical displacement or force, value of stock or popularity of politicians, reaction of the patient to medication, and essentially anything which is observable, measurable and quantifiable with reasonable accuracy. Observer will quite intuitively search for certain patterns and periods in signal if phenomena changes along the timeline. Some of them are easily detectable, such as periodic variation of the sun's activity; some might appear as random fluctuations at the first glance, such as parameters of the seismic waves in Earth's crust. Of course sources of noise and other disturbances are usually omnipresent and will make matters much more difficult to observe and explain.

It is well known already from early days of modern science that signals can be analyzed in frequency domain as well or instead of time domain analysis. It turns out that for observable events both domains can be easily interchanged or in other words transformed one into other without any loss of information, and even joint time-frequency analysis can be performed.

First known usage of what is essentially known today as fast Fourier transform (FFT) is contributed to Johann Carl Friedrich Gauss (Goldstine, 1977, as cited in Heideman et al., 1984). In 1805 Gauss describes his computationally efficient method for interpolation of the orbits of celestial bodies. Intriguingly it was almost forgotten. Few years later, in 1807, Jean Baptiste Joseph Fourier, while interested in heat propagation, claimed that any continuous periodic signal could be represented as the sum of properly chosen sinusoids during his presentation to the Academy of Sciences in Paris. That claim ignited dispute between the reviewers of his paper: Pierre-Simon Laplace, and Joseph-Louis Lagrange, and delayed the publication until 1822 (Heideman et al., 1984). Joseph-Louis Lagrange's protests were based on the fact that such an approach could not be used to represent signals with sharp corners, or in other words with discontinuous slopes, such as square waves. Dispute lasted almost hundred years until Maxime Bôcher gave a detailed mathematical analysis of the phenomenon and named it after the Josiah Willard Gibbs (Bôcher, 1906). In essence both Lagrange and Fourier were right. While it is not possible to construct signals with sharp corners from sinusoids it is possible to get so close that the difference in energy between these signals is zero. If real signals from nature are concerned instead of exact and purely

mathematical curiosities problem is even smaller. So for all practical signal processing tasks it is indeed possible to state that any real signal can be constructed from sinusoids.

While these transforms, or more precisely their modern counterparts, work very well for signals emanating from natural phenomena, matter can also be investigated by deliberately exciting it with known signals and analyzing its response. This process may be called system identification, synchronous measurement, lock in measurement, or in some application areas something entirely different. The task may even be reversed, in a sense that system can be designed or modified by knowing what output is desirable to certain excitation. It could be filter design in electronics or eigensystem realization algorithm (ERA) in civil engineering, to name a few.

Topic in general is too broad to discuss in one book or even in series of books, and many good papers and books are already written on the topic (Godfrey, 1993; Pintelon & Schoukens, 2001; etc. to name a few), however when low complexity, limited energy consumption and highly optimized measurement systems are targeted new solutions are often warranted, and some of them are briefly discussed in the following pages. So emphasis is given to properties and practical design of custom excitation signals.

2. General considerations

Any system has large set of different parameters, and many subsets of them to be characterized according to requirements imposed by task at hand. Sometimes they can be measured separately and sequentially one at a time, but quite often not. Bulk of system identification theory is based on an assumption that systems are linear and time invariant (LTI), which they are not. Generally certain set of measurements has to be conducted within short enough timeframe for the system to remain reasonable motionless, and with signals which will only very moderately drive system under investigation (SUT) into the non-linear region of operation. Limited magnitudes of excitation signals and the need to consider frequency spectrum of these signals very carefully are both among implications rising from the last requirement. Unfortunately there is also third factor to consider. Disturbances from surrounding world, and noise impact, can effectively render useless any and all of the measurement results, if not dealt with care. Signal to noise ratio (SNR) is often used for numeric quantification of the problem. Similarly known signal to noise and distortion ratio (SINAD) does a little better when covering the problem area and considering likely non-linear behavior of the SUT as well. Shorter excitation time and limited excitation signal energy are always paired with lower SNR.

Therefore it is clear that whenever real systems are characterized, then choice of excitation signals is always a subject of optimization and compromise. Effectiveness of said optimization depends on level of prior knowledge. Things to consider include measurement conditions, SUT itself and the cost of the measurement. Successive approximation and adaptation can be considered when prior knowledge is limited, unless it is one of a kind or very rare event. Fortunately these very rare cases are indeed rare, and furthermore there is usually at least some amount of general prior knowledge available.

Real objects and systems are seldom fully homogenous and isotropic; therefore several measurements from different locations might be warranted for sufficient characterization of the SUT. Possibility of sequential in time measurements from different locations depends

largely on time variance of the parameters of the system. Such a variance could be caused by slow ageing, rapid decay, and fast spatial movement of the SUT relative to the measurement system, as well as by modulation of some of the parameters of the SUT with outside signals. Ideally variance should be excluded by taking readings from different locations simultaneously and rapidly enough. One way to achieve this is to conduct measurements at several slightly differing frequencies. In this case system properties between different points can be separated, and values will not vary much due to almost identical frequencies:

Fig. 1. System identification in case of two simultaneous excitation signal with slightly differing frequencies f_1 , and f_2 , injected from different points

Similar but different task is accomplished when system properties at vastly different frequencies are of interest. Again sequential in time measurements can be considered, but in case of fast variations in system properties simultaneous multifrequency measurement is essential:

Fig. 2. System identification in case of two excitation waveforms with highly differing frequencies f_1 , and f_2 , injected from the same point

In even more complex cases multisite and multifrequency measurement are needed simultaneously. Question arises whether it is possible to optimize such a complex measurement by applying different signals and processing methods. Following matter discusses some suitable signals with increasing complexity. For comparison both waveform and frequency content of these signals is given.

Frequency response of any system is comparison of the magnitudes and phases of the output signal spectral components with the input signal components in real world measurement situations. It is best viewed as complex function of frequency. Instead of magnitudes and phases it is often useful to represent the result in Cartesian coordinates. Link is straightforward:

Fig. 3. Complex number with unity magnitude in polar and Cartesian coordinates

If sinusoidal signal with frequency ω , magnitude A and relative phase φ compared to reference signal is multiplied by orthogonal set of reference sinusoids, i.e. $\sin(\omega t)$ and $\cos(\omega t)$, then using simple trigonometric product-to-sum identity it is possible to write:

$$A \sin(\omega t + \varphi) \cdot \sin(\omega t) = \frac{A}{2} \cos(\varphi) - \frac{A}{2} \cos(2\omega t + \varphi) \quad (1)$$

And:

$$A \sin(\omega t + \varphi) \cdot \cos(\omega t) = \frac{A}{2} \sin(\varphi) + \frac{A}{2} \sin(2\omega t + \varphi) \quad (2)$$

From those equations it is clear that if double frequency component and factor of two is disregarded, then first equation can be viewed as giving real part of the complex response, and second equation an imaginary part. When response on only few frequencies is needed, then such a multiplication is often preferred signal processing method both in analog and digital domains. Of course it should be complemented with low pass filtering in order to remove 2ω component. If other than sinusoidal signals are used for excitation and reference, then more frequency components should be considered when calculating response. As long as the reference signal is kept sinusoidal result of the multiplication is still faithful representation of the complex response of the system at the frequency of the reference signal. Some scaling is needed though, due to the fact that magnitude of the fundamental frequency component of the non-sinusoidal signal is different from the magnitude of said signal itself. Unfortunately quite often the reference waveform is far from sinusoidal. Square wave is preferred simply because multiplication of two functions will be replaced by simple signed summing. In analog signal processing domain it will enable to use simple switches instead of sophisticated, error prone, and power hungry multipliers. Same is true in digital domain, where multiply accumulate operation can be replaced with simpler accumulation. Also in this case system response is correct if excitation signal is kept sinusoidal, and in addition to that the system is truly linear. Last part requires very careful analysis, since appearing higher harmonics might coincide with reference signal harmonics, and after multiplication became undistinguishable from the true response. Still, SINAD of the response will degrade even if system can be considered sufficiently linear. If nothing else, then noise will leak into the result, at frequencies which will coincide with higher harmonics of the reference signal. Therefore methods which enable reduction of higher harmonics in reference signal, while keeping it reasonable simple for signal processing, are of utmost importance.

3. Square waves for measurement at single frequency

First the simplest and spectrally worse signal is examined. It is square wave signal or signal which can be described as the sign of $\sin(\omega t)$:

Fig. 4. An odd square wave, with frequency $f = \omega / 2\pi = 1\text{Hz}$

Fourier series of this square wave function will contain only sinusoidal members, since function is odd, i.e. $-f(x) = f(-x)$:

$$f(t) = \frac{4}{\pi} \sum_{n=1}^{\infty} \frac{\sin((2n-1)\omega t)}{2n-1} = \frac{4}{\pi} (\sin(\omega t) + \frac{1}{3}\sin(3\omega t) + \frac{1}{5}\sin(5\omega t) + \dots) \quad (3)$$

It can be viewed as signal with fundamental frequency ω having higher harmonics on odd multiplies of ω . If multiplication of the system response signal with orthogonal references is chosen as signal processing means, and these reference signals are not sinusoids, then all the members of the Fourier series of the response signal are multiplied with all the members of the Fourier series of the reference signals, i.e. not only fundamental components of the signal, but also all higher harmonics. Therefore if those members of the Fourier series coincide in frequency, then the results of multiplication will be added together and become undistinguishable. The measurement is no longer conducted on single well defined frequency, but instead produces results also on all higher harmonics. As it was discussed above it could be largely ignored, if during signal processing multiplication is conducted with sinusoidal signals, unfortunately it is often accomplished with the same rectangular signal instead, and energy from higher harmonics is summed together. Also spectral impact due to non-linearity of the object (or apparatus) cannot be separated from desired response signal anymore. Worst case impact of the coinciding spectral components summed eventually all together can be seen on Fig. 5 with dotted line.

There is another way of looking at how the errors appear (Kuhlberg, Land, Min, & Parve, 2003), by considering phase sensitivity characteristics of the synchronous demodulator (SD). They are easy to draw by varying phase shift φ between two signals. For simple square waves such a characteristic is presented on Fig. 6. For comparison it is drawn together with ideal circle which appears when two sinusoids are multiplied.

Fig. 5. Worst case relative impact of the higher harmonics to the multiplication result compared to the level of first harmonic in dB. Case of ordinary square wave (dotted line), simple shortened square wave (white boxes), and multilevel shortened square wave (black boxes) (Annus, Min, & Ojarand, 2008)

Fig. 6. Quality of synchronous demodulation in case of square wave signals (bold line). For clarity only one quarter is shown

From the Fig. 6 the magnitude error can be easily computed, as the difference between two lines, and Fig. 7. shows this relative magnitude error when square waves are used instead of sinusoids.

Relative magnitude errors of such magnitude as shown in Fig. 7 are generally unacceptable. Not to mention large (several degrees) phase errors in addition to the magnitude error. Only in very specific cases is measurement with pure square waves useful, such as measurement of electrical bioimpedance in implanted pacemakers, where energy constraints are severe.

Fortunately there is very simple method for reducing errors introduced by higher harmonics. Let's consider sum of two square waves with same frequency and amplitude, one of them shifted in phase by β degrees, and another $-\beta$ degrees. Such a double shift is preferable, since resulting function is again odd. In signal processing odd functions are more

Fig. 7. Relative magnitude error when square wave signals are used during synchronous demodulation

natural, because negative time is usually meaningless, and signals start at $t=0$. Care must be taken that in many mathematical textbooks, and more importantly in different programs, even functions are considered instead. Should the summary phase shift 2β be equal to the half period or odd multiply of half periods of any of the higher harmonic, then such a harmonic will be eliminated from the signal, since sum of two equal sinusoids with 180 degree shift is zero. Such an operation is essentially comb filtering. Main difference of the resulting signal compared with simple square wave is in appearing third level with zero value, so it is reasonable to call them shortened square waves. More generally spectrum of these signals can be derived from Fourier series:

$$f(t)=\frac{4}{\pi}\sum_{n=1}^{\infty}\frac{\cos((2n-1)\beta)\sin((2n-1)\omega t)}{2n-1}=\frac{4}{\pi}(\cos\beta\sin(\omega t)+\frac{\cos3\beta}{3}\sin(3\omega t)+\dots) \tag{4}$$

Two of these shortened square waves are of special interest. In order to remove 3rd and 5th harmonics from the signal (as they cause most significant errors) 18 degree and 30 degree shifts are useful. First of them is void of 5th, 15th, etc. harmonics, and second 3rd, 9th, 15th etc. harmonics. It means that impact of these harmonics is drastically reduced. Both of these three level signals with amplitude A are shown on Fig. 8. The third level does not introduce

Fig. 8. 18, and 30 degree shortened signals with amplitude A (Min, Kink, Land, & Parve, 2006)

much added complexity from signal generation or processing point of view. Both generation with digital logic, and also synchronous rectification with CMOS switches is straightforward (Min, Kink, Land, & Parve, 2006). If one of them is used as excitation signal and other as rectifying reference, then the result will be much cleaner spectrally when compared to simple square waves, Fig. 5 white rectangles. These two waveforms were chosen, because complete elimination of certain harmonics was desired. Due to that quality of synchronous demodulation is drastically improved, as it can be seen on Fig. 9 and 10. Compared to Fig. 7 improvement is considerable. Not substantial, but still drawback is the need to generate different signals for excitation and synchronous demodulation. What would happen if both signals are shortened by the same amount, and is there an optimum?

Fig. 9. Quality of synchronous demodulation in case of shortened square wave signals bold line, compared with sinusoidal signals dashed line

Fig. 10. Relative magnitude error when 18 and 30 degrees shortened square wave signals are used during system identification

If complete elimination of some of the harmonics is not pursued, then arbitrary shift between component square waves is allowed. One possible optimization approach would be to find shortened square wave where minimal energy is leaked into higher harmonics. Relative dependence of the energy of the higher harmonics from shortening angle can be seen on Fig. 11:

Fig. 11. Relative dependence of the energy of the higher harmonics from shortening angle

Fig. 12. Quality of synchronous demodulation in case of shortened square wave signals when both are shortened by 22,5 degree bold line, compared with sinusoidal signals dashed line

Fig. 13. Relative magnitude error when 22,5 degrees shortened square wave signals are used during system identification

Such a signal could be used for excitation as well as for synchronous demodulator. Since curve is relatively flat around minimum then shortening angles between 22 and 24 degrees produce almost equally good results. Obvious choice for real system identification task would be 22,5 degrees. Compared with previously discussed pair of shortened square waves generation of such a signal requires much lower clock frequency. When 18 and 30 degree shortening is required, then the clock frequency must be at least 30 times higher than the frequency of the resulting signal. With 22,5 degree shortened signal only 8 times higher clock is required. It allows either better energy efficiency by lowering the system clock, or alternatively allows usage of higher frequency signals at the same clock rate. An added benefit is, that noise leakage into result is lower, then in case of 18 and 30 degree shortened pair.

While performing considerably better than simple square waves these shortened signals are far from ideal sinusoid. Could the same summing procedure produce further improvement without much added complexity, if more square waves are added together? The answer is yes. It is enough to add third member into palette consisting already from two square waves shortened by 18 and 30 degrees, namely square wave shortened by 42 degrees. By combining these three signals promising results can be achieved. Three interesting and still simple signals are considered as combinations of previously mentioned summed signals. First and perhaps most obvious is a sum of 18, 30, and 42 degrees shortened square wave signals with signs 1, -1, and 1. Resulting waveform is on Fig. 14, and spectrum of this waveform is on Fig. 15 (Annus, Min, & Ojarand, 2008). It is much cleaner compared to ordinary square wave.

Fig. 14. Resulting waveform from summing of three shortened signals with weights 1, -1, and 1 (Annus, Min, & Ojarand, 2008)

If on the other hand excitation is also shortened square wave, then following pair of signals is suggested (Annus, Min, & Ojarand, 2008). First of them is sum of all three components with coefficients 1,1, and 1, Fig. 16. Spectrum of this summed signal is on Fig. 17.

Fig. 15. Spectrum of the signal on Fig. 14 (Annus, Min, & Ojarand, 2008)

Fig. 16. Sum of three shortened waveforms with coefficients 1, 1, and 1 (Annus, Min, & Ojarand, 2008)

Fig. 17. Spectra of the signal on Fig. 16 (Annus, Min, & Ojarand, 2008)

Fig. 18. Sum of three shortened waveforms with coefficients 2, -1, and 1 (Annus, Min, & Ojarand, 2008)

Suitable counterpart summed with coefficients 2, -1, and 1 is on Fig. 18, and spectrum on Fig. 19:

Fig. 19. Spectrum of the signal on Fig. 18 (Annus, Min, & Ojarand, 2008)

Comparison of the worst case multiplication results (Fig. 5) shows significant improvement over previous result. Same improvement can be seen on Fig. 20 as well, where result of synchronous demodulation is shown. Relative magnitude error compared to sinusoid is given on Fig. 21. Nevertheless same clock speed penalty still applies as with simpler solution.

Fig. 20. Quality of synchronous demodulation in case of described multilevel shortened square wave signals

Different way of cleaning square wave spectrally it is described in (Min, Parve, & Ronk, Design Concepts of Instruments for Vector Parameter Identification, 1992). Simple piecewise (over number of system clock periods) constant approximation of the sine wave values is

Fig. 21. Dependence of the relative magnitude error from phase difference

used. Waveforms with relatively small number of different levels (3,4,5) are used, and as with already described shortened square wave method different waveforms are suggested for multiplication, resulting in cleaner multiplication product. Values of separate discrete levels are determined according to:

$$a_q = \sin\left(\frac{\pi}{4m}(2q-1)\right) \quad (5)$$

Where m is the total number of approximation levels, and $q = 0, 1, 2, \dots, m$ is the approximation level number. Spectral composition of these approximated harmonic functions can be found according to the following equation:

$$k_h = 4mi \pm 1 \quad (6)$$

Where k_h is the number of the higher harmonic, which exist in the spectra, and $i=1,2,3,\dots$. If two such signals with number of levels m_1 and m_2 are multiplied, then coinciding harmonics can be found according to:

$$k_c = 4m_1m_2j \pm 1 \quad (7)$$

Where $j=1,2,3,\dots$. If two waveforms with $m=3$, and $m=4$ are considered, then first coinciding harmonics are 47th, 49th, 95th, 97th, 143rd, 145th, etc. As with shortened square waves clock frequency should be relatively high, and furthermore these waveforms are relatively sensitive to level errors, which prohibit usage of higher m values, and manifest itself in reappearing higher harmonics.

4. Square waves for multifrequency measurement

If system parameters vary with frequency, as they usually do, single frequency measurement is not enough to fully describe object under test. In complex cases, like measuring electrical properties of biological specimens, sweeping over wide frequency band may be warranted. Sweeping on the other hand is slow, and prohibits examination of faster changes in object under investigation, since transfer function of dynamic system is time dependent. Measurements must be as short as possible to avoid significant changes during

the analysis and, at the same time, as long as possible for enlarging the excitation energy and improving signal to noise ratio.

Chirp signals can be considered to remedy those shortcomings. They will be mentioned here only briefly, since more comprehensive overview is in different chapter. Chirp signals, i.e. multi-cycle sine wave based signals in which the frequency increases ('up-chirp') or decreases ('down-chirp') continuously as a function of time, are widely used in radar and sonar applications, acoustic, ultrasonic, optical and seismological studies. The main advantage of chirp signals is their well-defined frequency range and predetermined power spectral density and good crest factor. Rectangular chirps can be used to further simplify signal generation and processing. Moreover, the rectangular waveform has the minimal possible value crest factor (ratio of a peak value to a root-mean-square level) of 1. Two signals are viewed here briefly: binary chirp and ternary chirp (Min et al., 2012). Binary or signum chirp is defined as signum of the sinusoidal counterpart. For linear chirps:

$$\text{sign}(ch(t)) = \text{sign}\left(\sin\left[2\pi \frac{B}{T} \frac{t^2}{2}\right]\right) \quad (8)$$

Where T is duration of the chirp signal, $0 \leq t \leq T$, and B is bandwidth of the signal. It has crest factor of 1, and it means that energetically the signum chirp is two times more powerful than sinusoidal chirp. Waveform of the binary chirp can be seen on Fig. 22. Third level can be introduced by comparing sinusoidal chirp with two levels instead of one as it is done in case of binary chirp (Fig. 23).

Fig. 22. NRZ or non-return-to-zero binary rectangular chirp pulse (Min et al., 2012)

Fig. 23. Return to zero (RZ) ternary rectangular chirp pulse (Min et al., 2012)

Instead of chirp signals other waveforms can be considered as well. A widely used method is to generate a pseudo-random maximum length sequence (MLS). Spectrum of the MLS signal follows a $\text{square}(\sin(x)/x)$ law. Signal processing is usually accomplished by taking circular cross-correlation of the output signal with the excitation MLS. MLS and chirp have however one serious disadvantage – their energy is distributed equally, or almost equally, over the whole frequency band of interest. Therefore, the power spectral density A^2/Hz is comparatively low.

In practice, there is seldom a need to measure at all of the frequencies within the bandwidth simultaneously, except perhaps when system under test is highly resonant. Usually it is enough to know the parameters at several arbitrarily spaced frequencies separately. Therefore, it is reasonable to concentrate the energy of the excitation signals to frequencies of interest instead of using uniform energy distribution over full measurement bandwidth. That can be achieved by summing up several sinusoids:

$$x(t) = \sum_{i=1}^{i=n} A_i \cdot \sin(2\pi f_i + \varphi_i) \quad (9)$$

Unfortunately while single sinusoid is technically feasible signal, and can be reproduced quite accurately, it becomes increasingly costly to use simultaneously many sinusoidal signals. There is another drawback associated with simultaneous use of multiple sinusoidal signals – crest factor. With single sinusoid the crest factor is $\sqrt{2} \approx 1,414$. By summing two or more sinusoidal signals together the crest factor can take many different values, generally bigger than 1,414. Why is crest factor so important? Two reasons are worth considering: nonlinear behavior of the object under investigation, and dynamic range of the measurement apparatus itself. Real objects can rarely be described as linear. It means that different excitation levels do not produce linearly dependent responses. For practical purposes measurement signals are usually kept within narrow range of amplitudes where object behaves approximately linearly. It is clear that such an approximation is better the narrower the range is kept. In worst case high energy pulses can even permanently alter or destroy the object under investigation, and that is certainly not acceptable when performing measurements for example on living human tissue. Also dynamic diapason of the apparatus is limited. From the lower side the limit is set by omnipresent noise signal. If the measurement signal is completely buried in the noise and cannot be restored any more the

Fig. 24. Multisinusoidal excitation signal (a), containing eleven equal amplitude components, with peak value of 1. For simplicity starting phases of all components are zero. Frequencies in Hz are: 1, 2, 4, 8, 16, 32, 64, 128, 256, 512, 1024. Same signal with some optimization of phases (b); achieved crest factor is 2,5. Please note that the peak values of these two signals are equal

measurement is void. Upper limit is ultimately determined by supply voltage. Form that it is clear that large peaks in excitation signal should be avoided, as well as very low level components which get lost in noise. Fig. 24 is to give an impression of what happens when just eleven sinusoidal signals, one octave apart from each other in frequency, each with equal amplitude, and with random initial phase are summed together. The crest factor of this noise like signal is 3,54, and it is clearly worse than crest factor of a single sinusoid.

Interestingly almost all the parameters of the multisinusoidal signal can be improved considerably by simplifying it drastically (Annus et al., 2011). Such a signal is derived from the original multisinusoidal signal by detecting its zero crossings:

$$\text{sign}(x(t)) = \text{sign} \sum_{i=1}^{i=n} A_i \cdot \sin(2\pi f_i + \varphi_i) \quad (10)$$

In case of $n=11$ the resulting waveform is shown on Fig. 25.

Fig. 25. Typical binary multifrequency signal with eleven components and random phases

Clearly multifrequency binary signal is easier to generate then multisinusoidal signal, and has far superior crest factor of 1. Minor drawback can be seen on Fig. 26, where spectrum of such a signal is shown. So called “snow” lines appear between the wanted frequency components, and roughly 30% of the total energy is lost for measurement:

Fig. 26. Spectrum of binary multifrequency signal with ten components and random phases

Truth is revealed when spectrum of optimized multisine is drawn together with spectrum of the binary counterpart:

Fig. 27. Magnitudes of spectral lines in binary multifrequency signal (grey squares) together with spectrum of multisinusoidal signal (black squares)

Useful spectral components in binary multifrequency signal contain 1,34927 times more energy than in case of optimized multisine, or 4,32933 times more than the multisine above without optimization. At first glance there is another drawback, since magnitudes of useful components are not exactly equal anymore. Fortunately said magnitudes can be easily equalized by iteratively manipulating magnitudes of the components of the original multisinusoidal signal. Residual error is generally well below one percent. Furthermore almost arbitrary magnitudes of spectral components can be achieved (Fig. 28).

Fig. 28. Magnitude control example

When two level comparison is introduced instead, then energy leakage into unwanted components can be reduced. Such a ternary multifrequency signal can be seen on Fig. 29.

Fig. 29. Ternary multifrequency signal. Comparison levels are set on +/- 0,23V

Comparison of the spectrum of the original multisinusoid with ternary signal from Fig. 29 can be seen on Fig. 30.

Fig. 30. Magnitude spectrum of the ternary multifrequency signal (grey rings) compared with magnitudes of multisinusoid components (grey lines), and binary multifrequency signal (crosses)

5. Conclusion

With some minor drawbacks it is possible to construct relatively simple square wave signals in order to replace more sophisticated sinusoidal or arbitrary waveforms when system identification is warranted. Simplest square wave nevertheless might not be sufficiently good measurement signal. By adding few more levels situation can be improved considerably. These signals can be used to replace single sinusoids, chirps, and arbitrary sums of sinusoids. Generally it is a good idea to choose best signal for given identification task. Some choices are shown, and reasoning behind them given.

6. Acknowledgment

This research was supported by the European Union through the European Regional Development Fund and through the projects of Eliko Competence Centre, and Centre for Research Excellence CEBE.

7. References

- Heideman, M.; Johnson, D.; Burrus, C. (1984). Gauss and the history of the fast fourier transform, *ASSP Magazine, IEEE*, vol.1, no.4, pp.14-21
- Maxime Bocher (1906). Introduction to the Theory of Fourier's Series. *The Annals of Mathematics, Second Series, Vol. 7, No. 3*, pp. 81-152
- Keith Godfrey (1993). *Perturbation Signals for System Identification* (Prentice Hall International Series in Acoustics, Speech, and Signal Processing)
- Rik Pintelon, Jhan Schoukens (2001). *System Identification: A Frequency Domain Approach*. Wiley-IEEE Press; 1 edition
- Annus, Paul; Min, Mart; Ojarand, Jaan (2008). Shortened square wave waveforms in synchronous signal processing. *IEEE International Instrumentation and Measurement Technology Conference Proceedings: 2008 IEEE International Instrumentation and Measurement Technology Conference (I2MTC 2008)*, 12-15 May 2008, Victoria, British Columbia, Canada, 2008. , 1259 - 1262.
- Min, M.; Parve, T.; & Ronk, A. (1992). Design Concepts of Instruments for Vector Parameter Identification. *IEEE Transactions on Instrumentation and Measurement*, Volume: 41 Issue:1, pp. 50 – 53

- Min, Mart; Parve, Toomas; Land, Raul; Annus, Paul (2012). Scalable Impedance Spectroscopy: comparative study of sinusoidal and rectangular chirp excitations. In: Lecture Notes on Impedance Spectroscopy v. 2 Measurement, Modeling and Applications – Olfa Kanoun (ed), Taylor & Francis Group, London, 33-38.
- Annus, P.; Min, M.; Ojarand, J; Paavle, T.; Land, R.; Ellervee, P.; Parve, T. (2011). Multisine and Binary Multifrequency Waveforms in Impedance Spectrum Measurement – A Comparative Study. 5th European IFMBE MBEC 2011, Budapest, 14-18 Sept., 2011, 1265 - 1268.

Fourier Transform - Signal Processing

Edited by Dr Salih Salih

ISBN 978-953-51-0453-7

Hard cover, 354 pages

Publisher InTech

Published online 11, April, 2012

Published in print edition April, 2012

The field of signal processing has seen explosive growth during the past decades; almost all textbooks on signal processing have a section devoted to the Fourier transform theory. For this reason, this book focuses on the Fourier transform applications in signal processing techniques. The book chapters are related to DFT, FFT, OFDM, estimation techniques and the image processing techniques. It is hoped that this book will provide the background, references and the incentive to encourage further research and results in this area as well as provide tools for practical applications. It provides an applications-oriented to signal processing written primarily for electrical engineers, communication engineers, signal processing engineers, mathematicians and graduate students will also find it useful as a reference for their research activities.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Paul Annus, Raul Land, Mart Min and Jaan Ojarand (2012). Simple Signals for System Identification, Fourier Transform - Signal Processing, Dr Salih Salih (Ed.), ISBN: 978-953-51-0453-7, InTech, Available from: <http://www.intechopen.com/books/fourier-transform-signal-processing/system-identification-simplified>

INTech
open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia
Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元
Phone: +86-21-62489820
Fax: +86-21-62489821

© 2012 The Author(s). Licensee IntechOpen. This is an open access article distributed under the terms of the [Creative Commons Attribution 3.0 License](https://creativecommons.org/licenses/by/3.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

IntechOpen

IntechOpen