We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,300

Open access books available

172,000

International authors and editors

190M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities

WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com

Practical Considerations of Liquid Handling Devices in Drug Discovery

Sergio C. Chai, Asli N. Goktug, Jimmy Cui,
Jonathan Low and Taosheng Chen

Additional information is available at the end of the chapter

<http://dx.doi.org/10.5772/52546>

1. Introduction

Automated liquid handling has become an indispensable tool in drug discovery, particularly in screening campaigns ranging millions of compounds. Intense innovations of these devices go hand in hand with the progression towards assay miniaturization, accelerating dramatically the discovery of drug candidates and chemical probes for querying biological systems. The advancement in this technology is driven in large part by much impetus in cost reduction and efficiency. In addition to increased throughput, streamlining screening operations using automated fluid devices ensures consistency and reliability while avoiding human error.

In this chapter, we provide a general overview of existing liquid handlers, with emphasis on their strengths and limitations. Notably, we discuss practical considerations in the implementation of these devices, methods to discern performance quality and potential sources of error.

2. Types of liquid handling devices

A whole array of liquid handlers has been developed for every aspect of drug discovery. These instruments encompass different technologies for distinct purposes. In terms of application, they are broadly classified as bulk liquid dispensers, transfer devices and plate washers (Rudnicki and Johnston 2009).

Based on the way the reagent is being transferred, these instruments can follow two dispensing modes: contact or non-contact (Kong et al. 2012). Contact-based devices allow the fluid to be transferred to touch the surface of the destination container or solution, offering a simple and dependable alternative to sub-microliter fluid handling. Non-contact devices utilize additional force other than gravity to eject liquids, as minute volumes cannot be dispensed efficiently with gravity alone (Kong et al. 2012). The process is faster than using permanent tips or pins (Fig.1), because there is no washing step between delivery, while reducing cross-contamination and evaporation (Dunn and Feygin 2000).

Figure 1. Various types of liquid handling tips, pins and heads from A) washer B) pintool C) peristaltic pump-based bulk dispenser D) liquid handler with single and 8-channel pipettors E) pipettor with 8-independent channels.

2.1. Peristaltic-based devices

The peristaltic pump is used for bulk reagent dispensing in conjunction with a nozzle head (Fig.1C) and a flexible tubing cartridge. The tubings stretch around a set of rollers connected to a motor. With the rotating motion of the motor, the rollers compress the tubings creating a continuous fluid motion due to positive displacement.

Typically, this type of dispenser is capable of handling volumes as low as 5 μL , offering a fast dispensing option for 96-/384-/1536-well plate formats. The disposable tubing cartridge is pre-sterilized, and the entire liquid path can be autoclaved. Additionally, these devices are normally equipped with programming capabilities that allow discrete column-wise dispensing, variable rolling speed settings and adjustable dispensing volume. The pump can roll both forward and backwards to execute priming and emptying functions, respectively. A major limitation is the lack of capabilities to dispense into individual wells.

2.2. Fixed-tip transfer devices

Fluid handlers that utilize fixed-tips (Fig.1E) are usually efficient at transferring relatively small volumes (100 μL or above) and have been largely used for compound pipetting (“cherry picking”) and serial dilutions. They incorporate 2-/ 4-/ 8-channel expandable liq-

liquid handling arms in addition to 96- and 384-channel heads. This type of liquid handling device functions based on air displacement mechanism. The dilutor or syringe plunger pulls system liquid from the pipette tubing to aspirate the sample, with an air gap separating both fluids. The plunger speed, syringe size and resolution are factors that affect pipetting flow rate.

2.3. Changeable-tip transfer devices

The use of disposable tips (Fig.1D) is a simple alternative to avoid washing steps required for fixed-tip based systems, while eliminating completely the risk of cross-contamination. These instruments employ a conventional air displacement mechanism. A wide array of commercially-available tip sizes, materials and molding qualities offers the scientist great flexibility. There are even specialized tips with nanoliter-scale transfer capabilities that can be used in any conventional pipettor (Murthy et al. 2011; Ramírez et al. 2008).

2.4. Pintool transfer devices

Pintool is a contact-based dispensing method widely used for handling volumes at the nanoliter scale (Cleveland and Koutz 2005). It consists of a set of stainless steel pins (Fig. 1B) carefully crafted for consistent dimensions. The bottom end of the pins can be solid, grooved or slotted, with the option of having a hydrophobic coating to prevent non-specific binding (Dunn and Feygin 2000; Rudnicki and Johnston 2009). Solutions are transferred through a combination of capillary action and surface tension, with the volume being highly dependent on the contact surfaces and solution properties (Dunn and Feygin 2000). The pin array is normally assembled in a floating pin cassette to ensure soaking of all the pins amid uneven surfaces, which also minimizes pin damage. After liquid transfer, the pins have to cycle through washing steps to prevent cross-contamination.

2.5. Piezoelectric devices

The piezoelectric dispenser is a non-contact technology, where solutions are delivered as multiple tiny drops of defined size (Niles and Coassin 2005). This technology has been utilized in contemporary inject printers and refined to be implemented in the biological sciences. Various biochemical solutions (DNA, RNA, proteins) and bacterial suspensions have been tested with no negative effects (Schober et al. 1993). The system is composed of a capillary tube made of quartz or steel, with one end connected to the reagent reservoir and the other end ending in an orifice from which droplets are ejected (Niles and Coassin 2005). A piezoelectric crystal collar is bound to the capillary, which is filled with solution. Upon voltage application, the piezoelectric element contracts causing pressure on the capillary to generate fine drops. The ejection is at high acceleration with minimal wetting of the nozzle (Schober et al. 1993). Several thousand drops can be dispensed per second, with attainable drop sizes spanning the picoliter and nanoliter range (Schober et al. 1993). Droplet volume depends on several factors, including bore diameter, solution viscosity and the voltage pulse amplitude and frequency (James and Papen 1998; Kong et al. 2012).

2.6. Solenoid-based devices

Solenoid-based devices are non-contact dispensers that use a positive displacement mechanism (Bateman et al. 1999). The flow of pressurized liquid is occluded by a solenoid valve, which is actuated by electric current to allow for liquid to pass through the valve. The dispensed volume is regulated by the fluid pressure, duration of the valve in the open position, solution properties and orifice diameter (Bateman et al. 1999; Niles and Coassin 2005). Depending on the time the valve stays in the open position, the device can eject droplets or a continuous stream (Niles and Coassin 2005).

2.7. Acoustic devices

Acoustic droplet ejection (ADE) is a recent touch-less technology that surges in popularity in recent years. It adopts acoustic energy to propel droplets from various types of solutions with good precision (Ellson et al. 2003; Harris et al. 2008; Rudnicki and Johnston 2009; Shieh et al. 2006). The source plate remains stationary as the transducer and destination plate shuffle to allow for solution transfer from any well in the source plate to any well in the destination plate, the latter one lying in an inverted position (Olechno et al. 2006). This system does not require any additional consumable other than microplates (Olechno et al. 2006), and it speeds up the process by avoiding washing steps and having the capability to prepare assay-ready plates (Turmel et al. 2010).

2.8. Microplate washers

Microplate washers are laboratory instruments designed to automate and expedite assay applications, where a washing step is essential. They play an important role in areas such as high-content screening and enzyme-linked immunosorbent assays (ELISA). In 1990, Stobbs developed the first multiple plate washer using readily available materials as a low cost alternative to the commercially available plate washers of the era (Stobbs 1990). Over the years, fully programmable plate washers have been developed with numerous features. The development of automated plate washers has decreased the time required for laborious washing steps involved in many screening assays and improved reproducibility through standardized plate handling across multiple wash cycles (defined as a single dispense and aspirate step per cycle).

The two most critical components of a plate washer are a plate carrier and a manifold containing a number of fixed stainless steel needle probes for solution dispensing (Fig.1A). This manifold (or a separate manifold depending on the design) aspirates the liquid from the wells after an optional soaking period, leaving a pre-defined residual volume in the wells. A third component is the vacuum/pump assembly, which supplies the necessary pressure differential to drive efficient aspiration. Sunghou Lee first developed an additional vacuum filtration system integrated with a conventional plate washer to speed up the wash process for applications involving filter plates (Lee 2006). Some plate washers have a built-in magnet or a vacuum filtration module for handling bead-based assays.

Microplate washers can be categorized into two types: strip washers, which wash a single column or row of a plate at a time, and full plate washers (Rudnicki and Johnston 2009). The availability of 8-/12-/16-channel manifolds for strip washers provides both single strip washing and full-plate washing capability in the same device, but at the cost of increased wash time for full plates. On the other hand, full plate washers with either a 96- or 384-channel manifold may be preferred for time-efficient wash operations (from a few seconds to a few minutes), but lack the flexibility of the 8-/12-/16-channel units.

The combination of plate washing and bulk dispensing features within the same device may be favored for a space-efficient solution. They are designed to dispense reliably low volumes and reduce prime volume (Rudnicki and Johnston 2009). A major advantage of the washer-dispenser combination comes into play with assay protocols that require the direct addition of fluid after or between the washing steps, such as cell fixation or microplate surface coating reagents.

3. Considerations for using liquid handling devices

3.1. Determination of quality assessment descriptors

Assessment of instrument performance has become important in order to minimize false-positive and false-negative rates in high-throughput screening (Taylor et al. 2002). One of the most important figures of merit in evaluating the performance of liquid handlers is accuracy, which is commonly reported as %bias (Rose 1999):

$$\%bias = 100 \times \left(\frac{V_M - V_T}{V_T} \right) \quad (1)$$

where V_M is the measured volume and V_T is the theoretical volume (desired). %bias represents the deviation from the desired volume, with a value of 0% indicating no deviation from the true value.

The precision, a measure of reproducibility, is calculated from the mean and standard deviation (SD) of a set of measurements, and it is reported as percent coefficient of variation (%CV) or relative standard deviation (RSD), as shown in Eq. 2. For most cases, it is adequate to have a bias value below 5% and a CV below 10% (Rose 1999).

$$\%CV = 100 \times \frac{SD}{\text{mean}} \quad (2)$$

There have been several approaches for volume verification, which typically consist of gravimetric or photometric methods. Gravimetric measurements utilize the mass and the density (ρ) of the dispensed solution to determine the volume. It has been used extensively to calibrate and verify the accuracy of liquid dispensers (Bergsdorf et al. 2006; Rhode et al. 2004; Taylor et al. 2002). Typically, the solution is dispensed across a pre-weighed microtiter

plate, which is weighed immediately after dispensing to prevent evaporation. %bias can be calculated based on the total weight of the dispensed solution (W_{total}) and the number of dispensed wells (n):

$$\% \text{bias per well (gravimetric)} = 100 \times \frac{\left(\frac{W_{\text{total}}}{n \times q} \right) - V_T}{V_T} \quad (3)$$

Environmental conditions (e.g. temperature and humidity) have major effects on the reliability of gravimetric methods, which facilitates evaporation and water uptake for hygroscopic solvents such as dimethyl sulfoxide (DMSO). These factors of variation can be minimized by placing gasketed lids on the microtiter plates immediately following dispense (Taylor et al. 2002).

Absorbance and fluorescence are the most common photometric methods utilized to test the accuracy and precision of the transferred volumes of a liquid handling device. In a study comparing the performance of the two methods on determining the precision in 96-/384-/1536-well plates, no significant difference was observed between the 96- and 384-well plates (Petersen and Nguyen 2005). However, to achieve similar results for both fluorescence and absorbance measurements in the 1536-well plate, a centrifugation step was required because of the irregular meniscus shape enhanced by the small well geometry. In another study performed on liquid handlers with two different mechanisms, absorbance was found to be a more reliable method as long as the pH stability of the dye-buffer solution is maintained (Rhode et al. 2004).

Fluorescence signal is also known to be susceptible to photobleaching, which can be prevented by shorter excitation times, suitable buffer solutions and adequate concentration of fluorophore (Diaspro et al. 2006; Harris and Mutz 2006). To overcome the problems encountered due to signal quenching in DMSO, sulforhodamine 101 was presented as an alternative fluorescence dye (Walling 2011). Fluorescein was found to be a suitable probe to use in liquid handling performance quantification as long as the DMSO concentration in the buffer solution does not exceed 1% and the stock solutions are stored in 70-100% DMSO in a dark environment (Harris and Mutz 2006). While photobleaching is not an issue in absorbance, the method is limited by high background levels and lower sensitivity compared to fluorescence (Bradshaw et al. 2007). Based on the physical characteristics of a transferred sample and the material of the consumables, unforeseen interactions may be observed influencing the assay results. Especially, DMSO-containing samples are highly affected by the hydroscopic properties of the solvent, which inflates sample volume (Berg et al. 2001).

3.2. Considerations for using bulk reagent dispensers: Peristaltic-based devices

A single screening experiment can be costly, requiring valuable compounds and biological reagents. Routine evaluation of liquid handlers, in particularly prior to each run, is a necessary mean for preventing disastrous outcomes. Simple procedures can be integrated to identify problems in a relatively short period of time, which in many instances, can be easily corrected. Routine analysis should be performed with the actual reagents, because there are

several factors that affect the dispensed volumes, including viscosity, density, and temperature (McGown and Hafeman 1998). General considerations to prevent undesirable dispensing performance and common sources of variations include:

3.2.1. Uneven dispensing

Tubings tend to stretch after certain period of use, affecting the intended volume to be delivered. When not in use, the cartridges should be placed in the “rest” position. In addition, autoclaving the cassettes should be minimized. Dispensing speed and the height of the tips in relation to the plate have to be optimized for the intended reagent, as viscous solutions could miss the targeted well at low dispensing speed and large spacing between the tips and microtiter plate. When working with cells, uneven dispensing can be reduced by increasing the prime volume, constant mixing/stirring the cell suspension source and minimizing cell clumps. Solutions should be dispensed in the center of the well, and plates have to be centrifuged when dispensing low volumes to force droplets at the walls to the bottom of the well. Cassettes should be calibrated regularly as recommended by the supplier and checked for tip clogging.

3.2.2. Protein binding

Protein binding to dispensing components is an important point to consider in the implementation of biochemical assays, particularly at low protein concentrations. In some instances, enzymes appear to be inactivated over time when dispensing multiple plates using a liquid handler, when in reality the enzymes have been depleted from the solution due to non-specific binding to plastic, silicone and other polymer-based surfaces. This effect is amplified when dispensing sizeable number of plates, as there is larger exposure time of the assay components to the surfaces of reagent reservoirs and dispensing cassette elements. In order to circumvent this problem, blocking reagents can be added to the buffer, plastic surfaces can be coated, or a combination of both. The two major types of blocking reagents are detergents and proteins. It is preferable to use non-ionic detergents such as Tween-20, Triton X-100 or Nonidet-P40. Among the most widely-used protein blockers are bovine serum albumin (BSA) and casein. Protein blockers are better suited for coating surfaces, as detergents can be easily washed away. Typical working concentrations for detergents range from 0.01 to 0.1%, while protein blockers are used between 0.1 to 3 %. The selection of the appropriate type of blocking reagent and concentration is central to a robust assay. Other less common blocking reagents include polyethylene glycol (PEG), polyvinyl alcohol (PVA) and polyvinyl pyrrolidone (PVP). Additionally, the use of glass reagent reservoirs is recommended.

3.2.3. Clogging

Particles can obstruct the flow of a dispensing cassette mainly by blocking the tips. Complete clogging is fairly easy to recognize, as the lack of fluid coming out of the tips can be visibly noticed. Depending on the degree of obstruction, partial clogging may not be easily perceived by the naked eye, and it is detected only by photometric or gravimetric testing. However, there are certain indications of partial clogging, such as slanted fluid spray or

drop formation at the tip. To prevent clogging, the tubing should be primed with deionized water shortly after use, especially prior to priming with alcohol, as salts in the buffer may precipitate and biological reagents may clump. When working with cells, it is recommended to wet the tubing with buffer or media before dispensing cells, and if possible, not to allow the cells to settle in the tubing by emptying the contents back to the reservoir immediately after dispensing (prime/empty cycle).

3.2.4. *Foaming*

Solutions with high protein content can cause frothing, including media containing serum and biochemical buffers with high percentage of BSA (used as blocking protein). To minimize frothing, it is recommended not to empty the tubing between dispensing (as ordinarily performed in fully automated platforms for large screenings). If tubing emptying is unavoidable, it is advisable to empty a volume smaller than the dead volume. Other means to reduce frothing involve decreasing dispensing speed and applying grease to the cassette tips. Torn or cracked tubing can pull air generating bubbles.

3.2.5. *Reservoir container*

The reservoir container is an important component of a liquid dispenser that is often neglected in troubleshooting. The material of the container can have a detrimental effect on the assay robustness, such as sticking of proteins to plastic surfaces. For peristaltic pump-based dispensers, we suggest using a jacketed glass flask connected to a water chiller (waterbath with adjustable temperature). Careful monitoring of the temperature in the flask using a thermometer is recommended, as the temperature set in the chiller is not always reflected in the container. Suspensions of cells, beads or nanoparticles have to be constantly stirred to prevent settling, which could result in uneven dispensing or clogging. The stirring speed needs to be optimized, as fast stirring can create bubbles and disturb biological components (cells). When working with large reagent volumes at the start of dispensing, the stirring may have to be reduced as the volume decreases to prevent foaming or bubble formation.

3.2.6. *Tubing extension*

Extensions can be implemented when the dispensing tubings cannot be immersed in the reservoir container because of its large dimensions. Some commercially available extensions allow for the 8 tubings of a standard cartridge to be coupled into single elongated tubing through metallic cannulas sticking out of a joint casing. For viscous solutions, these types of elongations can introduce bubbles due to the joint design, particularly during prime/empty cycles. The metallic cannulas can easily tear the tubing during fitting, which is ameliorated by using glycerol or alcohol to smoothen the surfaces. A better alternative is to build home-made extensions by attaching each of the new tubings to separate discarded tubings through connectors, which can be made by cutting the end of a pipette tip.

3.2.7. Routine quality assessment

During assay development and validation, factors affecting liquid dispenser performance are identified and corrected. However, setbacks can occur randomly regardless of detailed preparations ahead of the screens. For instance, torn tubing, tip blockage or incorrect cartridge setup cannot be prevented a priori. Therefore, it is recommended to rapidly monitor dispensing variations at the start of a screen, where problems encountered at this stage can be usually corrected fairly quickly.

We normally dispense a solution of fluorescein isothiocyanate (FITC) in PBS into a couple of 384-well plates. Fluorescence intensities are analyzed for signal variations corresponding to each cassette channel, as described by %CV and %bias' (Fig.2). Determination of %CV for the entire plate is frequently performed in many laboratories, but this approach cannot distinguish issues with individual channels. In addition, a flawed channel does not necessarily change drastically the %CV of the whole plate, as illustrated by Fig. 2A. The types of problems commonly associated to high %CV include improper cassette mounting, tubing stretching and damage.

There are instances when the tip is partially obstructed, leading to reduced volume delivered. Even when a channel displays low fluorescence counts, the signal can still have small %CV values (Fig. 2B). We have adapted the concept of %bias to detect significant deviations in signal intensity for each row (S_R) compared to that of the whole plate (S_T), resulting in %bias' (Eq. 4). Values lower than 10 %CV and 10 %bias' are acceptable.

$$\% \text{ bias ' } = 100 \times \left(\frac{S_R - S_T}{S_T} \right) \quad (4)$$

3.3. Considerations for using transfer devices: Pintool

The pintool has become a mature technology for transferring nanoliter to sub-microliter volumes. Even though the system is regarded as fairly simple and robust, there are a number of points to consider for a consistent and reliable performance:

3.3.1. Volume variation

The volume delivered by a pin can change due to a number of factors. To minimize volume variations, there should be consistency in immersion depth (Dunn and Feygin 2000). There is a minimum volume required in the source plate, and the destination plate should not be dry (Rudnicki and Johnston 2009). The dwell time that pins spend in the fluid and withdrawal speed from the liquid surface should be optimized for solutions of very different properties (e.g. viscosity).

The slot of a pin can be tainted by compound precipitation or formation of suspension deposits (Fig. 3B). Sufficient and robust washing and drying steps are effective in preventing deposition and being critical to avoid carry-over and cross-contamination. The pins can be physically damaged by dipping in highly uneven surfaces, particularly when using slotted

pins (Fig. 3C). Coated pins should avoid harsh washing procedures, such as going through powerful sonication washes.

Figure 2. A-B) Delivery variation by a bulk reagent dispenser distributing a FITC solution into 384-well plates. Certain dispensing cassette channels display either higher %CV or %bias' values than the anticipated cut-off of 10%. C) Cell settling in the reagent reservoir when transferring to a microtiter plate using an automated pipetting system with an 8-channel head, with 1 min delay between transfers to each column. Cell settling is uneven due to the v-shaped bottom of the reservoir, causing the intensity pattern observed in the plate. The cells (HEK293T) were incubated with Cell-Titer-Glo® for 20 min prior to luminescence reading.

Figure 3. Magnified view of FP1NS50H pins (V&P Scientific, Inc.) with A) clean slot B) dirty slot C) damaged slot.

3.3.2. Carry-over

After transferring compounds from one plate to another, the pins are washed in DMSO, alcohol, water or a combination of these solutions. The pintool protocol involves dipping the pins in each solution bath certain number of times, at a particular speed and soaking time. The pins are then dried on lint-free blotting paper. Protocols of pintool devices used on robotic platforms are optimized for effectiveness in removing previous transfers while spending the minimum time between wash cycles. In many cases, the drugging (i.e., addition of compound to assay well) step using pintool becomes the bottleneck in a screening campaign, and the washing step accounts for most of the time consumed. However, certain assays can be very sensitive to compound carry-over, particularly if the compounds are very potent modulators and bind avidly to the pin surface. In such cases, increasing the number of dips and soaking time can improve cleanliness, albeit at the cost of increasing total transfer time.

Fig. 4 illustrates the effect of four different wash protocols in a kinase assay using staurosporine as the inhibitor. After compound transfer by pintool to the first assay plate, the pins are immersed in DMSO and isopropanol reservoirs, followed by drying on blotting paper. Subsequently, the pins are dipped in a second assay plate containing the kinase system. Residual staurosporine in the pins increases the signal variation as determined by %CV of a set of multiple wells. Protocol 1 has the least number of dips and soaking time per bath, resulting in the most dramatic signal variation due to carry-over. This general approach is recommended for detecting carry-over and selecting the appropriate pintool wash.

Figure 4. General approach to detect compound carry-over and optimize pintool washing. A single wash cycle consists of dipping the pins in DMSO and isopropanol baths, followed by blotting on lint-free paper.

3.3.3. Routine quality assessment

Regular pintool calibration and quality assessment can considerably improve data quality. In screening runs at a single compound concentration, well-maintained pins can lead to a

reduction of false negative hits, as damaged or dirty pins would usually deliver lower volumes than anticipated. In dose-response analysis, the quality of the curve fit is highly dependent on the variability of the data points.

A good quality control procedure should provide the transferred volume and the variation associated with the pin set. We implemented a relatively quick and simple procedure using a fluorescent dye (FITC). Prior to the test, the pins are washed as described above. A calibration curve is generated of fluorescence intensity as a function of FITC concentration. Using the pintool, FITC in DMSO is transferred from a source plate to several destination plates containing PBS (the use of 4 plates was shown to be sufficient). The average transferred volume per pin is calculated using the fluorescence signal of the destination plates and the calibration curve. Volume variation across the microtiter plate can be readily appreciated by plotting volume against well position (Fig. 5, top charts). The pink and green solid lines represent the upper and lower boundaries within 10% CV of the average volume, where outliers can be clearly identified. The frequency chart (Fig. 5, bottom chart) displays outliers present in 1, 2, 3 or all of the 4 destination plates, and it can be used to identify pins that consistently provide volumes outside a specified range. In the example shown in Fig. 5, pins corresponding to positions A13, B21, D8, F13, K1, N14 and P20 will have to be replaced. Depending on the need, stringency can be adjusted by changing the boundaries as specified by %CV. It is highly recommended to utilize the same freshly prepared fluorescent dye and buffer solutions in all aspects of the protocol. A template for data analysis can be easily created in conventional software such as MS-Excel.

Figure 5. A simple and comprehensive approach to analyze pintool performance. Individual pins can be selected for replacement based on consistent variation across multiple transfers.

3.4. Considerations for using transfer devices: Pipettors

3.4.1. Pipette stations

The automation station is an integral part of any high throughput pipettor, regardless of the type of tips (fixed or disposable) it employs. It typically consists of ANSI/SBS standard compliant single or multiple deck positions on a stationary or moving platform to hold the labware, with a moving arm situated above the platform containing the single- or multi-channel pipette head. A major advantage of automated pipettor devices over manual or electronic multichannel hand-held pipettes is the elimination of inconsistency in the transfer process by minimizing human intervention, which also enables high throughput applications that are not otherwise feasible. The three major tasks that can be performed with suitable hardware settings are liquid transfer, cherry-picking and serial-dilution.

For plate-to-plate liquid transfers, 96- or 384-well pipette heads are preferred to work with 96-/384-/1536-well microplates to speed up the process and increase the throughput. While 4-/8-/12-/16-pipette heads can also be used for direct transfer applications, they are primarily used to perform serial-dilutions. On the other hand, a single channel pipette tip is an essential component to accomplish cherry-picking tasks.

The speed of an automated pipettor is important for time-sensitive experiments. Especially when performing small volume transfers into microplates, the amount of time spent to transfer liquids in a column-by-column or row-by-row manner may be problematic due to quick evaporation. If the speed of transfer is too slow, some evaporation in the first column or row may be observed before dispensing to the last column or row, causing inconsistent volume across the plate. To avoid evaporation issues during liquid transfers, deck size, pipettor speed, head type and the transfer volume should be considered.

3.4.2. Tip contamination

Sample carry-over is a common problem in liquid handling tasks requiring sequential dipping steps into various sample reservoirs. With fixed-tips, an adequate cleaning step is essential between two transfer operations to prevent sample carry-over. An on-deck cleaning protocol often consists of immersion in a bath (DMSO, alcohol and/or water) with optional sonication step. The tips should be allowed sufficient drying time to prevent sample dilution in the following transfer phase. Appropriate wash solutions should be selected and the optimum length of washing time should be determined during the assay development stage. Although fixed-tips may have the risk of carry-over, they enable more accurate and precise transfers in smaller volume ranges (Felton 2003).

Contamination can also be associated with disposable tips, especially when sterile and nuclease-free assay conditions are required. The speed at which the pipette tips are removed from a sample fluid was found to correlate to the amount of macroscopic droplets stuck to the outer surface of the polypropylene tips, which contributed to cross-contamination (Berg et al. 2001). It was also reported that to decrease this form of cross-contamination, which is

influenced by the tip shape and the sample-polypropylene interactions, the removal speed should be slow enough to diminish droplet generation.

Impurities can also leach out of the disposable tips when in contact with solvents such as DMSO. Studies have shown that bioactive compounds released from plastic labware may interfere with assay readouts causing misleading experimental results (McDonald et al. 2008; Niles and Coassin 2008; Watson et al. 2009). Consumable materials, especially polypropylene tips, tend to adsorb certain compounds, leading to unreliable concentrations in the destination plates (Harris et al. 2010). Therefore, it is recommended to test and validate the influence of consumables on an assay during assay development and whenever there is a change in labware.

3.4.3. *Foaming*

Pipetting viscous and “sticky” samples is challenging due to bubble formation. Among the most important parameters to consider in avoiding these issues are the speed that the tips exit the sample fluid and the aspirate/dispense rates; they should be slow enough to avoid residuals at the inside and outside of the tips. Pre- and post-air pipetting options should be avoided.

3.4.4. *Pipette behavior affecting dispensing variation*

Most pipettor systems provide pre- and post-air aspiration functions to ensure accurate liquid transfers. Introduction of air into the tips before or after the aspiration of the sample liquid is recommended to improve volume accuracy by forcing all the liquid out of the tips. In a study performed to optimize the automated parameters to achieve a 10 μL transfer volume in a sequential transfer experiment, introduction of a 5 μL pre-air gap significantly reduced the relative volume inaccuracy along with the CV of the final transferred volume in a 96-well plate (Albert et al. 2007). While this method may help to achieve more precise results especially with small volume transfers, bubble formation in the destination wells may be inevitable unless proceeded by a shaking or centrifugation step. Post-air aspiration may also be applied to create an air gap between liquids, preventing unsought contamination in the source reservoirs when multiple samples are picked up sequentially into a single tip before the delivery into the destination reservoir.

When small and repetitive volume transfers into multiple destinations are needed, it is a common practice to pick up a single large volume and deliver smaller amounts in a sequential mode. However, with this method, it is hard to achieve accurate delivery in each step. In a study of multi-sequential dispense accuracy, it was shown that the first and last dispense steps led to relatively higher and lower transferred volumes, respectively, in addition to increased relative inaccuracy (Albert et al. 2007). Therefore, it is recommended to dispense the first and last steps into the source reservoir to enhance the precision in the destination plate. Delivery performance of the dry versus pre-wetted tips may also exhibit differences in variability depending on the sample characteristics.

Droplet formation at the end of the pipette tips after a dispense action remains an issue for liquids with high viscosity or low densities. Besides the selection of the optimum dispense

speed, a “tip touch” function is a useful feature offered in some automated pipettors, where the tips contact the well wall at the end of a dispense step to force the release of the droplet. The path of the moving pipetting arm across the deck should be carefully determined to reduce the chance of contaminating other labware by hanging droplets.

Proper mixing of solutions in the source reservoir before aspiration and in the destination reservoir after dispensing may greatly affect the final assay quality due to the necessity of uniform sample concentrations. To avoid the formation of concentration gradient in wells, mixing can be performed by repetitive pipetting cycles. Mixing of the well contents by pipetting up and down is proven to be a quicker and more efficient method compared to free diffusion or shaking, which are not as successful due to the correlations between well size, content volume and the exerted capillary forces (Berg et al. 2001; Shieh et al. 2010; Travis et al. 2010). Mixing is necessary when dealing with suspensions (cells, beads, etc.). For instance, cell settling creates uneven cell density in the source reservoir, which would lead to aspiration of decreasing number of cells over time (Fig. 2C)

3.4.5. Routine quality assessment

Verification of transferred volumes and routine quality control (QC) are the most important and inevitable processes when working with liquid handling devices. While the verification method should be reliable enough to quantify the pipettor performance, it should also be easy and fast to be applied routinely. The performance assessment described for bulk liquid dispensers (section 3.2.7.) can also be applied to pipettors as long as the same volume is distributed throughout the plate for %CV and %bias' calculations.

As mentioned previously, liquid handlers are heavily used to perform serial dilutions, and suitable QC techniques should be employed to validate dilution performance, particularly when accurate compound potency is directly dependent on concentration accuracy. Dilution ratio, accuracy, precision and outlier distribution constitute the four major criteria that should be evaluated (Popa-Burke et al. 2009). Artel developed an approach to determine dilution and transferred volume accuracy by using dual-wavelength photometry, where two absorbance dyes with baseline resolved spectra are mixed at various ratios using a liquid handler (Albert 2007; Dong et al. 2007). This dual-dye ratiometric method can be applied by using a multichannel verification system (MVS) equipped with the necessary instrumentation and analysis (Bradshaw et al. 2005). Dual-dye photometry is also proven to be suitable to measure the efficiency of different mixing methods (Spaulding et al. 2006) and when pipetting non-aqueous solutions (Bradshaw et al. 2007).

3.5. Considerations for using microplate washers

3.5.1. General considerations

One of the major concerns with any high throughput microplate handler device is its compatibility with plates of various types and sizes. While most high throughput instruments

are designed to accommodate labware with dimensions conforming to ANSI/SBS standards, an ideal plate washer is also able to support flat, v-shaped and round-bottom plates.

Both the vacuum assembly and the bottle setup are also important aspects of the plate washer. Although most washers operate through changes in vacuum pressure, pump-based vacuum-free and pressure-free systems are also offered.

Plate washers functioning by positive displacement principle are also available, enabling non-contact washing with no residual volume (Rudnicki and Johnston 2009). For assays where more than one wash buffer may need to be used, plate washers with multiple dispense channels and automatic buffer switching capability are preferred to minimize both operation time and contamination. Examples of other optional features for safe instrument operation include waste liquid level sensors and plate detection sensors to avoid unwanted overflows and jams. For BSL2 or higher level experiments, a washer with aerosol cover should be chosen to prevent spread of the contagious material.

3.5.2. Washer performance

Although compatibility and control properties are important, plate washers are predominantly evaluated by their wash performance. Plate washers provide a range of user-defined dispense/aspirate heights, flow rates, and needle probe positioning in reference to the well walls. By adjusting these parameters for each step of the wash cycle, optimal wash performance can be ensured. On the other hand, an adequate wash quality needs to be reached to diminish extensive background signal and high signal variations amongst wells. This can be primarily achieved by minimizing the amount of liquid left inside each well at the end of the aspiration step. Besides their effects on wash power, the above-mentioned parameters also have an impact on the residual volume and need to be fine-tuned in conjunction with the vacuum/pump settings. Some plate washers may also provide multipoint, secondary, cross-wise or delayed aspiration modes aiming to deliver the best results. The number of wash cycles and the length of soaking time are other settings that can be modified to reduce background noise levels.

3.5.3. Washer maintenance

Since plate washers consist of tubing and needles which transport buffer solutions or waste liquid to or from the device, they require special cleaning processes as they are prone to be clogged by chemical residues such as salt and proteins from the wash liquids. Depending on the frequency of use, the fluid path may need to be rinsed daily to prevent blockage and contamination, especially if different buffers are being delivered through the same tubing. An efficient cleaning method alternates deionized water and a detergent such as Terg-a-Zyme®, which is highly recommended by plate washer manufacturers. Plate washers which provide an automatic cleaning feature or integrated ultrasonic washing technology are often easier to maintain. Models which do not contain built-in cleaning functionality are generally supplied with removable dispense/aspiration manifolds to ease the maintenance tasks.

Cleaning of the other detachable or fixed plate washer components should also be performed periodically.

3.5.4. Troubleshooting

Plate washers serve as an excellent alternative to time consuming manual wash procedures for many applications. Since all the wash parameters should be optimized for each specific application during the assay development stage, a tedious troubleshooting process may be inevitable while setting up wash protocols to meet specific assay needs. Table 1 presents a summary of wash parameters/components and their contributions to the wash performance along with various troubleshooting tips. Different assay types may require distinct considerations. With biochemical assays, minimizing the background signal and well-to-well variations are the most important tasks in the optimization process. Low background signal levels can be achieved by reducing the leftover liquid volume in each well. Decreasing the aspiration height and lowering the aspiration rate can greatly affect the residual volume leading to minimal liquid amounts in the wells. In order to prevent high standard deviations in the assay readouts, equal residual volumes should be attempted by optimizing the aspiration/dispense heights and rates. Depending on the viscosity of the wash buffer, high aspiration rates or low dispense rates may lead to unequal volumes. Inadequate priming volumes, unadjusted dispense or aspiration heights, clogged tubing, and physical misalignments between the manifolds and plate carrier should also be avoided to prevent high signal variations. The effect of the aspiration height on the final residual volume is presented in Fig. 6 for both 96- and 384-well black plates with clear bottom. The volume of the residual liquid (water) per well was measured with the gravimetric technique at several selected aspiration heights on a Biotek EL405 microplate washer, while all the other wash parameters were kept constant. A rising trend is observed in the final volume as the aspiration height is increased.

Figure 6. Effects of aspiration height on residual volume. Residual volume was measured in a) 96-well and B) 384-well plates at various aspiration heights. Residual volume was increased as the aspiration height from the bottom of the well was increased.

In cell-based assays, gentle cell washing is one of the most critical factors to produce reproducible assay results, and it can be controlled by several settings such as aspiration and dispense rates, heights and horizontal positions. For loosely-adherent cells, the cell layer attached to the bottom of the well may be easily disrupted by rigorous wash cycles, and the aspiration and dispense rates should be set low enough to prevent turbulence inside the wells. For the same purpose, wash fluid should be dispensed at a distance from the well bottom and may be even be aimed at the well walls when possible. To observe the consequences of inadequate washing and dispensing parameters on the cell layer endurance, a 3-cycle wash experiment was performed on HEK 293T cells, which are known for their low adherence and propensity to be frequently washed away in cell-based assays. The fixing solution was dispensed at medium speed, and the cells were washed before and after fixation. Representative images from wells containing an intact or damaged cell layer are presented in Fig. 7. When dealing with adherent cells, each step of the assay protocol should be optimized, including those involving other liquid handling devices such as bulk dispensers, pintools and pipettors.

Figure 7. Effects of non-optimized dispensing and washing on low-adherent cells. HEK293T cells were fixed, stained with Hoechst 33258 and imaged with Acumen eX3 in a 384-well black clear bottom plates. The fixing solution was dispensed by a Thermo Scientific Matrix® Wellmate®. Representative images (shown here in false color green) of A) an intact cell layer and B) disrupted cell layers indicated cell loss due to harsh dispense and wash settings.

As with most high throughput instrument operations, it is a common practice to perform a periodic quality check on plate washers to assure a satisfactory wash performance at each use. It is important to perform these assessments with a wash buffer that has a similar viscosity to the buffers used in most of the applications. For evaluations on the residual volume, one can perform a mock wash with a dummy plate and measure the leftover liquid volume inside the wells with a single or multichannel manual pipettor. For more accurate results, gravimetric or colorimetric techniques can be used to calculate the average volume per well. This way, one can also test if dispensing/aspiration is consistent in all the probes, and if there is any physical failure with any of the device components.

parameter/component	effect	troubleshooting tips
prime	dispense performance	<ul style="list-style-type: none"> • prevent air bubble formation or no/uneven dispensing with adequate priming
aspiration rate	residual volume, gentle/rigorous washing	<ul style="list-style-type: none"> • higher residual volume if too fast • perturbed cell layer if too fast • uneven aspiration if too fast
aspiration height	residual volume, gentle/rigorous washing	<ul style="list-style-type: none"> • higher residual volume if too high • uneven aspiration if too low or too high • perturbed cell layer if aspiration probes touch the well bottom • undisturbed cell layer if high enough
horizontal aspirate position	gentle/rigorous washing	<ul style="list-style-type: none"> • prevent bead loss by offsetting the aspirate position (for magnetic bead assays)
dispense flow rate	dispense volume, gentle/rigorous washing	<ul style="list-style-type: none"> • uneven dispensing if too slow • fluid overflow if too slow or too fast • perturbed cell layer if too fast • air bubble formation if too slow
dispense height	dispense volume, gentle/rigorous washing	<ul style="list-style-type: none"> • uneven dispensing if too low or too high • fluid overflow if too high
horizontal dispense position	gentle/rigorous washing	<ul style="list-style-type: none"> • undisturbed cell layer if dispense position is offset to aim the well walls
assay buffer properties	residual volume, aspiration/dispense performance	<ul style="list-style-type: none"> • optimize for viscous/non-viscous buffer solutions • add surfactant to the buffer solution to reduce surface tension
vacuum/pump assembly	aspiration/dispense performance	<ul style="list-style-type: none"> • no/uneven aspiration with insufficient vacuum supply • no/uneven aspiration or leakage if tubing is defective, bent or clogged
plate carrier	aspiration/dispense performance	<ul style="list-style-type: none"> • uneven aspiration/dispense if plate carrier is not leveled or movement is blocked • plate is placed on the carrier with A1 in the correct position • enough plate clearance to prevent jams • higher throughput with lower plate clearance

Table 1. Wash parameters and troubleshooting advices

4. Conclusion

In order to fulfill the need for higher throughput options, the technology behind liquid handling devices is in constant progression, with systems capable of delivering smaller volumes at a faster rate with accuracy and precision. These developments should consider cost reduction by minimizing reagent and solvent expenditure, as well as reducing consumables.

The main concerns and limitations that liquid handling systems face are reproducibility and reliability. The devices should be robust to execute extensive experiments in a daily basis with minimal downtime and maintenance. However, as a single screen can generate thousands of data points, the user is required to ensure all the devices are functioning up to standards by implementing routine quality assessments. Regardless of the technological innovations and advancements, scientists are compelled to spend significant amount of time optimizing the liquid handling parameters to suit specific assay conditions. A thorough understanding of the principles, strengths and limitations of the instruments is advantageous in preventing undesirable results and facilitating troubleshooting.

Acknowledgements

This work was supported by the American Lebanese Syrian Associated Charities (ALSAC), St. Jude Children's Research Hospital, and National Cancer Institute grant P30CA027165.

Author details

Sergio C. Chai, Asli N. Goktug, Jimmy Cui, Jonathan Low and Taosheng Chen

High Throughput Screening Center, Department of Chemical Biology and Therapeutics, St. Jude Children's Research Hospital, USA

References

- [1] Albert, KJ, Bradshaw, JT, Knaide, TR & Rogers, AL (2006). Verifying liquid-handler performance for complex or nonaqueous reagents: a new approach. *Journal of the Association for Laboratory Automation*, Vol. 11(4): pp. 172-80
- [2] Bateman, TA, Ayers, RA & Greenway, RB (1999). An engineering evaluation of four fluid transfer devices for automated 384-well high throughput screening. *Laboratory Robotics and Automation*, Vol. 11(5): pp. 250-59

- [3] Berg, M, Undisz, K, Thiericke, R, Zimmermann, P, Moore, T & Posten, C (2001). Evaluation of liquid handling conditions in microplates. *Journal of Biomolecular Screening*, Vol. 6(1): pp. 47-56
- [4] Bergsdorf, C, Gewiese, N, Stolz, A, Mann, R, Parczyk, K & Bomer, U (2006). A cost-effective solution to reduce dead volume of a standard dispenser system by a factor of 5. *Journal of Biomolecular Screening*, Vol. 11(4): pp. 407-12
- [5] Bradshaw, JT, Curtis, RH, Knaide, TR & Spaulding, BW (2007). Determining dilution accuracy in microtiter plate assays using a quantitative dual-wavelength absorbance method. *Journal of the Association for Laboratory Automation*, Vol. 12(5): pp. 260-66
- [6] Bradshaw, JT, Knaide, T, Rogers, A & Curtis, R (2005). Multichannel Verification System (MVS): a dual-dye ratiometric photometry system for performance verification of multichannel liquid delivery devices. *Journal of the Association for Laboratory Automation*, Vol. 10(1): pp. 35-42
- [7] Cleveland, PH & Koutz, PJ (2005). Nanoliter dispensing for uHTS using pin tools. *Assay and Drug Development Technologies*, Vol. 3(2): pp. 213-25
- [8] Diaspro, A, Chirico, G, Usai, C, Ramoino, P & Dobrucki, J (2006). Photobleaching. In: *Handbook of Biological Confocal Microscopy*. J. B. Pawley. pp 690-702, Springer Science +Business Media, LLC, New York
- [9] Dong, H, Ouyang, Z, Liu, J & Jemal, M (2006). The use of a dual dye photometric calibration method to identify possible sample dilution from an automated multichannel liquid-handling system. *Journal of the Association for Laboratory Automation*, Vol. 11(2): pp. 60-64
- [10] Dunn, DA & Feygin, I (2000). Challenges and solutions to ultra-high-throughput screening assay miniaturization: submicroliter fluid handling. *Drug Discovery Today*, Vol. 5(12): pp. S84-S91
- [11] Ellson, R, Mutz, M, Browning, B, Lee, L, Miller, MF & Papen, R (2003). Transfer of low nanoliter volumes between microplates using focused acoustics—automation considerations. *Journal of the Association for Laboratory Automation*, Vol. 8(5): pp. 29-34
- [12] Felton, MJ (2003). Liquid handling: dispensing reliability. *Analytical Chemistry*, Vol. 75(17): pp. 397a-99a
- [13] Harris, D, Mutz, M, Sonntag, M, Stearns, R, Shieh, J, Pickett, S, Ellson, R & Olechno, J (2008). Low nanoliter acoustic transfer of aqueous fluids with high precision and accuracy of volume transfer and positional placement. *Journal of the Association for Laboratory Automation*, Vol. 13(2): pp. 97-102
- [14] Harris, D, Olechno, J, Datwani, S & Ellson, R (2010). Gradient, contact-free volume transfers minimize compound loss in dose-response experiments. *Journal of Biomolecular Screening*, Vol. 15(1): pp. 86-94

- [15] Harris, DL & Mutz, M (2006). Debunking the myth: validation of fluorescein for testing and precision of nanoliter dispensing. *Journal of the Association for Laboratory Automation*, Vol.11(4): pp. 233-39
- [16] James, P & Papen, R (1998). Nanolitre dispensing - a new innovation in robotic liquid handling. *Drug Discovery Today*, Vol. 3(9): pp. 429-30
- [17] Kong, F, Yuan, L, Zheng, YF & Chen, W (2012). Automatic liquid handling for life science: a critical review of the current state of the art. *Journal of Laboratory Automation*, Vol. 17(3): pp. 169-85
- [18] Lee, S (2006). Development of a vacuum filtration system for the conventional microplate washer. *Journal of Pharmacological and Toxicological Methods*, Vol. 53(3): pp. 272-6
- [19] McDonald, GR, Hudson, AL, Dunn, SMJ, You, HT, Baker, GB, Whittall, RM, Martin, JW, Jha, A, Edmondson, DE & Holt, A (2008). Bioactive contaminants leach from disposable laboratory plasticware. *Science*, Vol. 322(5903): pp. 917-17
- [20] McGown, EL & Hafeman, DG (1998). Multichannel pipettor performance verified by measuring pathlength of reagent dispensed into a microplate. *Analytical Biochemistry*, Vol. 258(1): pp. 155-57
- [21] Murthy, TV, Kroncke, D & Bonin, PD (2011). Adding precise nanoliter volume capabilities to liquid-handling automation for compound screening experimentation. *Journal of the Association for Laboratory Automation*, Vol. 16(3): pp. 221-28
- [22] Niles, WD & Coassin, PJ (2005). Piezo- and solenoid valve-based liquid dispensing for miniaturized assays. *Assay and Drug Development Technologies*, Vol. 3(2): pp. 189-202
- [23] Niles, WD & Coassin, PJ (2008). Cyclic olefin polymers: innovative materials for high-density multiwell plates. *Assay and Drug Development Technologies*, Vol. 6(4): pp. 577-90
- [24] Olechno, J, Shieh, J & Ellson, R (2006). Improving IC₅₀ results with acoustic droplet ejection. *Journal of the Association for Laboratory Automation*, Vol. 11: pp. 240-46
- [25] Petersen, J & Nguyen, J (2005). Comparison of absorbance and fluorescence methods for determining liquid dispensing precision. *Journal of the Association for Laboratory Automation*, Vol. 10: pp. 82-87
- [26] Popa-Burke, I, Lupotsky, B, Boyer, J, Gannon, W, Hughes, R, Kadwill, P, Lyerly, D, Nichols, J, Nixon, E, Rimmer, D, Saiz-Nicolas, I, Sanfiz-Pinto, B & Holland, S (2009). Establishing quality assurance criteria for serial dilution operations on liquid-handling equipment. *Journal of Biomolecular Screening*, Vol. 14(8): pp. 1017-30
- [27] Ramírez, M, Tudela, C, Crespo, JB, Rubles, A, Jiménez, A, Tarazona, G, Fernández, L, Sánchez, P, Kroncke, D, Tabanera, N, Pérez, J, Sánchez, M, Tormo, JR & Peláez, F (2008). Automation of sample dispensing using pocket tips for evaluation of CYP2C9

isoenzyme reversible inhibition in an evolution precision pipetting platform. *Journal of the Association for Laboratory Automation*, Vol. 13: pp. 289-96

- [28] Rhode, H, Schulze, M, Renard, S, Zimmerman, P, Moore, T, Cumme, GA & Horn, A (2004). An improved method for checking HTS/uHTS liquid-handling systems. *Journal of Biomolecular Screening*, Vol. 9(8): pp. 726-33
- [29] Rose, D (1999). Microdispensing technologies in drug discovery. *Drug Discovery Today*, Vol. 4(9): pp. 411-19
- [30] Rudnicki, S & Johnston, S (2009). Overview of liquid handling instrumentation for high-throughput applications. *Current Protocols in Chemical Biology*, Vol. 1(1): pp. 43-54
- [31] Schober, A, Gunther, R, Schwienhorst, A, Doring, M & Lindemann, BF (1993). Accurate high-speed liquid handling of very small biological samples. *Biotechniques*, Vol. 15(2): pp. 324-29
- [32] Shieh, J, Ellson, RN & Olechno, J (2006). The effects of small molecule and protein solutes on acoustic drop ejection. *Journal of the Association for Laboratory Automation*, Vol. 11(4): pp. 227-32
- [33] Shieh, J, Jamieson, B & Vivek, V (2010). Evaluating five microplate mixing techniques: diffusion, centrifugation, shaking, pipetting and ultrasonic mixing. Poster presented at Society for Biomolecular Sciences, Phoenix, AZ
- [34] Spaulding, BW, Bradshaw, JT & Rogers, AL (2006). A method to evaluate mixing efficiency in 384-well plates. Poster presented at Society for Biomolecular Sciences, Seattle, WA
- [35] Stobbs, LW (1990). Construction of a simple, inexpensive multiple enzyme-linked immunosorbent assay microdilution plate washer. *Applied and Environmental Microbiology*, Vol. 56(6): pp. 1763-7
- [36] Taylor, PB, Ashman, S, Baddeley, SM, Bartram, SL, Battle, CD, Bond, BC, Clements, YM, Gaul, NJ, McAllister, WE, Mostacero, JA, Ramon, F, Wilson, JM, Hertzberg, RP, Pope, AJ & Macarron, R (2002). A standard operating procedure for assessing liquid handler performance in high-throughput screening. *Journal of Biomolecular Screening*, Vol. 7(6): pp. 554-69
- [37] Travis, M, Vivek, V, Jamieson, B, Shieh, J & Weaver, D (2010). Mixing effectiveness – a methodology and study of microplate mixing techniques including ultrasonic HEN-DRIX SM100. Poster presented at Association for Laboratory Automation, Palm Spring, CA
- [38] Turmel, M, Itkin, Z, Liu, D & Nie, D (2010). An innovative way to create assay ready plates for concentration response testing using acoustic technology. *Journal of the Association for Laboratory Automation*, Vol. 15(4): pp. 297-305

- [39] Walling, LA (2011). An inline QC method for determining serial dilution performance of DMSO-based systems. *Journal of the Association for Laboratory Automation*, Vol. 16(3): pp. 235-40
- [40] Watson, J, Greenough, EB, Leet, JE, Ford, MJ, Drexler, DM, Belcastro, JV, Herbst, JJ, Chatterjee, M & Banks, M (2009). Extraction, identification, and functional characterization of a bioactive substance from automated compound-handling plastic tips. *Journal of Biomolecular Screening*, Vol. 14(5): pp. 566-72