

We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,900

Open access books available

186,000

International authors and editors

200M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities



WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com



Introductory Chapter: History and Scope of Quality Control in Laboratories

Gaffar Sarwar Zaman

Additional information is available at the end of the chapter

<http://dx.doi.org/10.5772/intechopen.74593>

1. Introduction

Laboratory quality control encompasses the integral and essential monitoring of investigative analytics to find out investigative and interfering flaws, miscalculations and blunders that appear or materialize during the analytical procedures and conclusively avert the erroneous reporting of results to the patient. Quality control is usually used to monitor the accomplishment of a test, and, to find the accuracy. Levey and Jennings were the pioneers of using statistics for the standardization and calibration of analytical procedures. The substances used for quality control are known as quality control (QC) materials. They are usually aliquoted in a stable form. Nowadays, most laboratories purchase the QC material from reputed companies instead of making them. The QC material is usually in a powdered form and can be stored for more than a year. The concept of statistical quality control decreased to a great extent the cost of quality control by the help of the method of sampling.

2. Historical perspective of quality control in laboratories

Walter Andrew Shewhart (March 18, 1891–March 11, 1967) was an American engineer, statistician and physicist, also called “the Father of Modern Quality Control,” and he is also credited as the founder of the “Shewhart cycle” [1] (**Figure 1**). Walter Shewhart was instrumental in the introduction and development of “Process Control” in the year 1924. Prevention of the manufacture of defective products was the main aim of this method. For



Figure 1. Walter Andrew Shewhart (March 18, 1891–March 11, 1967) [reference: <http://www.york.ac.uk/depts/maths/histstat/people/shewhart.gif>] [4].




Figure 2. William Edwards Deming (October 14, 1900–December 20, 1993). <http://www.fda.gov/oc/initiatives/criticalpath/stanski/stanski.html> [5].

this, he also suggested and developed control charts, known as “Average Shewhart Chart” and “R Shewhart Chart or Range Shewhart Chart”. W. Edwards Deming was another pioneer in this field [2–5] (**Figure 2**). He developed 14 rules, which comprised of a series of successive steps of testing. His contributions boosted Japan’s rapid industrial growth in the post-war period.

2.1. Levey-Jennings chart

It is a chart in which the data from a quality control are plotted and from which visually we can find out whether a particular laboratory test is working or not. The name of the chart is given after Stanley Levey and Elmer R. Jennings who first introduced this chart in the 1950s. It has become so popular that nowadays that it is even used for automated analyzers.

When Rausch and Freier introduced that serum pools should be used in place of samples from patients, the chart of Stanley Levey and Elmer R. Jennings became even more popular. Thus, these samples that came to be called as “standards” were ultimately known as “control samples” [6].

In 1954, E.S. Page of the University of Cambridge introduced the sequential analysis technique known as the cumulative sum control chart (CUSUM). CUSUM was devised as a method to find out changes in the techniques or quality and also the exact time as to when to take corrective actions. But the CUSUM chart was only analyzed for feasible use in the laboratory many years later. It was followed by the invention of the “Exponentially Weighted Moving Average (EWMA)” chart in 1959 by the American S.W. Roberts [7]. However, the EWMA chart was adapted for use in medical laboratory applications much later [8].

However the use of patient results for quality control only started during 1960–1970. One of the pioneers in this field was the Japanese statistician Kaoru Ishikawa. Quality control was used for patients’ tests involving hematology and biochemistry laboratories [9]. Dennis Dorsay, in 1963, was one of the pioneers who stressed the importance of erythrocyte indexes for quality control in various hematology analyzers [10].

Repeated assaying of whole blood samples from two successive days was advocated by Frank Ductra in 1966; this was to be done in place of control samples and also revolutionized the quality control process [11].

In 1965, Michael Waid and Robert Hoffmann introduced the unique “Average of normals” (AON) method where systematic errors can be detected by the arithmetic average of normal test results produced by biochemical analyses [12].

Quality control in hematology analyzers, called “Bull’s algorithm” or “ X_B ”, was introduced in 1974 by the American hematologist Brian Bull [13].

The introduction of “Computer simulations” brought about a big change in the issues of quality control. It was introduced by the Swedish clinical engineer Torsten Aronson, medical doctor Carl-Henric de Verdier and a physicist Torgny Groth [10, 14]. In the same year, Arthur Gottmann and Jerome Nosanchuk [13, 14] utilized the new method of comparing each patient’s results with previous results, within a specified time, to detect any errors made by analyzers. It proved to be a reliable and cost-effective quality method. The distinctive feature of this method was the use of patient’s results instead of control samples, and it does not make any discrimination between normal and pathological values.

The “delta check method” (in comparison with the previous record) and the “rate check method” in which the time elapsed between measurements were being considered and were suggested by Nosanchuk and Gottmann [15]. Usefulness of the moving average was elucidated by the Canadian clinical chemist George Cembrowski and the American clinical pathologist James

Westgard in 1975 (**Figure 3**). A year after, the use of anion gap equation for automated blood gas and electrolyte analyzers quality control was advocated [16] by David Witte and co-workers.

“A multi-rule Shewhart Chart for quality control in clinical chemistry” was published during the 1980s by Westgard, marking a major breakthrough in quality control for laboratories. The simple rules explaining implementation of the Levy-Jennings chart were given in this chapter.

The initial international quality standard for operations in a clinical laboratory was also established during the 1980s. During the 1990s, the theoretical and practical application of biological variances as analytical targets in clinical chemistry [18, 19] were worked upon by Fraser and his co-workers, distinguishable among them being Eugene Harris, the American clinical chemist, who was instrumental in contributing to the formulation of the theory of biological variances through his expertise of statistics and informatics [20]. Another notable contribution is that of Carmen Ricos [21–23] and her group of Spanish researchers (majority), who were responsible for collecting data on quality specifications and biological variances a number of biochemical parameters.

The “OPSpecs charts” [24] concept was proposed by Westgard in 1994. Non-analytical errors, that is, errors that occur before or after analysis, were also discussed extensively during the 1990s. Configuration of laboratory information systems (LISs) led to the prevention of post-analytical errors and some types of pre-analytical errors.



Figure 3. Professor James O. Westgard is President of Westgard QC, Inc., a small business providing education and training for laboratory quality management. He is an Emeritus Professor in the Department of Pathology and Laboratory Medicine at the University of Wisconsin Medical School [reference: <https://www.labqualityconfab.com/speakers/james-o-westgard>] [17].

Later on, Westgard introduced the six sigma theory in clinical chemistry, which proved to be another method of establishing quality specifications [25].

3. Salient features

The QC should proceed through three parts, mainly:

1. Each analytic method should have its own statistical limits of variation.
2. These limits should be utilized for finding out the QC data which is generated for each type of test.
3. Elimination of the various errors, and if found:
 - (I) The cause of the error should be found out.
 - (II) Action should be taken to correct the error.
 - (III) The patients' data should be re-analyzed.

Multirule procedure: this includes decision criteria to determine if an analytic run is in control; it is used to detect random and systemic error over time and is developed by Westgard and Groth [26].

Proficiency testing, internal quality control, laboratory inspections, clinical utilization and quality assurance monitoring play an important role as indicators of analytic performance. Management of quality consists of quality design, quality control and quality improvement [26] (**Figure 4**).




Figure 4. Relation between various aspects of quality control [26].

Use of automated analyzers in clinical laboratories: nowadays, almost every laboratory uses automated analyzers. The reason is that they are more reliable, can process more samples at a time, and are time saving and also cost saving in the long run. Most companies provide the quality control material along with the quality control guide. This has made it easier for laboratories to assess quality of the various types of parameters.

Author details

Gaffar Sarwar Zaman

Address all correspondence to: gffrzaman@gmail.com

King Khalid University, Abha, KSA

References

- [1] Shewhart WA. A Study of the Accelerated Motion of Small Drops Through a Viscous Medium. Lancaster, PA: Press of the New Era Printing Company; 1917. 433 p
- [2] Shewhart WA. Economic Control of Quality of Manufactured Product. New York: D. Van Nostrand Company. 1931. 501 p. ISBN: 0-87389-076-0
- [3] Shewhart WA. Statistical method from the viewpoint of quality control. Edwards Deming W, editor. Washington, DC: The Graduate School, the Department of Agriculture; 1939. 155 p. ISBN: 0-486-65232-7
- [4] <http://www.york.ac.uk/depts/maths/histstat/people/shewhart.gif>
- [5] <http://www.fda.gov/oc/initiatives/criticalpath/stanski/stanski.html>
- [6] Freier E, Rausch V. Quality control in clinical chemistry. American Journal of Medical Technology. 1958;5(4):309-319
- [7] Roberts S. Control charts tests based on geometric moving averages. Technometrics. 1959;1:239-250
- [8] Neubauer A. The EWMA control chart: Properties and comparison with other quality-control procedures by computer simulation. Clinical Chemistry. 1997;43:594-601
- [9] Karkalousos P, Evangelopoulos A. The history of statistical quality control in clinical chemistry and haematology (1950-2010). International Journal of Biomedical Laboratory Science (IJBLS). 2015;4(1):1-11
- [10] Dorsey D. Quality Control in Haematology. American Journal of Clinical Pathology. 1963;40(5):457-464

- [11] Ductra F. Monitoring the quality of blood cell counts with replicate determinations on routine samples. *American Journal of Clinical Pathology*. 1966;**46**(2):286-288
- [12] Hoffmann R, Waid M. The “average of normals” method of quality control. *American Journal of Clinical Pathology*. 1965;**43**:134-141
- [13] Bull B. A study of various estimators for derivation of quality control procedures from erythrocyte indices. *American Journal of Clinical Pathology*. 1974;**61**:473-481
- [14] Aronson T, de Verder C-H, Groth T. Factors influencing the quality of analytical methods—A systems analysis with use of computer simulation. *Clinical Chemistry*. 1974;**20**:738-748
- [15] Nosanchuk J, Gottmann A. Cums and delta checks. *American Journal of Clinical Pathology*. 1974;**62**:707-712
- [16] Witte D, Rodgers J, Barrett D. The anion gap: Its use in quality control. *Clinical Chemistry*. 1976;**22**:643-646
- [17] <https://www.labqualityconfab.com/speakers/james-o-westgard>]
- [18] Fraser C, Harris E. Generation and application of data on biological variation in clinical chemistry. *Critical Reviews in Clinical Laboratory Sciences*. 1969;**27**:409-437
- [19] Harris E, Kanofsky P, Shakarji G, Cotlove E. Biological and analytic components of variation in long-term studies of serum constituents in normal subjects. *Clinical Chemistry*. 1970;**16**(12):1022-1027
- [20] Ricós C, Alvarez V, Cava F, García-Lario JV, Hernández A, Jiménez CV, Minchinela J, Perich C, Simón M. Current databases on biological variation: pros, cons and progress. *Scandinavian Journal of Clinical and Laboratory Investigation*. 1999;**59**:491-500
- [21] Westgard J, Quam E, Barry P. QC Selection grids for planning QC procedures. *Clinical Laboratory Science*. 1990;**3**:271-278
- [22] <http://www.jano.es/noticia-carmen-ricos-premiada-con-el-18844>
- [23] Westgard J, Burnet R, Bowers G. Quality management science in clinical chemistry: A dynamic framework for continuous improvement of quality. *Clinical Chemistry*. 1990;**36**:1712-1716
- [24] Lippi G. Governance of preanalytical variability and error detection. *JMB*. 2008;**25**(3): 337-338
- [25] Westgard JO, Groth T. Power functions for statistical control rules. *Clinical Chemistry*. 1979;**25**:863-869
- [26] Bishop ML, Fody EP, Schoeff LE. *Clinical Chemistry: Principles, Techniques, and Correlations*. 7th ed. Lippincott Williams and Wilkins. 2013. ISBN-10: 1451118694, ISBN-13: 978-1451118698

