

# We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,900

Open access books available

185,000

International authors and editors

200M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities



WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index  
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?  
Contact [book.department@intechopen.com](mailto:book.department@intechopen.com)

Numbers displayed above are based on latest data collected.  
For more information visit [www.intechopen.com](http://www.intechopen.com)



# Control of a Prosthetic Arm Using fNIRS, a Neural-Machine Interface

*Usama Ali Syed, Zareena Kausar and Neelum Yousaf Sattar*

## Abstract

Development in the field of bio-mechatronics has provided diverse ways to mimic and improve the function of human limbs. Without an elbow joint, the hand remains stiff because all the muscles tension passes through this joint. Advanced myoelectric prosthetic devices are limited due to the lack of appropriate signal sources on residual amputee muscles and insufficient real-time control. Neural-machine interfaces (NMI) are representing a recent approach to develop effective applications. In this research study, an NMI is designed that presents real-time signal processing for command generation. The human brain hemodynamic responses are, therefore, translated into control commands for people suffering from transhumeral amputation. A novel and first of its kind scheme is proposed which utilizes functional near-infrared spectroscopy (fNIRS) to generate the control commands for a three-degree-of-freedom (DOF) prosthetic arm. The time window for fNIRS signals was set to 1 second. The average accuracy was found to be 82% which is a state-of-the-art result for such a technique. The accuracy ranged from 65 to 85% subject-wise. The data were trained and tested on both artificial neural network (ANN) and linear discriminant analysis (LDA). Eight out of 10 motions were correctly predicted in real time by both classifiers.

**Keywords:** functional near-infrared spectroscopy (fNIRS), real-time signal processing, upper-limb prosthesis, transhumeral amputees, artificial neural network, linear discriminant analysis

## 1. Introduction

Amputation is taken from the Latin terminology “amputare” meaning to cut out. It is a removal of a limb due to medical reasons such as diseases and accidents. After this, an artificial device (prosthetics) is provided to fulfill all the desired needs. A prosthesis is an artificial device that replaces a missing body part that may be lost due to any traumatic accidents or medical reasons.

In the nineteenth century, hooks and wooden limbs were used as a replacement to fulfill the supporting needs to overcome support as well as the psychological effects experienced during the time. Prosthetic arm is a biomedical device consisting of links and joints in an open or closed system, which is also a combination of electronics. Thus, there is a need for a specified prosthetic which would help in fulfilling the requirements of the patient.

Types of amputation	Types of prosthesis
Shoulder disarticulation	From shoulder below the elbow
Transhumeral	Above elbow
Transradial	Below elbow
Transcarpal	Below elbow

**Table 1.**  
*Types of upper-limb amputation and respective prosthesis type.*


Prosthetics also come under different categories concerning the patient’s demand and desirable need. **Table 1** lists the type of amputation and their respective prosthesis type.

Controlling a prosthetic arm could be done in several ways. Some of them are by using Invasive Methods, which reflects a process in which an instrument is introduced in the human body. In such a process, electrodes are implanted inside the body, which would receive and implement the process. The other referred to as the non-invasive method. A non-invasion process does not introduce instruments into the body but uses the surface information to get its details and the desired output that is to be determined.

In the previous studies, various strategies have been practiced to monitor muscle activations all through activities, as reported by Lobo Prat et al. [1] To carry out a valuation of muscular contraction, sonomyography (SMG), mechanomyography (MMG) [2, 3], miokinematic (MK), and electric impedance estimations are classically applied. Though muscular intentions and/or contraction are often determined using electromyography (EMG) and near-infrared spectroscopy (NIRS) [4], it allows continuous monitoring of the muscle during motor actions or rehabilitative movements. Further practices, such as ultrasonography [5] and lactate sampling, offer only a representation of the muscular status at the moment of the study, and not an effective trace in time.

Optical brain imaging is a frequently applied methodology in human-machine interaction technically acknowledged as functional near-infrared spectroscopy (fNIRS). It allows them to monitor the quantification of the relative changes in concentration of oxygenated and deoxygenated hemoglobin in tissue blood based on artificial diffuse spectroscopy. Functional neuroimaging suggests a non-invasive method of indirect as well as direct monitoring of brain activity. The hardware involved is portable hence making it easy to carry out experiments in any environment. fNIRS is a non-invasive brain imaging method including the quantification of chromophore concentration determined from the measurement of near-infrared (NIR) light attenuation or time-based changes [6]. It exploits the optical window in which the fundamental elements found in the human body typically cause no hindrance to infrared light of small wavelength range that is 700–900 nm. In addition to that, oxygenated hemoglobin (Hb) and deoxygenated-hemoglobin (deoxy-Hb) are strong absorbers of light which are the key components to translate the brain response [7]. The peculiarity in the absorption bands of deoxy-Hb and oxy-Hb permit the estimation of near changes in hemoglobin concentration by methods of estimating light attenuation at a couple of known wavelengths [8]. The reason behind choosing more than one wavelength is to take care of the isosbestic point that occurs at around 810 nm. At this value of light wavelength both the absorbing coefficients are indistinguishable [9].

Via the improved Beer-Lambert law [10], relative concentration is evaluated concerning the entire length covered by the light photon [11, 12]. Now, for an



**Figure 1.**  
 Raw fNIRS light intensity.

incident ray of light an emitter and to detect that reflected light, a detector is positioned. The distance between them is also defined. Hence, a state of brain hemodynamic is captured. The raw light concentrations that are further transformed into hemodynamic responses by the implementation of renowned Beer-Lambert Law and further utilized for feature extraction and classification. The extracted light intensity patterns can be viewed in **Figure 1**.

Several research institutions have undertaken the design and construction of robotic arms. These structures diverge depending upon the proposed utilization of the human hand. Diverse knowledge of actuation approaches has been considered and implemented. Earlier design approaches have focused on the mechanical problems of the construction and operation of the prosthetic devices. Most of these hardware devices are controlled via methods that are not natural, such as using the contraction of muscles of the opposite arm. This research attempts to lay the foundation for a scheme that can offer functionality similar to the human arm, with an intuitive technique of control.

A search carried out using Web of Science engine, to review work done in this area along with the gap identification, revealed no work done so far in this field of study.

## 2. Materials and methods

The human arm is capable of performing seven basic motions associated with joints in the human arm. To account for transhumeral amputation, three of the main arm motions are considered i.e. one elbow and two motions affiliated with the wrist joint. These motions comprise wrist extension (WE), elbow flexion (EF), wrist supination (WS), wrist flexion (WF), elbow extension (EE), and wrist pronation (WP).

This section elaborates on the data acquisition of the defined motions. These motions were captured using The NIRsport manufactured by NIRx Technologies. It is an accessible, segmental, and robust functional near-infrared spectroscopy (fNIRS) machine that measures hemodynamic responses generated by neuro-activation of the inside brain via oxy-, deoxy-, and total hemoglobin variations in the cerebral cortex.

The updated version, NIRsport-2 proposes a host prepared to implement advancements and units to meet the requirements of wide-ranging cognitive

neuroscience applications. The major advantage of this device is that it is designed to work in a rugged environment and also this device is portable. This feature aids in the development of a wearable and portable system.

2.1 Sample

A total of 15 healthy subjects were engaged who were reported to be right-hand dominant males having a mean age of 30 with a standard deviation of 4. Righthanders had been pursued to confine any dissimilarities in the hemodynamic responses because of the hemispheric domination difference [13]. All participants were chosen wisely as no one of the selected subjects had engaged in any prior study associated with brain signal acquisition experiments. None was accounted for to have a past filled with any mental, neurological, or visual affliction. Every one of them had an ordinary vision, and all signed and agreed to a composed consent after being briefed in detail regarding the test procedure. Three amputee subjects also participated in the study. Their demographics are given in **Table 2**.

Trials employing fNIRS were permitted by the Air University Human Research Ethics Committee (HREC). These research experiments were held regarding ethical standards dictated by the world medical association in the recent declaration of Helsinki [14].

The generic methodology can be seen in **Figure 2**.

2.2 fNIRS data acquisition

2.2.1 Preparation

The specially designed fNIRS headset i.e. Easycap by NIRx technologies follows the international standard of source-detector separations i.e. 3 cm [15–24]. After the subject wears the cap, the optodes are calibrated. The result of this calibration can be analyzed as in **Figure 3**. The faulty setting is shown in **Figure 3(b)**. The boxes represent optodes. The color bar indicates if the optodes are in contact with the scalp or not and hence the colors are assigned. The white color depicts no connection between scalp and optodes. The red color indicated that the connection between the scalp and optode is critical, i.e. it needs to be adjusted. Sometimes hair comes as a hindrance and just by plugging the optode again in the cap would help establish a better connection. If the issue is not resolved by then, a clinical gel (EASYCAP Supervisc, high-viscosity electrolyte-gel) is used to make sure no hair absorbed the light. The gel is rated healthy and is safe to use with optodes. The yellow color indicates that the connection is acceptable. The signals can be acquired. In this scenario, the machine conditions are calibrated by the machine itself. The machine adds a gain factor to the optodes where the connection is acceptable and it

Patient ID	A1	A2	A3
Gender	Male	Male	Male
Age	23	32	42
Amputated side	Right	Left	Right
Residual length (cm)	14	18	10
Time of amputation (months)	7	24	145

**Table 2.**  
*Demographic characteristics of amputated subjects.*






Figure 2.  
Generic methodology.





Figure 3.  
Visual representation of optode settings. The perfect setting is illustrated in (a) meanwhile the faulty optode settings are shown in (b). The color bar on the right side represents the signal quality class.

is saved in conditions file which is used in signal processing later. The green color shows that the optodes are perfectly placed on the head surface and an excellent connection is established for data acquisition. It can be analyzed in **Figure 3(a)**.

2.2.2 Acquisition

When the optodes are aligned, signal acquisition is started. The test strategy was segmented into training and testing. The subjects were asked to complete six tasks that were identified by fNIRS.

A comfortable chair was set up roughly 100 cm away from the subjects so that the motion cues are visible to them while the screen backlight does not interfere with the optical sensors [25–38]. This environment was set up for signal extraction. The Easycap was prepared in advance to minimize the time consumption during the optode placement process. Yet some of the detectors/sources had to be optimized during the calibration process by fixing hair via gel. The session commenced with an undeveloped span of 30 seconds to create a reference point. Later the screen indicated the participants to perform one of six definite tasks the training was



**Figure 4.**  
*Experimental model for signal acquisition.*

additionally divided into two parts. In the first part, all the tasks were recorded sequentially, i.e. the sequence of the tasks was pre-defined. In the second stage, the subjects were demanded to perform similar motions but with random intentions. All these six tasks were logged by fNIRS. Each task comprised of 10-second trials separated by a 20-second rest session. Particulars about the experimental model are given below in **Figure 4**.

The acquired data was then processed and is briefly illustrated in the coming section.

### 3. Data analysis


#### 3.1 Pre-processing

This section explores and explains the signal processing which includes signal pre-processing, optode selection criteria, statistical feature computation, and the signal classification method to generate a control command for the control of a 3-DOF prosthetic arm designed for transhumeral amputees.

Functional near-infrared spectroscopy is the raw light intensity values recorded during a change in oxygenation and de-oxygenation of the blood in the human brain. With the help of dual-tip optodes, this concentration is recorded with two different wavelengths i.e. 760 and 850 nm. In the nirsLAB environment, the data is further processed. nirsLAB is the signal processing software that comes with the machine. nirsLAB is fully aware of the specifications and conditions applied during signal acquisition, hence the best choice for signal processing. The unwanted data is truncated along with unusual spikes or discontinuities that occurred during acquisition. It is then filtered to compute the hemodynamic states. These hemodynamic states are now utilized to extract the features.

As soon as the light intensities are acquired, they are fed to nirsLAB where first the time of stimulus is defined which in our case is 10 seconds as per the designed experimental paradigm. The data is further marked according to the conditions i.e. motion and rest.

**Figure 5** represents the raw fNIRS data of both wavelengths i.e. 760 and 850 nm. It is evident that the amplitude for both data sets is different as the concentration of hemodynamic response is varying from healthy to the amputated subject. This is because of the absence of an arm. As the brain generates these signals, the unwanted responses die due to the absence of neuron carrier in the brain. The connection of arm and brain is cut because there is nothing present at the receiving end. Further, the discontinuities are removed along with the spikes if there exists any. The data is then fed for filtration. nirsLAB provides the commonly practiced filters for fNIRS data. Band-pass filter was implemented to smooth the acquired light intensities. The filtered and raw data at both the wavelengths are illustrated below in **Figure 6**.



**Figure 5.**  
*The first chunk of fNIRS data according to the defined experimental paradigm is illustrated. The initial and final rest is truncated. (a) fNIRS signals of healthy subjects and (b) fNIRS signals acquired from amputee subject.*


nirsLAB makes use of `firls` and `filtfilt` MATLAB® commands to filter the data. `firls` returns the parameters of a linear-phase filter, while `filtfilt` applies the filter parameters into the data. The latter is set to work as finite impulse response (FIR). The roll-off defines the width of the transition frequency band, i.e. how steep the transition between frequencies which are cut and frequencies which are passed for each of the upper and lower limits of frequency. The width of the transition band is calculated as Eqs. (1) and (2):

$$\text{Upper limit} = 1 + \frac{\text{Roll-off}}{2} \tag{1}$$

$$\text{Lower limit} = 1 - \frac{\text{Roll-off}}{2} \tag{2}$$

This noise-free and minimum artifact data are then used to find the hemodynamic states by using the modified Beer-Lambert Law [10, 39–41]. This light intensity raw data is then used to compute the hemodynamic response of the brain. The hemodynamic changes computed offline in nirsLAB are based on the modified Beer-Lambert law for scattering media, as mentioned above. While in nirsLAB the operator can modify all input parameters of the Beer-Lambert law





**Figure 6.**  
(a) The filtered data for both wavelengths after filtering can be seen in this illustration. Band-pass filter has range of 0.01–0.2 Hz. The roll-off was set to 15 as a default value. (b) The same signal is smoothed and as it is a large time-series, (c) a chunk of signal for a closer look is illustrated which is for the first 5 seconds of the activity during the task.

(absorption coefficients and inter-optode distance), in NIRStar®, these are fixed, as they are calculated real-time. More precisely, the values for real-time  $\Delta\text{HbO}$  and  $\Delta\text{Hb}$  computation are as follows:

Absorption coefficients are 3.843707 and 1.4865865 for 760 nm, deoxy and oxy, respectively, and for 850 nm, 1.798643 and 2.526391 deoxy and oxy, respectively.

The default inter-optode distance is set to 3.0 cm and the absorption coefficient unit is millimole per liter per centimeter (1/cm)/(mmol/L).

Mathematically, it is defined as Eq. (3)

$$\Delta A(\lambda) = \varepsilon(\lambda) \cdot \Delta c \cdot d \cdot \text{DPF}(\lambda) + g(\lambda) \quad (3)$$

where the variables can be defined as A: light reduction, or  $\Delta A(\lambda)$ : changes in light reduction at a given wavelength ( $\lambda$ );  $\varepsilon(\lambda)$ : loss of the chromophore at a certain wavelength ( $\lambda$ );  $\Delta c$ : changes observed in the chromophore absorption; d: distance between source and detector;  $\text{DPF}(\lambda)$ : differential path length factor (DPF) for a certain wavelength ( $\lambda$ );  $g(\lambda)$ : the scattering of the light wave at a certain wavelength ( $\lambda$ ), where  $g$  is annulled since it is presumed to be insignificant when only light attenuation (as in continuous-wave NIRS) is considered [20, 36, 42–47].

The differential path length factor (DPF) is a dimensionless modification factor that takes care of the increase in the optical path length that is produced by the scattering of light in organic tissue. The product of DPF and source-detector separation evaluates the “true” path length that the light has traveled inside the biological tissue cell [37, 38, 48, 49]. For NIRx technologies, this value is set constant for wavelengths 7.25/6.38 for 760/850 nm respectively.

### 3.2 Feature extraction

The mathematical representation of statistical features extracted during the study is given as follows.

Signal mean (SM) was computed as Eq. (4)

$$SM = \frac{1}{N} \sum_{i=1}^N X_i \quad (4)$$

where  $N$  denotes the length of the data points within a segment and  $X_i$  represents the signal values.


Signal peak (SP) is defined by the change in signals amplitude among two adjacent segments which surpass a predefined threshold to reduce noise. It is given by Eq. (5)

$$SP = \sum_{i=1}^N f(|X_i - X_{i+1}|) \quad (5)$$

where  $N$  represents the samples while  $X_i$  and  $X_{i+1}$  represent the successive peaks in the signal. These features are extracted and fed to the classifier to predict the motion.

## 4. Classification process

The statistical features extracted from the data sample are then fed to the classifier. Classifying methods are employed to predict the motion intention. To comprehensively evaluate the performance of features, the two widely used classifiers in pattern recognition were implemented, namely, linear discriminant analysis (LDA) and artificial neural network (ANN). A generic yet comprehensive process is illustrated in **Figure 7**.



**Figure 7.**  
A generic classification process.

#### 4.1 Linear discriminant analysis

Fisher's discriminant analysis or linear discriminant analysis is a method used to dimensionally contract samples of two or more classes to separate them, linearly. This classification method projects all the samples on an imaginary line which is useful for data classification. To cater for the word linear, it suggests that the classifier will dimension the given samples to represent the class information. It characterizes the resulting combinations to reduce the number of arbitrary samples by tracing a set of values in a distinct form. It anticipates the sample information so that each class is isolated without any problem. It decreases intraclass variance and increases the inter-class mean. By doing this, unlike data samples become segmented from each other and their set point shrinks together so that they cannot be mixed with other classes.

LDA is commonly used for pattern classification in offline and online systems. This technique projects all the data points on a line in such a way that each data sample that corresponds to a class is separated effectively. It decreases the intra-class variance and increases the inter-class mean. By doing this, different classes become separated from each other, and their data points get closer together so that they cannot be mixed with other classes. LDA works by maximizing the Fisher's criterion given in Eq. (6)

$$J(\mathbf{v}) = \frac{\mathbf{v}^t \mathbf{S}_B \mathbf{v}}{\mathbf{v}^t \mathbf{S}_w \mathbf{v}} \quad (6)$$

Between classes scatter matrix  $\mathbf{S}_B$  is defined as in Eq. (7)

$$\mathbf{S}_B = \sum_{x_i}^c n_i (\mu_i - \mu)(\mu_i - \mu)^t \quad (7)$$

where  $n_i$  represents several samples that belong to class  $i$ , the class scatter matrix  $\mathbf{S}_w$  is defined as in Eq. (8)

$$\mathbf{S}_w = \sum_{x_i}^c \mathbf{S}_i = \sum_{x_i}^c \sum_{x_k \in \text{Class}(i)} (x_k - \mu_i)(x_k - \mu_i)^t \quad (8)$$

A generalized eigenvector problem can be represented as Eq. (9)

$$S_B v = \lambda S_w v \quad (9)$$

The optimal  $v$  is the eigenvector corresponding to the largest eigenvalue can be written represented as in Eq. (10) provided that  $S_w$  is nonsingular.

$$v = S_w^{-1} (\mu_i - \mu) \quad (10)$$

The classifier results were validated using the cross-validation scheme. The number of folds/layers was set to 10. It means that the entire data was mixed randomly into 10 groups, out of which nine took part to train the classifier while one remains untouched for testing purposes. This process was repeated 10 times until all groups were tested against each other.

As an initial measure, the attributes of the dataset which need to be classified or dimensionally contracted will lead to the choice of applying this method as a classifier or a dimensionality reduction algorithm to play out any desired task. The primary thought of Fisher's analysis is fundamentally to isolate sample classes linearly moving them to an alternate feature-space. In this way, if the considered data set is linearly distinguishable, just using the algorithm as a classifier will yield better results. In any case, if the dataset is not truly distinct the classifier will attempt to sort out this dataset in another space. Yet despite every measure, the classes sample data may overlap due to the non-linear characteristic present in the sampled dataset. For this situation, there emerges a need to utilize another grouping model to manage nonlinearities governing the dataset. Hence, a neural network that comprise of hidden layers is also implemented. As for the neural network, raw data is used as input rather than featured data. This will give a broader idea of how to predict any output based on the input that have non-linear characteristic.

## 4.2 Artificial neural network (ANN)

ANN utilizes multiple neuron layers to map data from one distribution to another for better and optimized classification. A technique called backpropagation helps ANN to learn the relationship between input and output class label. The neural network toolbox provided by MATLAB® was utilized to train the classifier. First, network topology and an activation function were defined and then weights were randomized. The model uses all training data to approximate the error of the predicted output as compared to the actual output. Then it uses the error to adjust the weights so that it could be minimized for the next training data and this process was repeated until the error was minimized. For this network we employed Relu as the activation function; the weights were initialized using the Xavier distribution, the network utilized the Adam optimizer function for gradient descent. We used 60% of data for training, and 20% for testing and validation each. A confusion matrix was generated afterward, which had a class number corresponding to each arm motion. The number of hidden layers was specified i.e. 10, and system training was initiated. Ten neurons were present in each of the intermediate hidden layer. The number of neurons in output or last layer was set to be 6, which is equal to the number of elements in the target vector.

After classifying the information, their real-time testing was performed to ensure the behavior of both classifying techniques. But bear in mind that both of

these classifiers have different parameters and methodologies. They are not compared with each other here but they are implemented to grasp a comprehensive idea of how these different brain hemodynamic intentions can be evaluated. LDA and ANN were both applied separately and the outcomes are discussed in next section.

## 5. Results and discussions


In research, the neural-machine interfaces can have a control foundation of either a single modality or via hybrid activity. This present study dwells on capturing hemodynamic responses from the human brain and generating the control command that can be translated to activate a prosthetic arm for transhumeral amputees. The results found by this particular research are discussed as follows.

These hemodynamic states are mapped using nirsLAB in **Figure 8**. The color bar represents the concentration of oxygenation.

### 5.1 Channel selection


The changes in oxygenated hemoglobin  $\Delta\text{HbO}$  for all 20 channels and six activities of subject 6 are demonstrated in **Figure 9**. All of the optodes were not capturing the true concentration changes while brain activity was performed by the subject. Nevertheless, it was observed that similar channels were active when identical motions were executed while signal acquisition.

The channel outputs in **Figure 8** serve to highlight the need for choosing good channels for recognizing true brain activities. According to our channel choice standard, signal averaging was used. It is understandable by human brain studies that when the right side of the human body is in motion, the left hemisphere of



**Figure 8.** The color bar represents the concentration of oxygenation. Brain activity is shown in (a) the illustration represents condition 1 i.e. motion. It can be seen that the motor region is “hot” when compared to the color bar. It depicts motion state whereas the (b) shows that the brain is undergoing significantly low or minimal hemodynamic changes hence the “cool” values depicting the condition 2 i.e. rest. The red and yellow dots seen in the motor cortex region are representations of optodes and they were positioned according to the international system.





**Figure 9.**  
*Optode-wise hemodynamic status visualization.*

the human brain is activated. As in our case, the subjects were asked to move their right hand, it is obvious that the hemodynamic patterns occurring in the right hemisphere are merely noise. It can be seen in **Figure 8** that all the channels of right hemisphere i.e. channels 1–10 do not show significant activity and later they were discarded while classification. The left side of the motor cortex was however active and the channels from 11 to 20 were used to extract the statistical features which took part in the classification. Features were computed spatially which allowed the overall brain activity on the left side of the motor cortex region of the human brain.


Window sizing of diverse spans has been utilized in several studies for the detection of fNIRS features [50, 51]. It is intended to minimize the window size to generate a fast response for real-time applications. So, the time spans of 0–0.5, 0–1, and 0–2 second windows were selected. These split seconds were employed for investigation of hemodynamic features to secure the best window size that will aid in decreased calculation time.

### 5.2 Classification accuracy

The stated performance outcomes were statistically evaluated based on the number of correctly predicted samples while the activity was completed during the period of 0–10 seconds. These actions were assessed by MATLAB® implementing the 10-fold cross-validation course. Student's t-test was performed to establish the statistical significance of the obtained results. The confidence interval was set to 95% ( $p < 0.05$ ). The quantitative comparison between healthy subjects and amputees was not possible due to a limited number of amputees. The computed p-value was 0.0337 considering a 95% confidence interval. The classification accuracies of the subjects are shown in **Figure 10**.

A confusion matrix is illustrated in **Table 3** where it is evident that the wrist pronation and supination cause the most confusion. From time to time the subject executes the actions and sometimes put a break to them. In so doing the muscle intention power descend below the threshold and is not detected, and subsequently tagged as uncounted or undetected.

In common practices, in addition to muscular fatigue, mental fatigue could likewise show up. This would affect the unwavering quality of the fNIRS signal. Frequent use of nicotine substances involving tea or coffee and weak eyesight are



**Figure 10.** A representation of subject-wise accuracies. Healthy subjects are presented as S1–S15 whereas amputees are labeled A1–A3. All concerning classifying technique as LDA and ANN.

Output class	1	2	3	4	5	6
Target class						
1	5010 100.0%	0 0.0%	0 0.0%	0 0.0%	0 0.0%	0 0.0%
2	0 0.0%	4978 99.3%	12 0.3%	0 0.0%	0 0.0%	0 0.0%
3	0 0.0%	0 0.0%	3648 73%	720 14.4%	0 0.0%	0 0.0%
4	0 0.0%	20 0.4%	630 12.6%	3617 72.1%	0 0.0%	0 0.0%
5	0 0.0%	0 0.0%	0 0.0%	0 0.0%	5001 99.8%	0 0.0%
6	0 0.0%	0 0.0%	0 0.0%	23 0.5%	9 0.2%	4978 99.3%

**Table 3.** Confusion matrix of motion prediction.

accounted for in such reliability. The motion prediction accuracy of an individual can be changed under the influence of such conditions.

5.3 Control command generation

After the results from the classifier are returned, they are translated into a control command. The motions are assigned abbreviations identical as used in Section 2. They are listed here as in **Table 4**.

While testing, the machine responds as the variables illustrated in **Table 4** that are assigned to each class motion. The controllers further single-out one definitive motion which can be analyzed while testing the data.

These control commands can be further translated to motor action via a controller such as Arduino, Raspberry Pi, Odroid, etc. as the program routine is written in a language supported by all these controllers. A three degree of freedom device can be actuated using this neural-machine interface scheme.

To eliminate the channel selection part, manufacturers are working on bundled optodes. Using these bundled optodes will not only eliminate the channel selection

Motion	Notation
Extension of elbow	EE
Flexion of elbow	EF
Pronation of wrist	WP
Supination of wrist	WS
Flexion of wrist	WF

**Table 4.**  
*Class labels.*

complications but it will also help in physical attributes as the whole head will not be covered. It will be easy to wear the head cap because of a smaller number of optodes and wires going around.

Also, using a different feature set may aid in the increase in accuracy. Rather than extracting three or four features, only one optimal feature can be evaluated. Hybridization of bio-signals can be done using advanced probability models or neural networks can be trained and implemented to hybridize different modalities.

## 6. Conclusion

fNIRS signals were acquired using the NIRSport machine developed by NIRx technology. These signals were recorded for six motions i.e. elbow extension (EE), elbow flexion (EF), wrist supination (WS), wrist pronation (WP), wrist extension (WE) and wrist flexion (WF) and were further analyzed. Mean and peak feature was extracted from the hemodynamic response of the brain. Also, minimum values were extracted for channel selection. The hemodynamic responses acquired from the brain were trained and tested by two widely used classifiers in pattern recognition i.e. LDA and ANN. The highest value of accuracy for an individual subject was recorded at 85% which is not yet achieved with six control commands employed by fNIRS. Both the classifiers were also active for real-time analysis. As a result of such high value of training accuracy, 8 out of 10 motions were correctly predicted in real-time setting. Possible extension of this work could be to hybridize these fNIRS signals together with another signal modality to not only increase the accuracy but also the number of control commands. Arm movement pattern for different age groups can be further explored. The number of amputated subjects could be increased to acquire data which will aid in better understanding of hemodynamic behavior of human brain and how it can be used to predict the arm motions.

## Acknowledgements

We would like to mention the funding body, i.e. Higher Education Commission (HEC) of Pakistan who awarded the grant under NRPU project number 10702. We would also like to show gratitude to the friends who connected us with the amputees.

IntechOpen

IntechOpen

### **Author details**

Usama Ali Syed\*, Zareena Kausar and Neelum Yousaf Sattar  
Air University, Islamabad, Pakistan

\*Address all correspondence to: syed.a.usama@outlook.com

### **IntechOpen**

© 2020 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/3.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited. 

## References

- [1] Lobo-Prat J, Kooren PN, Stienen AH, Herder JL, Koopman BF, Veltink PH. Non-invasive control interfaces for intention detection in active movement-assistive devices. *Journal of Neuroengineering and Rehabilitation*. 2014;**11**(1):168
- [2] Fang Y, Hettiarachchi N, Zhou D, Liu H. Multi-modal sensing techniques for interfacing hand prostheses: A review. *IEEE Sensors Journal*. 2015;**15**(11):6065-6076
- [3] Guo W, Sheng X, Liu H, Zhu X. Mechanomyography assisted myoelectric sensing for upper-extremity prostheses: A hybrid approach. *IEEE Sensors Journal*. 2017;**17**(10):3100-3108
- [4] Guo W, Sheng X, Liu H, Zhu X. Toward an enhanced human-machine interface for upper-limb prosthesis control with combined EMG and NIRS signals. *IEEE Transactions on Human-Machine Systems*. 2017;**47**(4):564-575
- [5] Reaz MB, Hussain MS, Mohd-Yasin F. Techniques of EMG signal analysis: Detection, processing, classification and applications (correction). *Biological Procedures Online*. 2006;**8**(1):163
- [6] Batula AM, Mark J, Kim YE, Ayaz H, editors. Developing an optical brain-computer interface for humanoid robot control. In: *International Conference on Augmented Cognition*. Springer; 2016
- [7] Coyle SM, Ward TE, Markham CM. Brain-computer interface using a simplified functional near-infrared spectroscopy system. *Journal of Neural Engineering*. 2007;**4**(3):219
- [8] Sitaram R, Zhang H, Guan C, Thulasidas M, Hoshi Y, Ishikawa A, et al. Temporal classification of multichannel near-infrared spectroscopy signals of motor imagery for developing a brain-computer interface. *NeuroImage*. 2007;**34**(4):1416-1427
- [9] Lambert J-H. *Photometria, sive de Mensura et gradibus luminis, colorum et umbrae*. sumptibus viduae E. Klett; 1760
- [10] Abitan H, Bohr H, Buchhave P. Correction to the Beer-Lambert-Bouguer law for optical absorption. *Applied Optics*. 2008;**47**(29):5354-5357
- [11] Beer A. Bestimmung der absorption des rothen lichts in farbigen flussigkeiten. *Annalen der Physik*. 1852;**162**:78-88
- [12] Lambert JH. *Lamberts Photometrie: Photometria, sive De mensura et gradibus luminis, colorum et umbrae* (1760). W. Engelmann; 1892
- [13] Herold F, Wiegel P, Scholkmann F, Müller NG. Applications of functional near-infrared spectroscopy (fNIRS) neuroimaging in exercise-cognition science: A systematic, methodology-focused review. *Journal of Clinical Medicine*. 2018;**7**(12):466
- [14] Aresté N, Salgueira M. World Medical Association Declaration of Helsinki: Ethical principles for medical research involving human subjects. *Journal of the American Medical Association*. 2013;**310**(20):2191-2194
- [15] Yanagisawa H, Dan I, Tsuzuki D, Kato M, Okamoto M, Kyutoku Y, et al. Acute moderate exercise elicits increased dorsolateral prefrontal activation and improves cognitive performance with Stroop test. *NeuroImage*. 2010;**50**(4):1702-1710
- [16] Hyodo K, Dan I, Suwabe K, Kyutoku Y, Yamada Y, Akahori M, et al. Acute moderate exercise enhances compensatory brain activation in older adults. *Neurobiology of Aging*. 2012;**33**(11):2621-2632
- [17] Byun K, Hyodo K, Suwabe K, Ochi G, Sakairi Y, Kato M, et al. Positive



effect of acute mild exercise on executive function via arousal-related prefrontal activations: An fNIRS study. *NeuroImage*. 2014;**98**:336-345

[18] Dupuy O, Gauthier CJ, Fraser SA, Desjardins-Crepeau L, Desjardins M, Mekary S, et al. Higher levels of cardiovascular fitness are associated with better executive function and prefrontal oxygenation in younger and older women. *Frontiers in Human Neuroscience*. 2015;**9**:66

[19] Hyodo K, Dan I, Kyutoku Y, Suwabe K, Byun K, Ochi G, et al. The association between aerobic fitness and cognitive function in older men mediated by frontal lateralization. *NeuroImage*. 2016;**125**:291-300

[20] Albinet CT, Mandrick K, Bernard PL, Perrey S, Blain H. Improved cerebral oxygenation response and executive performance as a function of cardiorespiratory fitness in older women: A fNIRS study. *Frontiers in Aging Neuroscience*. 2014;**6**:272

[21] Hyodo K, Suwabe K, Soya H, Nagamatsu T. The effect of an acute bout of slow aerobic dance on mood and executive function in older adults: A pilot study. *Bulletin of the Physical Fitness Research Institute*. 2017;**115**:35-41

[22] Wang W, Qiu C, Ota T, Sawada M, Kishimoto N, Kishimoto T. Effects of Tai Chi exercise on attention in healthy elderly subjects as measured by near-infrared spectroscopy during the Stroop task. *Journal of Nara Medical Association*. 2013;**64**(5):79-86

[23] Faulkner J, Stoner L, Grigg R, Fryer S, Stone K, Lambrick D. Acute effects of exercise posture on executive function in transient ischemic attack patients. *Psychophysiology*. 2017;**54**(8):1239-1248

[24] Moriya M, Aoki C, Sakatani K. Effects of physical exercise on working

memory and prefrontal cortex function in post-stroke patients. In: *Oxygen Transport to Tissue XXXVIII*. Springer; 2016. pp. 203-208

[25] Tsujii T, Komatsu K, Sakatani K. Acute effects of physical exercise on prefrontal cortex activity in older adults: A functional near-infrared spectroscopy study. In: *Oxygen Transport to Tissue XXXIV*. Springer; 2013. pp. 293-298

[26] Liao LD, Tsytsarev V, Delgado-Martinez I, Li ML, Erzurumlu R, Vipin A, et al. Neurovascular coupling: In vivo optical techniques for functional brain imaging. *Biomedical Engineering Online*. 2013;**12**(1):38

[27] Chen T, Yue GH, Tian Y, Jiang C. Baduanjin mind-body intervention improves the executive control function. *Frontiers in Psychology*. 2016;**7**:2015

[28] Mücke M, Andrä C, Gerber M, Pühse U, Ludyga S. Moderate-to-vigorous physical activity, executive functions and prefrontal brain oxygenation in children: A functional near-infrared spectroscopy study. *Journal of Sports Sciences*. 2018;**36**(6):630-636

[29] Kato K, Iwamoto K, Kawano N, Noda Y, Ozaki N, Noda A. Differential effects of physical activity and sleep duration on cognitive function in young adults. *Journal of Sport and Health Science*. 2018;**7**(2):227-236

[30] Sudo M, Komiyama T, Aoyagi R, Nagamatsu T, Higaki Y, Ando S. Executive function after exhaustive exercise. *European Journal of Applied Physiology*. 2017;**117**(10):2029-2038

[31] Matsuda K, Ikeda S, Mitsutake T, Nakahara M, Nagai Y, Ikeda T, et al. Factors influencing executive function by physical activity level among young adults: A near-infrared spectroscopy study. *Journal of Physical Therapy Science*. 2017;**29**(3):470-475

- [32] Makizako H, Doi T, Shimada H, Park H, Uemura K, Yoshida D, et al. Relationship between going outdoors daily and activation of the prefrontal cortex during verbal fluency tasks (VFTs) among older adults: A near-infrared spectroscopy study. *Archives of Gerontology and Geriatrics*. 2013;**56**(1):118-123
- [33] Kujach S, Byun K, Hyodo K, Suwabe K, Fukuie T, Laskowski R, et al. A transferable high-intensity intermittent exercise improves executive performance in association with dorsolateral prefrontal activation in young adults. *NeuroImage*. 2018;**169**:117-125
- [34] Giles GE, Cantelon JA, Eddy MD, Brunyé TT, Urry HL, Mahoney CR, et al. Habitual exercise is associated with cognitive control and cognitive reappraisal success. *Experimental Brain Research*. 2017;**235**(12):3785-3797
- [35] Yamazaki Y, Sato D, Yamashiro K, Tsubaki A, Yamaguchi Y, Takehara N, et al. Inter-individual differences in exercise-induced spatial working memory improvement: A near-infrared spectroscopy study. In: *Oxygen Transport to Tissue XXXIX*. Springer; 2017. pp. 81-88
- [36] Lambrick D, Stoner L, Grigg R, Faulkner J. Effects of continuous and intermittent exercise on executive function in children aged 8-10 years. *Psychophysiology*. 2016;**53**(9):1335-1342
- [37] Quaresima V, Ferrari M. Functional near-infrared spectroscopy (fNIRS) for assessing cerebral cortex function during human behavior in natural/ social situations: A concise review. *Organizational Research Methods*. 2019;**22**(1):46-68
- [38] Perrey S. Non-invasive NIR spectroscopy of human brain function during exercise. *Methods*. 2008;**45**(4):289-299
- [39] Pinti P, Aichelburg C, Gilbert S, Hamilton A, Hirsch J, Burgess P, et al. A review on the use of wearable functional near-infrared spectroscopy in naturalistic environments. *Japanese Psychological Research*. 2018;**60**(4):347-373
- [40] Holtzer R, Mahoney JR, Izzetoglu M, Wang C, England S, Verghese J. Online fronto-cortical control of simple and attention-demanding locomotion in humans. *NeuroImage*. 2015;**112**:152-159
- [41] Holtzer R, Verghese J, Allali G, Izzetoglu M, Wang C, Mahoney JR. Neurological gait abnormalities moderate the functional brain signature of the posture first hypothesis. *Brain Topography*. 2016;**29**(2):334-343
- [42] Cui X, Bray S, Reiss AL. Speeded near infrared spectroscopy (NIRS) response detection. *PLoS One*. 2010;**5**(11):e15474
- [43] Strangman G, Boas DA, Sutton JP. Non-invasive neuroimaging using near-infrared light. *Biological Psychiatry*. 2002;**52**(7):679-693
- [44] Strangman G, Franceschini MA, Boas DA. Factors affecting the accuracy of near-infrared spectroscopy concentration calculations for focal changes in oxygenation parameters. *NeuroImage*. 2003;**18**(4):865-879
- [45] Scholkmann F, Spichtig S, Muehlemann T, Wolf M. How to detect and reduce movement artifacts in near-infrared imaging using moving standard deviation and spline interpolation. *Physiological Measurement*. 2010;**31**(5):649
- [46] Scholkmann F, Wolf M. Measuring brain activity using functional near infrared spectroscopy: A short review. *Spectroscopy Europe*. 2012;**24**(4):6
- [47] Ando S, Yamada Y, Kokubu M. Reaction time to peripheral visual

stimuli during exercise under hypoxia.  
*Journal of Applied Physiology*.  
2010;**108**(5):1210-1216

[48] Villringer A, Chance B. Non-invasive optical spectroscopy and imaging of human brain function. *Trends in Neurosciences*. 1997;**20**(10):435-442

[49] Issard C, Gervain J. Variability of the hemodynamic response in infants: Influence of experimental design and stimulus complexity. *Developmental Cognitive Neuroscience*. 2018;**33**:182-193

[50] Demandt E, Mehring C, Vogt K, Schulze-Bonhage A, Aertsen A, Ball T. Reaching movement onset-and end-related characteristics of EEG spectral power modulations. *Frontiers in Neuroscience*. 2012;**6**:65

[51] Herold F, Wiegel P, Scholkmann F, Thiers A, Hamacher D, Schega L. Functional near-infrared spectroscopy in movement science: A systematic review on cortical activity in postural and walking tasks. *Neurophotonics*. 2017;**4**(4):041403