We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,900
Open access books available

185,000
International authors and editors

200M
Downloads

Our authors are among the

154
Countries delivered to

TOP 1%
most cited scientists

12.2%
Contributors from top 500 universities



WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com



Chapter

A Brief Summary of EEG Artifact Handling

İbrahim Kaya

Abstract

There are various obstacles in the way of use of EEG. Among these, the major obstacles are the artifacts. While some artifacts are avoidable, due to the nature of the EEG techniques there are inevitable artifacts as well. Artifacts can be categorized as internal/physiological or external/non-physiological. The most common internal artifacts are ocular or muscular origins. Internal artifacts are difficult to detect and remove, because they contain signal information as well. For both resting state EEG and ERP studies, artifact handling needs to be carefully carried out in order to retain the maximal signal. Therefore, an effective management of these inevitable artifacts is critical for the EEG based researches. Many researchers from various fields studied this challenging phenomenon and came up with some solutions. However, the developed methods are not well known by the real practitioners of EEG as a tool because of their limited knowledge about these engineering approaches. They still use the traditional visual inspection of the EEG. This work aims to inform the researchers working in the field of EEG about the artifacts and artifact management options available in order to increase the awareness of the available tools such as EEG preprocessing pipelines.

Keywords: Artifact, Artifact removal methods, EEG, EEG preprocessing, Muscular artifacts, Ocular artifacts, Preprocessing pipelines

1. Introduction

A signal is a function that conveys information about the behavior or attributes of some phenomenon [1]. On the other hand, information can be anything. A waveform can have multiple overlapping information in the same space–time. The signal in a waveform is subjective, it can be color for one and shape for the other. In electrophysiology, waveform under inspection can be separated into two as the signal of interest and noise. The signal can be electrocardiography (ECG), Electroencephalogram (EEG), or any other physiological signal, noise is any unwanted wave source ınterfering with the signal. If we consider EEG as the signal, it is recorded from the scalp by electrodes and consists of the overall electrical activities of neural populations and a contribution of glial cells [2]. EEG has a wide range of use in both clinical practice and engineering applications in medicine, particularly neurology, sleep, and epilepsy research.

2. Background

The EEG recording environment and subject related electrical activities during recording deteriorate the signal quality. Artifacts are undesired signals that may introduce changes in the measurements and affect the signal of interest [3]. EEG can be contaminated in frequency or time domain by artifacts that are resulted from internal sources of physiologic activities and movement of the subject and/or external sources of environmental interferences, equipment, movement of electrodes and cables [4]. Artifact types and sources are listed in the **Table 1**. External artifacts can be prevented by proper shielding, grounding cables, isolating and moving cables away from recording sites since they act as antennas during operation. On the other hand, internal or physiological artifacts are challenging for researchers because of their inclusion of signal or resemblance to the signals. The most important artifacts in a typical EEG recording are ocular electro-oculogram (EOG) artifacts and muscular (EMG) artifacts.

2.1 Ocular artifacts

Electrical potentials due to eye opening/closure, blinks, eyelid flutter and eye movements propagate over the scalp and produce hostile EOG artifacts in the

Artifact	Type	Source
Eye blink	Ocular	Internal/Physiological
Eye movement	Ocular	Internal/Physiological
REM Sleep	Ocular	Internal/Physiological
Scalp contractions	Muscle	Internal/Physiological
Glossokinetic artifact	Muscle	Internal/Physiological
Chewing	Muscle	Internal/Physiological
Talking	Muscle	Internal/Physiological
EKG	Cardiac	Internal/Physiological
Swallowing	Muscle	Internal/Physiological
Respiration	Respiratory	Internal/Physiological
Galvanic Skin Response	Skin	Internal/Physiological
Sweating	Skin	Internal/Physiological
Electrode movement	Instrumental	External/Extra-physiological
Electrode Impedence Imbalance	Instrumental	External/Extra-physiological
Cable movement	Instrumental	External/Extra-physiological
Electromagnetic coupling	Electromagnetic	External/Extra-physiological
Powerline	Electrical	External/Extra-physiological
Head movement	Movement	External/Extra-physiological
Body movement	Movement	External/Extra-physiological
Limbs movement	Movement	External/Extra-physiological

Table 1.
EEG artifact types and sources. Adapted from [4, 5].

recorded EEG. Eye movements are major sources of contamination of EEG. The origin of this contamination is disputable. Cornea-retinal dipole movement, retinal dipole movement and eyelid movement are the three main proposed causes of the eye movement related voltage potential [6]. The direction of eye movements affects the shape of the EOG waveform while a square-like EOG wave is produced by vertical eye movements and blinks which leads to a spike-shaped waveform [7]. Blinks which are attributable to the eyelid moving over the cornea, occurring at intervals of 1-10s, generate a characteristic brief potential of between 0.2 s and 0.4 s duration due to eyelid movement over cornea [8, 9]. The blinking artifact generally has an amplitude much larger than that of the background EEG [6]. It is advantageous to have a reference EOG channel during EEG recording for the cancellation of ocular artifact from EEG activity [3].

2.2 Muscular artifacts

Electrical activity on the body surface due to the contracting muscles are recorded via Electromyogram (EMG) [3]. Since independent myogenic activities of head, face and neck muscles are conducted through the entire scalp, it can be monitored in the EEG [10, 11]. The amplitude of this type of artifact is dependent on the type of muscle and the degree of tension [3, 12]. The frequency range of EMG activity is wide, being maximal at frequencies higher than 30 Hz [13, 14].

2.3 Cardiac artifacts

The electrical potential due to cardiac activity can exhibit itself in the EEG as ECG artifacts. Typical high frequency waveforms similar to EKG P-QRS-T shape are characteristics of EKG artifacts in EEG [15].

2.4 Other artifacts

Head, body and limb movements cause irregular high voltage artifacts. Artifacts can be produced by tremors in patients such as Parkinson disease and movement disorders. Changing patient position into a calm comfortable stable position helps reducing artifacts. Another prevention for respiratory related movement artifacts is to use a towel or a firm material support for the neck. The changes in the impedance or electrical potential between scalp and electrode may cause electrode artifacts. These can result from poor electrode contact, broken lead, electrolyte gel insufficiency. This type of artifact usually exhibits itself in sudden electrode pops. These electrode artifacts can be eliminated by using proper electrolyte gel, checking electrode impedance, changing the broken electrodes, and shifting the electrode position slightly.

3. Artifact handling methods

A typical EEG recording system is shown in **Figure 1**. At the heart of a recording setup is the biopotential amplifier. It should have high common mode rejection ratios, however it should not have high gains, this can saturate the signal due to large half-cell potentials at the electrodes. Unequal electrode impedances are major sources of common mode artifacts such as powerline.




Figure 1.
EEG recording system and experiment setup.

Environmental artifacts can be eliminated by bringing the electrodes leads closer together, moving the electrodes and subject away from the noise sources, using single isolated earth for the whole setup, and shielding the cables, machines and artifact sources with a metal tape connected to the common earth. Moreover, the environmental conditions should satisfy the following requirements for proper recordings. These can be listed as, quiet atmosphere, comfortable temperature and humidity, controlled proper lighting, using a comfortable bed or chair, and separating the powerline of the EEG system from the other machines in the lab.

3.1 Averaging methods to suppress ERP artifacts

Event Related Potentials (ERP) are electrical signals generated in response to internal or external events and they are recorded by EEG [16]. In evoked potentials, each stimulus produces an evoked potential embedded in EEG. However, since the ERP or evoked potential signals are generally subtle in EEG, averaging of many epochs are needed to make them distinguishable. An ensemble averaging method to enhance the ERPs was defined by [17]. This relies on the assumption that by synchronous averaging of each epoch, signal ERP amplitude adds constructively and EEG background noise diminishes destructively.

In ERP and evoked potential research, artifacts contaminate the final ensemble average signal of interest. One method to overcome this adverse effect is to benefit from a weighted averaging [18]. In weighted averaging technique each epoch is weighted inversely with the non-stationary noise maximum amplitude in the epoch. In [19], each trial's contribution to ensemble average is multiplied by a weight according to its correlation with the rest of the data. This factor is inversely related to its probability of being an artifact. For example, a large amplitude EEG is likely to be an artifact and the contribution factor for the trial involving large amplitudes will be low whereas the factor for a small amplitude EEG is high (**Figure 2**). Davila and Mobin [20] showed that weighted averaging of auditory EP has higher SNR than conventional ensemble averaging. John et al. [21] studied the effects of such techniques as sample-weighted averaging, noise-weighted averaging, amplitude based artifact rejection, percentage based artifact rejection, and normal averaging on the steady state auditory evoked potentials. It concluded in favor of weighted averaging for better SNR of steady state responses. On the other hand, according to [22], weighted averaging underestimates the ERP signal amplitude. Determination of the optimal weighting factor is not straightforward and this limits the performance of




Figure 2.
Various EEG artifacts are shown.

the weighting averaging method. Mühler and Specht [23] developed a method called 'sorted averaging'. In sorted averaging, epochs are sorted with RMS values from small to large, since noisy artifactual epochs have large RMS values compared to low noise signals. The signal averaging is performed by addition of epochs from the low noise RMS to large RMS sorted order until a maximum peak of SNR^2 is obtained [24]. This eliminates the high RMS noisy epochs and yields a better ERP waveform. Compared to weighted averaging, sorted averaging had significantly higher SNR^2 [23].

Median averaging is another approach to ERP artifact handling and it is based on taking the median points of all the epochs and adding them to form a median average instead of classic mean average [25]. Some advantages of the median averaging are that; it elicits hidden signals more clearly and it is not affected by infrequent large artifacts that much compared to mean averaging [25]. Özdamar and Kalayci [26] supported the advantages of median averaging over the conventional mean averaging in a study on the ABR signals. Median averaging is an efficient way to remove adverse effects of the outliers on the final averaged signal, yet it also removes the valuable data in the outliers causing significant loss of information [27, 28].

3.2 Artifact handling methods for EEG

Artifact avoidance, artifact rejection, manual rejection, automatic rejection, and artifact removal are the common methods to deal with artifacts [29]. Although it seems a simple solution to cancel EOG and EMG artifacts by instructing subject to avoid blinking or movement, it can result in change of amplitudes in evoked potentials as well as the additional cognitive load [29–31]. On the other hand, artifact rejection or manual rejection may require a person dedicated to this purpose of eliminating artifacts visually one by one in an EEG. Moreover, the artifact detection by an expert may be subjective, tedious, and time consuming. In addition, it can not be applicable to online removal [3]. However, automatic rejection can automate this artifact rejection procedure but it can eliminate non-artifact signals if not properly tuned. The automatic rejection of artifact containing EEG can depend on artifact amplitude based or EEG segment RMS based artifact detection and rejection. An example of a

simple blink artifact removal is depicted in **Figure 3**. Since blinks have low frequency content compared to EEG, by low pass filtering, EEG can be reduced while blink artifact still remains at a high voltage level. Thus, an amplitude threshold based artifact rejection can be applied. As seen from **Figure 3**, red traces are the EEG and blue are the low pass filtered EEG signal. While a simple artifact rejection (without low pass filtering) using a threshold of 20 μV will produce false positives (red traces over 20 μV), in the low pass filtered EEG these false positives are prevented.

Usually one or two channels are dedicated to detect EOG artifacts. There are two widely used procedures for EOG artifacts, first EOG rejection where EEG trials with EOG artifacts having VEOG greater than a preset threshold are omitted, and second EOG correction where the effect of eye movement is tried to be removed from EEG [6].

Artifacts can distort the EEG in a way that the electrophysiologists or physicians can be misled in their clinical interpretation [32]. This makes artifact removal critical in the pre-processing phase prior to analysis. There are many methods to remove artifacts such as Artifactual Segment Rejection, Filtering, Wiener filtering, Adaptive Filtering, Time-Frequency Representation, Wavelet Transform, Discrete Wavelet Transform (DWT), Adaptive Noise Cancelation (ANC), Wavelet Packet Transform (WPT), Kalman Filtering, Linear Regression, Blind Source Separation (Principal Component Analysis (PCA), Independent Component Analysis (ICA), Canonical Correlation Analysis (CCA), Minor Components Analysis (MCA)), Source Decomposition, Empirical Mode Decomposition (EMD), Support Vector Machine (SVM), and hybrid methods [3, 4, 29, 33–38]. A functional dedicated artifact channel which provides complementary aid to identify ECG/EOG is required to remove ocular or cardiac artifacts in the most of the available methods [4].

Regression is a common and well established technique in artifact removal, yet it cannot be used to remove muscle noise or line noise, since these type of artifacts have no reference channels [39]. Having a good regressor (e.g., an EOG) is critical in both time and frequency domain regression methods. It is an inherent weakness that eye movements and EEG signals are bidirectional. When unacceptable amount of data are lost in artifact rejection, delicate artifact removal methods which will preserve




Figure 3.
Low pass filtering based EEG blink rejection. Red is raw EEG, blue is low pass filered EEG with 6th order Butteworth low pass filter at 8 Hz cut off. The detected artifact containing EEG epochs are shown in dashed rectangles.

the essential EEG signals while removing artifacts are necessary [39]. One of the most important artifacts is EOG. EEG regions infected with EOG can be rejected from overall EEG signal with simplest artifact rejection where these portions are detected by EOG channels, however these regions still carry brain signals in addition to ocular artifacts and total rejection or subtraction of EOG from them results in loss of brain data [40–42].

Blind Source Separation (BSS) algorithms utilize multiple channels in an unsupervised learning algorithm to extract brain related activity from the ensemble EEG signal which can be assumed a linear superposition of brain signals, noise and artifacts [38]. Three common BSS algorithms are Independent Component Analysis (ICA), Principal Component Analysis (PCA) and Canonical Correlation Analysis (CCA).

ICA, a BSS method, is often used to remove EEG artifacts based on statistical approach of spatial filtering and separation of multiple channel EEG data into spatially fixed and temporally independent components [39, 43, 44]. Since the EEG sources and artifacts are usually of different origins, they can be assumed to be linear summation of each independent components. ICA method finds these statistically independent components and enable us to eliminate artifactual ones from the desired EEG [45]. On the other hand, ICA provides extraction of the eye related signals present in the EOG, and removal of this information or artifact, rather than the complete EOG which still has some brain activity [40], is possible. However, detection and removal of transient artifacts such as head and neck muscle contractions and movement are difficult with ICA [46]. Moreover, adapting ICA as an online method requires high computational power [46]. On the other hand, an advantage of ICA is that it does not rely on a reference channel [39]. However, many artifact removal algorithms are compared in [3], and Revised Aligned-Artifact Average (RAAA) and Second Order Blind Identification (SOBI) and Adaptive Mixture of Independent Component Analyzers (AMICA) are the preferred artifact removal methods for EOG, EMG and ECG artifacts.

PCA uses orthogonal transform of correlated time domain signal into linearly uncorrelated principal components (PCs) [47]. These principal components possess as much as variance of the EEG as possible. Artifact containing PCs can be eliminated if they are uncorrelated with the brain EEG. Application of PCA into ocular artifacts was provided in [48].

CCA is also another method utilized in removing artifacts. In CCA second order statistics are employed, correlation between two multivariate datasets are maximized by canonical variables. CCA offers shorter computational time compared to ICA [38].

Another method is filtering in frequency domain. Usually a high-pass filter starting from 0.5-1 Hz is applied for baseline drift removal. Notch filters are used to remove powerline-noise. Another one, EMG activity of contracting scalp sites can hinder the signals of interest in the EEG recordings during an epileptic seizure [49]. It was possible to remove this high frequency content EMG activity from EEG spectra by filtering out signals over 25 Hz. Adaptive Filters, Wiener Filtering and Bayesian Filters are three filtering methods applied in EEG signal preprocessing. Adaptive Filters are the most commonly used for artifact removal [47]. In Adaptive Filtering a reference channel for artifacts is subtracted from the EEG recursively. This reference is multiplied by a weight factor obtained from the output of the filter by a learning algorithm and this weighted reference is subtracted from the recorded EEG yielding output artifact free EEG changing adaptively [50].

In wavelet transform, many scaled and time shifted wavelets are used to produce coefficients for the particular signal and wavelet type by convolution of the signal and

wavelets. These coefficients indicate similarity between the corresponding wavelet and the signal. In artifact removal via wavelet transform, the main idea is that the signal which can be highly correlated with a basis mother wavelet and can be separated from artifacts which might have no correlation to the principal mother wavelet [50]. Some examples of Wavelet Transform in artifact removal are for ocular artifact removal as in [51, 52].

3.3 EEG pre-processing pipelines available

Recently many preprocessing pipelines have been introduced in order to reduce the burden of artifact handling by an expert one by one visual inspection. This laborious task can be fastened by using existing automatized preprocessing methods in order. An efficient pre-processing pipeline not only helps the artifact management time but also provides objective evaluation with predefined criteria compared to highly subjective artifact handling by a human expert. The preprocessing pipelines usually consist of the combination of the following stages; filtering, re-referencing, bad channel identification (and interpolation), bad channel and epoch removal, artifact detection using ICA, artifact correction and removal [53], see **Figure 4**.

Fully Automated Statistical Thresholding for EEG artifact Rejection (FASTER) [54] algorithm is a state of the art method which is available in EEGLAB toolbox [55]. FASTER has filtering, line noise removal, bad channel detection and interpolation, segmentation, and artifact rejection on segments by identifying bad channels, blinks, eye movements and muscular artifacts using combination of statistical thresholding and ICA [56]. It requires an extra EOG channel. The Automatic Pre-processing Pipeline (APP) removes powerline noise, bad channels, eye movements, blinks and muscular artifacts using ICA to identify artifactual components [53], see **Figure 4**. However, it also requires extra EOG channels. Da Cruz et al. [53] has found that APP performs better than FASTER yielding higher amplitude in ERP study. Another pipeline is Tool for Automated Processing of EEG data (TAPEEG) [57]. It uses automated routines of FASTER and Fieldtrip for artifact identification and performed similar to visually analysis by an expert [58]. TAPEEG handles the resting state EEG data as well. Both FASTER and TAPEEG are based on z- scores and have difficulty in handling outliers, this leads to loss of signal content due to false positive artifact detection and rejections [53]. Another standardized preprocessing method for large EEG datasets, PREP pipeline, handles line noise removal, bad channel detection, and referencing to standardize and normalize the data before processing [58]. It is also available as plug-in in EEGLAB toolbox.

Automagic is a toolbox developed for standardized handling of large growing EEG/ERP datasets by time [56]. The power of Automagic comes from the fact that it exploits many existing pipelines and methods, such as PREP pipeline for bad channel identification and for average referencing, Cleanline [59] to remove power line noise, EOG regression [60], Multiple Artifact Rejection Algorithm (MARA), ICA or robust PCA for artifact correction [61]. MARA is a plug-in available in EEGLAB which automatically identifies artifacts not only ocular or muscular but also any general artifactual source component in ICA [61]. Pedroni et al. [59] showed that combination of a preprocessing pipeline to identify bad channels and MARA method is efficient to remove most of the artifacts.

None of the methods offers a perfect robust and high accurate management of all types of artifacts. In general, they are all limited with the training dataset and fail to achieve high success with new type of artifactual data.




Figure 4.
APP artifact management flow diagram from [53].

3.4 Simultaneous EEG and f-MRI artifact handling

Since EEG is widely used as a clinical tool to monitor or diagnose patients, doctors can be misguided in case of artifacts and EEG can be misinterpreted. For this reason, artifact removal becomes a crucial point for some cases such as epilepsy monitoring in an EEG/fMRI recording room. Today EEG and fMRI are two distinct but closely related and complementary methods. While fMRI provides high spatial resolution for localization of phenomena in the brain, EEG on the other hand results in better temporal resolution [62–65]. One should be careful about the experiments involving both fMRI and EEG because there are many unwanted electromagnetic sources interfering with EEG. For example, the false identification of spikes are highly possible since residuals of Ballistocardiogram (BCG) artifacts have similar shapes as epileptic spikes [66]. The factors that can lead to differences in the artifact are linked

to the subject and experimental setup, [67]. There are imaging artifacts, cardiac related Ballistocardiogram artifacts (BCG), EOG and EMG artifacts in an EEG inside MRI [44]. Static field (B0) and the time-varying fields of radio-frequency excitations and of imaging gradients, generate artifacts in the EEG known as Ballistocardiogram (BCG) and imaging artifacts [44, 68–70]. The pulse artifact which can be observed in EEGs recorded inside MR scanners easily, is due to a fundamental cause that any movement of electrically conductive muscles in a static magnetic field generates electromagnetic induction and it is proportional to the static field, generally larger at higher field strengths [67, 71]. Pulsations of the scalp arteries are the main cause of this type of BCG artifact [72, 73]. The study of Grouiller et al. [44] compared different imaging artifact removal techniques and various cardiac artifact correction techniques in both simulated EEG data and in real experimental data. They concluded that there is no key for every door, some algorithms work well for some case and others might work well for other cases. Certain algorithms may be preferred depending on the type of data and analysis method [44]. Another algorithm, adaptive Optimal Basis Set (aOBS), automatically eliminates BCG artifacts yet preserving the neural origin signals in EEG [74]. It can be used efficiently for simultaneous fMRI and EEG recordings.

3.5 Sleep stage classification artifact handling

Manual artifact detection is still the most common method for artifact handling for sleep stage classification, however, the long time required and the difficulty to apply it to large datasets poses the main disadvantages [75]. Malafeev et al. [75] compared 12 simple algorithms that are applicable with a single EEG channel for ease of use. It was found that automatic artifact detection in EEG during sleep within large datasets is possible with simple algorithms. Among these, Power thresholding 25–90 Hz (PT25), Power thresholding 45–90 Hz (PT45) and Autoregressive (AR) models had Reciever Operating Characteristic (ROC) areas above 0.95. In addition, online detection is also possible with the majority of these simple algorithms.

3.6 BCI Artifact handling

Artifact removal in BCI applications are getting more attention. By studies it was shown that artifacts generated by EOG and EMG activities affect the neurological signals utilized in a BCI system [10, 76]. Although there are extensive researches into artifact removal for BCIs and developed efficient methods such as Fully Online and Automated Artifact Removal (FORCe), Lagged Auto-Manual Information Clustering (LAMIC), Fully Automated Statistical Thresholding for EEG artifact Rejection (FASTER) and K-Singular Value Decomposition (K-SVD), the field lacks an effective artifact removal [12, 54, 77–82]. The surrogate-based artifact removal (SuBAR) technique proposed by Chavez et al. [33] effectively cancels EOG and EMG artifacts from single-channel EEG. Chang et al. [83] proposed a method for detection of eye artifact from single prefrontal channel which is useful for headband-type wearable EEG devices with a few frontal EEG channels. Compared to conventional methods the accuracy of detecting ocular artifact contaminated epochs was significantly better. Daily-life EEG-BCIs are getting popular and artifact removal techniques for these BCIs must have some critical features such as; must be performed outdoor, with portable wearable wireless device, with real EEG signals, compatible with daily life tasks, must have simple electrical montage, must use dry electrodes, must remove complex

artifacts, must work only EEG without reference, must work online and must work with single electrode channel. More research into artifact removal other than ocular and cardiac artifacts is necessary especially for those daily-life EEG BCIs [36].

While ICA and PCA are common artifact removal methods, Artifact Subspace Reconstruction (ASR), which is a powerful automated artifact removal method available for both online real-time and offline, can be applied to prevent transient and large artifact [46, 84]. It also does not require additional channel and cleans the data from artifacts.

4. Conclusion

The number of artifact handling techniques and algorithms are increasing drastically, however the artifact problem is still challenging for many applications. Particularly, the internal or physiologic artifacts are difficult to distinguish and remove. While simple measures such as artifact avoidance and artifact rejection can be utilized in some applications, most of the cases require special methods dedicated to handle artifacts in order to significantly reduce their harmful effects on signal of interest. Due to the varying nature of artifacts a generic method for all sorts of artifacts is still missing. However preprocessing pipelines provides some efficient approaches to this challenge. In future, the progress in machine learning and deep learning based approaches may yield more efficient, accurate and robust artifact removal options. Online artifact removal methods such as ASR must be developed to overcome various artifacts in daily life to be efficient for BCIs.

Author details

İbrahim Kaya
Department of Biomedical Engineering, Izmir Katip Celebi University, Izmir, Turkey

*Address all correspondence to: ibrahimkaya21@yahoo.com

IntechOpen

© 2021 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the Creative Commons Attribution License (http://creativecommons.org/licenses/by/3.0), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited. (cc) BY

References

[1] Priemer R. Introductory signal processing. Vol. 6. World Scientific; 1991.

[2] Da Silva FL. EEG: origin and measurement. InEEg-fMRI 2009 (pp. 19-38). Springer, Berlin, Heidelberg.

[3] Urigüen JA, Garcia-Zapirain B. EEG artifact removal—state-of-the-art and guidelines. Journal of neural engineering. 2015 Apr 2;12(3):031001.

[4] Islam MK, Rastegarnia A, Yang Z. Methods for artifact detection and removal from scalp EEG: A review. Neurophysiologie Clinique/Clinical Neurophysiology. 2016 Nov 1;46(4-5): 287-305.

[5] Sazgar M, Young MG. Absolute epilepsy and EEG rotation review. Spinger; 2019.

[6] Croft RJ, Barry RJ. Removal of ocular artifact from the EEG: a review. Neurophysiologie Clinique/Clinical Neurophysiology. 2000 Feb 1;30(1):5-19.

[7] Vigon L, Saatchi MR, Mayhew JE, Fernandes R. Quantitative evaluation of techniques for ocular artefact filtering of EEG waveforms. IEE Proceedings-Science, Measurement and Technology. 2000 Sep 1;147(5):219-28.

[8] Vigon L, Saatchi MR, Mayhew JE, Fernandes R. Quantitative evaluation of techniques for ocular artefact filtering of EEG waveforms. IEE Proceedings-Science, Measurement and Technology. 2000 Sep 1;147(5):219-28.

[9] Matsuo F, Peters JF, Reilly EL. Electrical phenomena associated with movements of the eyelid. Electro-encephalography and clinical neuro-physiology. 1975 May 1;38(5):507-11.

[10] Goncharova II, McFarland DJ, Vaughan TM, Wolpaw JR. EMG contamination of EEG: spectral and topographical characteristics. Clinical neurophysiology. 2003 Sep 1;114(9):1580-93.

[11] McMenamin BW, Shackman AJ, Maxwell JS, Bachhuber DR, Koppenhaver AM, Greischar LL, Davidson RJ. Validation of ICA-based myogenic artifact correction for scalp and source-localized EEG. Neuroimage. 2010 Feb 1;49(3):2416-32.

[12] Sweeney KT, Ayaz H, Ward TE, Izzetoglu M, McLoone SF, Onaral B. A methodology for validating artifact removal techniques for physiological signals. IEEE transactions on information technology in biomedicine. 2012 Jul 10;16(5):918-26.

[13] Anderer P, Roberts S, Schlögl A, Gruber G, Klösch G, Herrmann W, Rappelsberger P, Filz O, Barbanoj MJ, Dorffner G, Saletu B. Artifact processing in computerized analysis of sleep EEG–a review. Neuropsychobiology. 1999; 40(3):150-7.

[14] McFarland DJ, McCane LM, David SV, Wolpaw JR. Spatial filter selection for EEG-based communication. Electroencephalography and clinical Neurophysiology. 1997 Sep 1;103(3): 386-94.

[15] Tamburro G, Stone DB, Comani S. Automatic Removal of Cardiac Interference (ARCI): a new approach for EEG data. Frontiers in neuroscience. 2019 May 8;13:441.

[16] Luck, S. J. Event-related potentials. In H. Cooper, P. M. Camic, D. L. Long, A. T. Panter, D. Rindskopf, & K. J. Sher

(Eds.), (2012). APA handbooks in psychology®. APA handbook of research methods in psychology, Vol. 1. Foundations, planning, measures, and psychometrics (p. 523-546). American Psychological Association. https://doi.org/10.1037/13619-028

[17] Dawson GD. A summation technique for the detection of small evoked potentials. Electroencephalography & clinical neurophysiology. 1954.

[18] Hoke M, Ross B, Wickesberg R, Lütkenhöner B. Weighted averaging—theory and application to electric response audiometry. Electroencephalography and clinical neurophysiology. 1984 May 1;57(5):484-9.

[19] Bezerianos A, Laskaris N, Fotopoulos S, Papathanasopoulos P. Data dependent weighted averages for recording of evoked potential signals. Electroencephalography and Clinical Neurophysiology/Evoked Potentials Section. 1995 Sep 1;96(5):468-71.

[20] Davila, C. E., & Mobin, M. S. (1992). Weighted averaging of evoked potentials. IEEE Transactions on Biomedical Engineering, 39(4), 338-345

[21] John MS, Dimitrijevic A, Picton TW. Weighted averaging of steady-state responses. Clinical Neurophysiology. 2001 Mar 1;112(3):555-62.

[22] Lütkenhöner B, Hoke M, Pantev C. Possibilities and limitations of weighted averaging. Biological cybernetics. 1985 Oct;52(6):409-16.

[23] Mühler R, Specht HV. Sorted averaging-principle and application to auditory brainstem responses. Scandinavian audiology. 1999 Jan 1;28(3):145-9.

[24] Rahne T, von Specht H, Mühler R. Sorted averaging—application to auditory event-related responses. Journal of neuroscience methods. 2008 Jul 15;172(1):74-8.

[25] Yabe H, Saito F, Fukushima Y. Median method for detecting endogenous event-related brain potentials. Electroencephalography and clinical Neurophysiology. 1993 Dec 1;87(6):403-7.

[26] Özdamar Ö, Kalayci T. Median averaging of auditory brain stem responses. Ear and hearing. 1999 Jun 1;20(3):253-64.

[27] Leonowicz Z, Karvanen J, Shishkin SL. Trimmed estimators for robust averaging of event-related potentials. Journal of neuroscience methods. 2005 Mar 15;142(1):17-26.

[28] Leski JM, Gacek A. Computationally effective algorithm for robust weighted averaging. IEEE transactions on biomedical engineering. 2004 Jun 21;51(7):1280-4.

[29] Fatourechi M, Bashashati A, Ward RK, Birch GE. EMG and EOG artifacts in brain computer interface systems: A survey. Clinical neurophysiology. 2007 Mar 1;118(3):480-94.

[30] Ochoa CJ, Polich J. P300 and blink instructions. Clinical Neurophysiology. 2000 Jan 1;111(1):93-8.

[31] Verleger R. The instruction to refrain from blinking affects auditory P3 and N1 amplitudes. Electroencephalography and Clinical Neurophysiology. 1991 Mar 1;78(3):240-51.

[32] Hagemann D, Naumann E. The effects of ocular artifacts on (lateralized) broadband power in the EEG. Clinical Neurophysiology. 2001 Feb 1;112(2):215-31.

[33] Chavez M, Grosselin F, Bussalb A, Fallani FD, Navarro-Sune X. Surrogate-based artifact removal from single-channel EEG. IEEE transactions on neural systems and rehabilitation engineering. 2018 Jan 22;26(3):540-50.

[34] Chen Y, Zhao Q, Hu B, Li J, Jiang H, Lin W, Li Y, Zhou S, Peng H. A method of removing ocular artifacts from EEG using discrete wavelet transform and Kalman filtering. In2016 IEEE International Conference on Bioinformatics and Biomedicine (BIBM) 2016 Dec 15 (pp. 1485-1492). IEEE.

[35] Lins OG, Picton TW, Berg P, Scherg M. Ocular artifacts in recording EEGs and event-related potentials II: Source dipoles and source components. Brain topography. 1993 Sep;6(1):65-78.

[36] Minguillon J, Lopez-Gordo MA, Pelayo F. Trends in EEG-BCI for daily-life: Requirements for artifact removal. Biomedical Signal Processing and Control. 2017 Jan 1;31:407-18.

[37] Shao SY, Shen KQ, Ong CJ, Wilder-Smith EP, Li XP. Automatic EEG artifact removal: a weighted support vector machine approach with error correction. IEEE Transactions on Biomedical Engineering. 2008 Oct 3;56(2):336-44.

[38] Jiang X, Bian GB, Tian Z. Removal of artifacts from EEG signals: a review. Sensors. 2019 Jan;19(5):987.

[39] Jung TP, Humphries C, Lee TW, Makeig S, McKeown MJ, Iragui V, Sejnowski TJ. Extended ICA removes artifacts from electroencephalographic recordings. Advances in neural information processing systems. 1998 Nov 30:894-900.

[40] Vigário RN. Extraction of ocular artefacts from EEG using independent component analysis. Electroencephalography and clinical neurophysiology. 1997 Sep 1;103(3):395-404.

[41] Barlow JS. Computerized clinical electroencephalography in perspective. IEEE Transactions on Biomedical Engineering. 1979 Jul(7):377-91.

[42] Verleger R. Valid identification of blink artefacts: are they larger than 50 μV in EEG records?. Electroencephalography and clinical Neurophysiology. 1993 Dec 1;87(6):354-63.

[43] Jung TP, Makeig S, Humphries C, Lee TW, Mckeown MJ, Iragui V, Sejnowski TJ. Removing electro-encephalographic artifacts by blind source separation. Psychophysiology. 2000 Mar;37(2):163-78.

[44] Grouiller F, Vercueil L, Krainik A, Segebarth C, Kahane P, David O. A comparative study of different artefact removal algorithms for EEG signals acquired during functional MRI. Neuroimage. 2007 Oct 15;38(1): 124-37.

[45] Makeig S, Bell AJ, Jung TP, Sejnowski TJ. Independent component analysis of electroencephalographic data. Advances in neural information processing systems. 1996 Dec 2:145-51.

[46] Chang CY, Hsu SH, Pion-Tonachini L, Jung TP. Evaluation of artifact subspace reconstruction for automatic EEG artifact removal. In2018 40th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC) 2018 Jul 18 (pp. 1242-1245). IEEE.

[47] Mannan MM, Kamran MA, Jeong MY. Identification and removal of physiological artifacts from electro-encephalogram signals: A review. Ieee Access. 2018 May 31;6:30630-52.

[48] Berg P, Scherg M. Dipole modelling of eye activity and its application to the removal of eye artefacts from the EEG and MEG. Clinical Physics and Physiological Measurement. 1991;12(A):49.

[49] Gotman J, Ives JR, Gloor P. Frequency content of EEG and EMG at seizure onset: possibility of removal of EMG artefact by digital filtering. Electroencephalography and clinical neurophysiology. 1981 Dec 1;52(6):626-39.

[50] Kim SP. Preprocessing of eeg. InComputational EEG Analysis 2018 (pp. 15-33). Springer, Singapore.

[51] Krishnaveni V, Jayaraman S, Anitha L, Ramadoss K. Removal of ocular artifacts from EEG using adaptive thresholding of wavelet coefficients. Journal of neural engineering. 2006 Nov 23;3(4):338.

[52] Zikov T, Bibian S, Dumont GA, Huzmezan M, Ries CR. A wavelet based de-noising technique for ocular artifact correction of the electroencephalogram. InProceedings of the Second Joint 24th Annual Conference and the Annual Fall Meeting of the Biomedical Engineering Society] [Engineering in Medicine and Biology 2002 Oct 23 (Vol. 1, pp. 98-105). IEEE.

[53] da Cruz JR, Chicherov V, Herzog MH, Figueiredo P. An automatic pre-processing pipeline for EEG analysis (APP) based on robust statistics. Clinical Neurophysiology. 2018 Jul 1; 129(7):1427-37.

[54] Nolan H, Whelan R, Reilly RB. FASTER: fully automated statistical thresholding for EEG artifact rejection. Journal of neuroscience methods. 2010 Sep 30;192(1):152-62.

[55] Delorme A, Makeig S. EEGLAB: an open source toolbox for analysis of single-trial EEG dynamics including independent component analysis. Journal of neuroscience methods. 2004 Mar 15;134(1):9-21.

[56] Pedroni A, Bahreini A, Langer N. Automagic: Standardized preprocessing of big EEG data. NeuroImage. 2019 Oct 15;200:460-73.

[57] Hatz F, Hardmeier M, Bousleiman H, Rüegg S, Schindler C, Fuhr P. Reliability of fully automated versus visually controlled pre-and post-processing of resting-state EEG. Clinical Neurophysiology. 2015 Feb 1;126(2):268-74.

[58] Bigdely-Shamlo N, Mullen T, Kothe C, Su KM, Robbins KA. The PREP pipeline: standardized preprocessing for large-scale EEG analysis. Frontiers in neuroinformatics. 2015 Jun 18;9:16.

[59] Mullen T. CleanLine EEGLAB plugin. San Diego, CA: Neuroimaging Informatics Toolsand Resources Clearinghouse (NITRC). 2012.

[60] Parra LC, Spence CD, Gerson AD, Sajda P. Recipes for the linear analysis of EEG. Neuroimage. 2005 Nov 1;28(2): 326-41.

[61] Winkler I, Haufe S, Tangermann M. Automatic classification of artifactual ICA-components for artifact removal in EEG signals. Behavioral and Brain Functions. 2011 Dec;7(1):1-5.

[62] Huster RJ, Debener S, Eichele T, Herrmann CS. Methods for simultaneous EEG-fMRI: an introductory review. Journal of Neuroscience. 2012 May 2;32(18):6053-60.

[63] Vanni S, Warnking J, Dojat M, Delon-Martin C, Bullier J, Segebarth C. Sequence of pattern onset responses in the human visual areas: an fMRI constrained VEP source analysis. Neuroimage. 2004 Mar 1;21(3):801-17.

[64] Wibral M, Bledowski C, Kohler A, Singer W, Muckli L. The timing of feedback to early visual cortex in the perception of long-range apparent motion. Cerebral cortex. 2009 Jul 1; 19(7):1567-82.

[65] Wibral M, Bledowski C, Turi G. Integration of separately recorded EEG/MEG and fMRI data. Simultaneous EEG and fMRI: recording, analysis, and application (Ullsperger M, Debener S, eds). 2010 May 28:209-34.

[66] de Munck JC, van Houdt PJ, Gonçalves SI, van Wegen E, Ossenblok PP. Novel artefact removal algorithms for co-registered EEG/fMRI based on selective averaging and subtraction. Neuroimage. 2013 Jan 1;64:407-15.

[67] Debener S, Kranczioch C, Gutberlet I. EEG quality: origin and reduction of the EEG cardiac-related artefact. InEEG-fMRI 2009 (pp. 135-151). Springer, Berlin, Heidelberg.

[68] Bonmassar G, Purdon PL, Jääskeläinen IP, Chiappa K, Solo V, Brown EN, Belliveau JW. Motion and ballistocardiogram artifact removal for interleaved recording of EEG and EPs during MRI. Neuroimage. 2002 Aug 1;16(4):1127-41.

[69] Allen PJ, Josephs O, Turner R. A method for removing imaging artifact from continuous EEG recorded during functional MRI. Neuroimage. 2000 Aug 1;12(2):230-9.

[70] Felblinger J, Slotboom J, Kreis R, Jung B, Boesch C. Restoration of electrophysiological signals distorted by inductive effects of magnetic field gradients during MR sequences. Magnetic Resonance in Medicine: An Official Journal of the International Society for Magnetic Resonance in Medicine. 1999 Apr;41(4):715-21.

[71] Debener S, Mullinger KJ, Niazy RK, Bowtell RW. Properties of the ballistocardiogram artefact as revealed by EEG recordings at 1.5, 3 and 7 T static magnetic field strength. International Journal of Psychophysiology. 2008 Mar 1;67(3):189-99.

[72] Allen PJ, Polizzi G, Krakow K, Fish DR, Lemieux L. Identification of EEG events in the MR scanner: the problem of pulse artifact and a method for its subtraction. Neuroimage. 1998 Oct 1;8(3):229-39.

[73] Ives JR, Warach S, Schmitt F, Edelman RR, Schomer DL. Monitoring the patient's EEG during echo planar MRI. Electroencephalography and clinical neurophysiology. 1993 Dec 1;87(6):417-20.

[74] Marino M, Liu Q, Koudelka V, Porcaro C, Hlinka J, Wenderoth N, Mantini D. Adaptive optimal basis set for BCG artifact removal in simultaneous EEG-fMRI. Scientific reports. 2018 Jun 11;8(1):1-1.

[75] Malafeev A, Omlin X, Wierzbicka A, Wichniak A, Jernajczyk W, Riener R, Achermann P. Automatic artefact detection in single-channel sleep EEG recordings. Journal of sleep research. 2019 Apr;28(2):e12679.

[76] McFarland DJ, Sarnacki WA, Vaughan TM, Wolpaw JR. Brain-computer interface (BCI) operation: signal and noise during early training sessions. Clinical Neurophysiology. 2005 Jan 1;116(1):56-62.

[77] Chen X, Liu A, Peng H, Ward RK. A preliminary study of muscular artifact cancellation in single-channel EEG. Sensors. 2014 Oct;14(10):18370-89.

[78] Chen X, Liu A, Chiang J, Wang ZJ, McKeown MJ, Ward RK. Removing

muscle artifacts from EEG data: Multichannel or single-channel techniques?. IEEE Sensors Journal. 2015 Dec 8;16(7):1986-97.

[79] Daly I, Nicolaou N, Nasuto SJ, Warwick K. Automated artifact removal from the electroencephalogram: a comparative study. Clinical EEG and neuroscience. 2013 Oct;44(4):291-306.

[80] Daly I, Scherer R, Billinger M, Müller-Putz G. FORCe: Fully online and automated artifact removal for brain-computer interfacing. IEEE transactions on neural systems and rehabilitation engineering. 2014 Aug 13;23(5):725-36.

[81] Khatun S, Mahajan R, Morshed BI. Comparative study of wavelet-based unsupervised ocular artifact removal techniques for single-channel EEG data. IEEE journal of translational engineering in health and medicine. 2016 Mar 22; 4:1-8.

[82] Sreeja SR, Sahay RR, Samanta D, Mitra P. Removal of eye blink artifacts from EEG signals using sparsity. IEEE journal of biomedical and health informatics. 2017 Nov 13;22(5):1362-72.

[83] Chang WD, Lim JH, Im CH. An unsupervised eye blink artifact detection method for real-time electroencephalogram processing. Physiological measurement. 2016 Feb 19;37(3):401.

[84] Kothe CA, Jung TP, inventors. Artifact removal techniques with signal reconstruction. United States patent application US 14/895,440. 2016 Apr 28.