We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,300

Open access books available

172,000

International authors and editors

190M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities

WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com

An Analogue Front-End System with a Low-Power On-Chip Filter and ADC for Portable ECG Detection Devices

Shuenn-Yuh Lee¹, Jia-Hua Hong¹, Jin-Ching Lee² and Qiang Fang³

¹Department of Electrical Engineering, National Chung-Cheng University, Taiwan

²Department of Biotechnology, College of Life Science, Kaohsiung Medical University,

³Department of Electrical and Computer Engineering, RMIT University,

^{1,2}Taiwan

³Australia

1. Introduction

Medical diagnostic instruments can be made into portable devices for the purpose of home care, such as the diagnosis of heart disease. These assisting devices are not only used to monitor patients but are also beneficial as handy and convenient medical instruments. Hence, for reasons of both portability and durability, designers should reduce the power consumption of assistant devices as much as possible to extend their battery lifetime. However, achieving the low power requirement of the ECG sensing and the processing board for the ECG with commercial discrete components (A21-0003) is difficult because the low power consumer electronics for ECG acquisition systems are not yet available. With the help of the integrated circuit technology, the power-saving requirement of portable and durable equipment gives circuit designers the impetus to reduce the power consumption of analogue front-end circuits in ECG acquisition systems. In addition, the analogue front-end circuits, which are the interface between physical signals and the digital processor, must be operated at a low-supply voltage to be integrated into the low-voltage system-on-a-chip (SOC) system (Eshraghian, 2006). Therefore, the chapter will present two design examples of low-voltage (1 V) and low-power ($<1 \mu\text{W}$) on-chip circuits including a low-pass filter (LPF) and an analogue-to-digital converter (ADC) to demonstrate the possibility of developing the low-voltage low-power ECG acquisition SOC.

Fig. 1. An analog front-end system for portable ECG detection devices

In clinical electrocardiogram (ECG) acquisition, several leads combined with signals from different body parts (i.e., from the right wrist and the left ankle) are utilised to trace the electric activity of the heart. In this system, as shown in Fig. 1, the ECG acquisition board translates the body signal to six leads and processes the signal using a low-pass filter (LPF) and a successive-approximation analogue-to-digital converter (SAADC). The acquisition board is composed of some discrete components including an instrumentation amplifier, a high-pass filter, a 60-Hz notch filter, and a common-level adjuster. The main function of the acquisition board is to pre-amplify the weak ECG signal whose amplitude is between 100 μV and 4 mV (Webster, 1995). The range of the ECG signal means that this system requires a signal-to-noise and distortion-ratio (SNDR) of at least 32 dB (that is, 6 bits) to detect heart activities precisely. There is more than one sensing channel on the board, and thus this system suffers from some problems such as crosstalk, settling time, and dispensable switch-induced noise (Olsson et al., 2005). The frequency range of the ECG signal is between 0.1 Hz to 250 Hz. Therefore, an on-chip low-power LPF behind the acquisition board provides a low cut-off frequency (250 Hz) to decrease the out-of-band high-frequency noise. On the other hand, the noise under 0.1 Hz will be eliminated by a high-pass filter on the acquisition board. To compensate for the in-band signal attenuation in the LPF, an adjustable compensation amplifier located between the filter and the SAADC was designed. It can decrease the influence of the switch-induced noise caused by the sampling behaviour of the SAADC. Because the total power of the ECG acquisition board will be dominated by the high-order LPF and ADC integrated by off-chip components, in this chapter, low-power integrated-circuit design techniques are proposed and adopted to implement these two chips under the 0.18- μm TSMC CMOS process. It reveals that the low-power miniature ECG acquisition system is realisable, and it can be integrated into wearable devices for ECG signal acquisition.

The whole analogue front-end system will be introduced in detail. First, a multi-function acquisition board for the ECG signal and a low-power anti-aliasing operational transconductance amplifier-C (OTA-C) filter without the off-chip capacitors under low-frequency operation are described in Section 2. Furthermore, the design and utilisation of the low-power SAADC are also presented in Section 2. Finally, the practical human-body measurement results of the whole system and the conclusions are presented in Sections 3 and 4, respectively.

2. Design

2.1 ECG signal acquisition board

Human-body signals are too complex to be directly fed into on-chip analogue circuits including a LPF and a SAADC, and hence the Wilson circuit on board is used to transfer human-body signals to six leads. In addition to the Wilson circuit, other elements including an instrumentation amplifier, an isolator, a high-pass filter, and a 60-Hz notch filter on the acquisition board are required to capture the ECG signal, as shown in Fig. 2. They will be introduced in the next subsections.

2.1.1 Wilson circuit

In normal ECG signal detection, the Wilson circuit is commonly used. As shown in the left part of Fig. 1, the electrodes are stuck on the right wrist, left wrist, and left ankle, and each node connects with a resistor to a common node called the Wilson central terminal. The three main leads (Lead I, II, and III) and three minor leads (aV_R , aV_F , and aV_L) are formed by these terminals and some nodes in the circuit.

Fig. 2. Analogue blocks on the ECG signal acquisition board and analogue filters implemented by discrete components.

2.1.2 Instrumentation amplifier and isolation circuit

To obtain a high common-mode rejection ratio (CMRR), an instrumentation amplifier is adopted as the preamplifier in the analogue front-end system. In this case, an instrumentation amplifier with chip INA-128 is used. It not only provides a high CMRR (at least 120 dB) but also high precision, low power consumption, and low quiescent current. Aside from these, the gain of this amplifier can be adjusted to an appropriate level to fit the operation condition of the chip. Because the experimentation in this chapter is done on an actual human body, for safety considerations, a high linear isolator with an ISO122 chip is utilised to prevent electric shock.

Fig. 3. Ladder type fifth-order passive Butterworth filter

2.1.3 High pass and 60-Hz notch filters

Behind the isolator, a high-pass filter with a cut-off frequency of 0.1 Hz is implemented by an active RC circuit to reject the low-frequency noise. The last stage is a notch filter with a twin-T structure; it protects the circuits against the 60-Hz noise produced by the AC-110V power supply. The realisations of the high-pass filter and notch filter are shown in Fig. 2, respectively.

2.2 Anti-aliasing OTA-C filter

For the long-term physical signal detection and monitor system, the use of a switched capacitor (SC) is a popular technique (Lasanen & Kostamovaara, 2005). However, the low sampling frequency in the kilohertz range will result in leakage, and the power consumption will be increased by the operational amplifiers in the SC circuits. Hence, continuous-time operational transconductance amplifier (OTA)-based filters are preferred in low-frequency applications, and the transistors inside a filter can be operated in the sub-threshold region to save power and to achieve ultra-low transconductance (a G_m of the order of a few nano-amperes per volt to save the capacitive area) (Salthouse & Sarpeshkar, 2003). The performance of the OTA-based filter is dominated by OTA, and the time constant of the OTA-C integrators is determined by the ratio of the capacitor value to the small transconductance.

Fig. 4. Fully-differential operational transconductance amplifier (OTA)

2.2.1 Filter selection for ECG detection

The detection circuits must attenuate the out-of-band interference before the ADC to avoid aliasing and to diagnose the disease precisely. For these reasons, a fifth-order ladder-type Butterworth filter, as shown in Fig. 3, with a maximum flat response and a cut-off frequency of 250 Hz, is selected for this design. Adopting the ladder-type filter makes the transfer function more robust and more insensitive in terms of component variations. In addition, to implement the filter with the integrated circuits (ICs), the filter synthesis using the signal flow graph (SFG) mapping method should be applicably used (Schaumann & Valkenburg, 2001).

2.2.2 Nonlinearity and input referred noise of the OTA

An ultra-low- G_m OTA can be implemented as the low-frequency filter with an on-chip capacitor (Bustos et al., 2000). Fig. 4 shows an OTA circuit that uses two techniques to reduce the transconductance, including current cancellation and current division.

Furthermore, the fully differential structure provides a higher capability in terms of common-mode rejection and an increase of 3 dB in the dynamic range rather than the single-end structure. In addition, all transistors in the OTA are operated in sub-threshold region to save the power consumption.

According to the analysis of the nonlinearity and input referred noise of this filter (Lee & Cheng, 2009), the dominated third harmonic distortion ($HD3$) caused by the device M_R can be expressed as

$$HD3 \cong \frac{\gamma \cdot v_{SD}^2}{96(v_{SG} - V_{th})(2\phi_F - v_{SB})^{\frac{3}{2}}} \quad (1)$$

where ϕ_F represents the Fermi potential. In this design, $HD3$ is to be suppressed below -50 dB with a differential input level of 100 mV. Second harmonic distortion ($HD2$) is almost cancelled by using a fully differential structure. As previously mentioned, the system detecting the ECG signal should possess an SNDR greater than 32 dB. Because the distortion is below 50 dB (that is, it is sufficient), the other issues in terms of equivalent input noise should be considered in the filter design.

By selecting the PMOS as the input stage of Fig. 4, the flicker noise should be low. Given this, the required input referred noise of the filter can be expressed as

$$Noi_{in,ref}(rms) = \frac{V_{in,rms}}{10^{SNR(dB)/20}} \quad (2)$$

According to the description, with an SNR of 42 dB (7 bits), the input referred noise must be less than 560 μV_{rms} for an input voltage of 100 mV_{rms}.

Fig. 5. Active circuit realization of the fifth-order Butterworth filter

2.2.3 Fifth-order OTA-C low-pass filter

The realisations of the fifth-order OTA-C filter with common-mode feedback (CMFB) circuits are illustrated in Fig. 5. To contrast with the ladder type, the overall circuit is composed of two grounded resistors G_{m0} and G_{m6} , two gyrators A and B, which implement the equivalent inductors L_2 and L_4 , respectively, and five capacitors $C_1 \sim C_5$. To reduce the

power consumption, five common-mode feedback circuits are shared by the eleven OTAs of the filter. The common-mode feedback circuits provide sensing of the common output voltages on nodes $a\sim e$ to control the bias voltage $vb3$ of the OTA.

2.3 Low-power successive approximation ADC

For the required performance of ECG signals with an amplitude between 100 μV and 4 mV (Webster, 1995), the resolution of the ADC with analogue filter in this ECG signal processing system usually only has to be between 6 to 8 bits. In this chapter, a low-power SAADC with 8-bit resolution and 10-KHz sampling frequency is designed. It is not only applied to the ECG signal, but also used for other physical signals, such as Electroneurography (ENG). The basic architecture of a SAADC is illustrated in Fig. 6. The converter consists of a sample/hold (S/H) circuit, a comparator, a successive approximation register (SAR) controller, and an 8-bit digital-to-analogue converter (DAC). Using a binary searching algorithm, the input sample voltage can be successively approximated by the DAC output voltage. For a N-bit SAADC, N cycles are required to convert the analogue signals into digital codes. Obviously, the DAC dominates the accuracy and the speed of the SAADC. To conform to the system specifications, a low-power, opamp-free, capacitor-based DAC with an 80-kHz sampling rate is implemented. The sub-circuit design of the SAADC will be described in the following subsections in detail.

Fig. 6. Block diagram of a successive approximation ADC.

Fig. 7. (a) Passive S/H circuit with dummy switch, and (b) Comparator circuit

2.3.1 Sample/hold circuit

To decrease the power consumption of the SAADC, a passive S/H circuit illustrated in Fig. 7(a) is adopted. It consists of the NMOS switch S_1 and the sampling capacitor C_H . A dummy switch S_2 is adopted to circumvent the problem of the charge injection and the clock feedthrough and to compensate for the charge error. It will meet the requirements of 8-bit resolution.

2.3.2 Comparator circuit

The comparator used in the SAADC is illustrated in Fig. 7(b); it is a track-and-latch stage. Because the accuracy of the comparator plays a critical role in the SAADC, the transistors M_{s1} and M_{s2} are included to avoid hysteresis or delayed response when resetting the phase. The operational principle is as follows. When the clock is high, the comparator is operated in the resetting mode, and both outputs (V_{OUT+} and V_{OUT-}) are pulled to V_{DD} (high). On the other hand, when the clock is low, the circuit will execute the comparison of differential input, and the outputs level (V_{OUT+} or V_{OUT-}) of the comparator will depend on the difference between V_{IN+} and V_{IN-} .

The design of the bias current I_b is critical for the performance of the comparator, including speed, noise, and power consumption. For speed considerations, the frequency response of the comparator depending on the dominant pole should be analysed. The dominated pole of the comparator is located at node P and can be described as follows:

$$\omega_P = \frac{g_{m1}}{C_P}, \quad C_P = C_{gs3} + C_{db2} + C_{db4} + C_{db_{s2}} + C_{buffer} \quad (3)$$

where C_P is the total capacitance at node P. To fit the 10 kS/s sampling rate of the 8-bit SAADC, the speed of the comparator must be operated at no less than 80 kHz. However, for settling within the 0.1% accuracy, the required unity-gain bandwidth of the comparator must satisfy seven times its time constant. Therefore, a comparator with a unity-gain bandwidth of 1 MHz is implemented. Hence, a comparing time of less than 1 μ s can be achieved.

Fig. 8. Architecture of an 8-bit capacitor-based DAC.

Because the operational speed of the comparator is at a low frequency, flicker noise will dominate the input-referred noise. Hence, a low bias current with large channel width is appropriate to decrease power consumption. In this chapter, a bias current with only 400 nA is adopted to let the comparator operate in the sub-threshold region under the 8-bit and 10-kHz requirements.

2.3.3 Capacitor-based DAC

There are various structures, including resistor strings, the current-mode approach, and capacitor arrays, that can be adopted to implement the internal DAC in the SAADC. Among these, the capacitor-array structure is most suitable for the low-power approach (Hong &

a finite state machine (FSM) is used to generate the control signal. At the beginning of conversion, the most significant bit (MSB) is set to one, whereas the remaining bits are set to zero. The initial value of the DAC output is then set to 0.5 V (1/2 full scale). If the comparator output is low, the MSB will be set to 0 and saved in the output of the SAR. If the output is high, the MSB will remain 1. The residue bits will be processed in the same operations until the least significant bit (LSB) is determined.

3. Experimental results

The three main elements introduced above were integrated to a low-power analogue front-end system on a detection board; the measured analogue front-end system was set up as shown in Fig. 10. The picture also shows the practical measurement conditions: the electrodes are stuck on both wrists and both ankles. The circuits of the OTA-C filter and SAADC were fabricated in a 0.18- μm TSMC process with metal-insulator-metal (MIM) capacitors. The die area of the OTA-C filter and SAADC are 0.135 mm² and 0.12 mm², respectively. Figure 10 also shows the microphotographs of the two chips.

Fig. 10. Chip-microphotographs of an OTA-C filter and SAADC, respectively, and the measurement setup of an analog front-end ECG detection system

3.1 Filter measurement

A differential sinusoidal wave with a magnitude of 100 mV_{PP} is fed into the chip to measure the frequency response and the power spectrum with an input frequency of 50 Hz. Referring to Fig. 11(a), the -3 dB frequency is around 240 Hz, where the inband gain degradation (from -6 dB to -10 dB) arises from the finite-gain effect of the OTA. The measured third harmonic distortion (HD3) in Fig. 11(b) is well below -49 dB, which is close to the simulation estimation. In addition, the integrated input referred noise from 1 - 250 Hz is 340 μV_{rms} . The power consumption is 453 nW at a supply voltage of 1 V.

Fig. 11. Measurement of the OTA-C filter. (a) Frequency response. (b) Output power spectrum density.

3.2 SAADC measurement

The 1-kHz input signal with a 500-mV_{pp} full-swing magnitude sinusoidal wave is fed into the SAADC to measure integral nonlinearity (INL) and differential nonlinearity (DNL). Moreover, the sampling rate is 10 kHz. Fig. 12 shows the maximum DNL is +0.38/-0.41 LSB, whereas the maximum INL is +0.6/-0.89 LSB. Moreover, a full-scale 100-Hz sine-wave spectrum measured at a 1-kHz sampling rate is illustrated in Fig. 13 to demonstrate the low-frequency performance. The signal-to-noise distortion ratio (SNDR) in the ECG bandwidth (250 Hz) is 48.46 dB, and the spurious free dynamic range (SFDR) is 57 dB. Meanwhile, the effective number of bits (ENOB) defined as follows is 7.76 bits:

$$ENOB = \frac{SNDR - 1.76}{6.02} \tag{6}$$

Fig. 12. Measured DNL and INL of the SAADC

Fig. 13. Measured output spectrum at 100Hz input frequency and 1 kHz sampling rate.

Fig. 14. Measurement results of the ECG analog front-end system

3.3 Real ECG signal testing

Corresponding to the system shown in Fig. 1, the measurement results are also illustrated in Fig. 14. They are measured from each output node of the three main partitions on the system, including (a) the ECG acquisition board, (b) the OTA-C filter, and (c) the SAADC. We can observe the relations and functions between each output. Channel 1 in Fig. 14 is the initial pre-amplified ECG signal with obvious high-frequency noise generated by the human body. Because some tiny physical waves are covered, this kind of ECG signal is inconvenient for the diagnosis of heart disease. With the help of the low-pass OTA-C filter, the high frequency noise can be significantly attenuated, and the tracing signal with a clear baseline is shown in channel 2. To demonstrate the operation of the SAADC, the waveform view showing in the logic analyser is adopted to present the conversion result. Channel 3 in Fig. 14 shows the real-time human-body digital ECG waveform reconstructed by the 8-bit decimal codes of the SAADC. Similarly, a clear baseline is observed in this graph, and the 8-bit digital codes can be accepted by the post digital processor to diagnose the abnormal heart activities precisely.

4. Conclusions

A low-power analogue front-end system for ECG detection consisting of an acquisition board and two low-power on-chip components is presented. The design issues including an off-chip ECG signal acquisition board, an on-chip analogue filter, and an on-chip ADC are introduced in this chapter. The result reveals that developing an ultra low-power ECG acquisition SOC is possible. In the future, the entire elements of these three partitions will be integrated into a single chip to save area and achieve a fully low-voltage and low-power ECG acquisition SOC for wearable applications.

5. Acknowledgment

The authors would like to thank the Chip Implementation Center (CIC), the National Science Council (NSC) of Taiwan, R. O. C. under Grant NSC 99-2628-E-194-032, NSC 99-2220-E-194-001, and NSC 99-2220-E-194-006, the Australian Research Council, Australia for their support of this work, and the help of CBIC Laboratory members, Yu-Cheng Su, Chih-Jen Cheng, Kun-Min Huang, and Wei-Chun Kao.

6. References

- A21-0003. ECG acquisition board. <http://www.icci.com.tw>
- Bustos S. S., Martinez J. S., Maloberti F., Sinencio E. S., (2000). A 60-dB dynamic-range CMOS sixth-Order 2.4Hz low-pass filter for medical application. *IEEE Trans. Circuits Syst. II: Analog and Digital Signal Processing*, Vol. 47, No.12, (2000), pp. 1391-1398, ISSN 1057-7130.
- Eshraghian K. (2006). SoC emerging technologies. *Proc. IEEE*, Vol. 94, No. 6, (2006), pp. 1197-1213, ISSN 0018-9219.
- Hong H. C., Lee G. M. (2007). A 65-fJ/conversion-step 0.9-V 200-kS/s rail-to-rail 8-bit successive approximation ADC. *IEEE J. Solid-State Circuits*, Vol. 42, No. 10, (2007), pp. 2161-2168, ISSN 0018-9200.
- Lasanen K., Kostamovaara J. (2005). A 1-V analog CMOS front-end for detecting QRS complexes in a cardiac signal. *IEEE Trans. Circuits Syst. I: Regular Papers*, Vol. 52, No. 12, (2005), pp. 2584-2594, ISSN 1057-7122.
- Lee S. Y., Cheng C. J. (2009). Systematic design and modeling of a OTA-C filter for portable ECG detection. *IEEE Trans. Biomed. Circuits Syst.*, Vol. 3, No. 1, (2009), pp. 53-64, ISSN 1932-4545.
- Olsson R. H., Buhl D. L., Sirota A. M. Guzski G., Wise K. D. (2005). Band-tunable and multiplexed integrated circuits for simultaneous recording and stimulation with microelectrode arrays. *IEEE Trans. Biomed. Eng.*, Vol. 52, No. 7, (2005), pp. 1303-1311, ISSN 0018-9294.
- Rossi A., Fucili G. (1996). Nonredundant successive approximation register for A/D converters. *Electronics Letters*, Vol. 32, No. 12, (1996), pp. 1055-1056, ISSN 0013-5194.
- Salthouse C.D., Sarpeshkar R. (2003). A practical micropower programmable bandpass filter for use in bionic ears. *IEEE J. Solid-State Circuits*, Vol. 38, No. 1, (2003), pp. 63-70, ISSN 0018-9200.
- Schaumann R., Valkenburg M. E. V. (2001). *Design of Analog Filters*. New York: Oxford.
- Webster J. G. (1995). *Design of Cardiac Pacemakers*. Piscataway, NJ: IEEE Press.

Advances in Electrocardiograms - Methods and Analysis

Edited by PhD. Richard Millis

ISBN 978-953-307-923-3

Hard cover, 390 pages

Publisher InTech

Published online 25, January, 2012

Published in print edition January, 2012

Electrocardiograms are one of the most widely used methods for evaluating the structure-function relationships of the heart in health and disease. This book is the first of two volumes which reviews recent advancements in electrocardiography. This volume lays the groundwork for understanding the technical aspects of these advancements. The five sections of this volume, Cardiac Anatomy, ECG Technique, ECG Features, Heart Rate Variability and ECG Data Management, provide comprehensive reviews of advancements in the technical and analytical methods for interpreting and evaluating electrocardiograms. This volume is complemented with anatomical diagrams, electrocardiogram recordings, flow diagrams and algorithms which demonstrate the most modern principles of electrocardiography. The chapters which form this volume describe how the technical impediments inherent to instrument-patient interfacing, recording and interpreting variations in electrocardiogram time intervals and morphologies, as well as electrocardiogram data sharing have been effectively overcome. The advent of novel detection, filtering and testing devices are described. Foremost, among these devices are innovative algorithms for automating the evaluation of electrocardiograms.

How to reference

In order to correctly reference this scholarly work, feel free to copy and paste the following:

Shuenn-Yuh Lee, Jia-Hua Hong, Jin-Ching Lee and Qiang Fang (2012). An Analogue Front-End System with a Low-Power On-Chip Filter and ADC for Portable ECG Detection Devices, *Advances in Electrocardiograms - Methods and Analysis*, PhD. Richard Millis (Ed.), ISBN: 978-953-307-923-3, InTech, Available from: <http://www.intechopen.com/books/advances-in-electrocardiograms-methods-and-analysis/an-analogue-front-end-system-with-a-low-power-on-chip-filter-and-adc-for-portable-ecg-detection-devi>

INTECH
open science | open minds

InTech Europe

University Campus STeP Ri
Slavka Krautzeka 83/A
51000 Rijeka, Croatia
Phone: +385 (51) 770 447
Fax: +385 (51) 686 166
www.intechopen.com

InTech China

Unit 405, Office Block, Hotel Equatorial Shanghai
No.65, Yan An Road (West), Shanghai, 200040, China
中国上海市延安西路65号上海国际贵都大饭店办公楼405单元
Phone: +86-21-62489820
Fax: +86-21-62489821

© 2012 The Author(s). Licensee IntechOpen. This is an open access article distributed under the terms of the [Creative Commons Attribution 3.0 License](https://creativecommons.org/licenses/by/3.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

IntechOpen

IntechOpen