We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,300

Open access books available

172,000

International authors and editors

190M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities

WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com

Introductory Chapter: Gas Chromatography - The Most Versatile Analytical Technique

Peter Kusch

1. Definition and short history of gas chromatography

Gas chromatography (GC) is a type of chromatography. According to the International Union of Pure and Applied Chemistry (IUPAC) recommendation, gas chromatography is defined as a separation technique in which the mobile phase is a gas. Gas chromatography is always carried out in a column [1]. GC is a separation and detection method for sample mixtures, whose components can be volatilized without thermal decomposition. The analytical procedure is used for the determination of organic substances; usually molecules have a molecular mass of less than 500 g/mol and a boiling point of less than 400°C. GC is a technique used to separate mixtures of gaseous chemical compounds based on differences in the compounds' relative affinities for a solid (gas-solid chromatography) or liquid (gas-liquid chromatography) stationary phase held within a column.

Gas-liquid partition chromatography was invented by Martin and James from the National Institute for Medical Research, London, in 1952. The invention of this technique is generally attributed to the inventors in their 1952 published paper in the *Biochemical Journal* [2]. In this publication the theory of the partition column has been extended to cover a compressible mobile phase, and gas-liquid partition columns were described for the separation of volatile fatty acids. In the same year, the *Nobel Prize* in chemistry was awarded jointly to Martin and Synge for their *invention of partition chromatography*. Starting from this time, gas chromatography has become one of the most important and widely applied analytical techniques in modern chemistry. The first commercial gas chromatograph was introduced in 1955 by Perkin-Elmer (USA). Subsequently, this method was used to study petrochemical products. Today, gas chromatography is one of the most widespread investigation methods of instrumental analysis. This technique is used in the laboratories of the chemical, petrochemical, and pharmaceutical industries, in the research institutes and also in the clinical and environmental and food and beverage analyses. Recent developments in GC have resulted in the introduction of better and selective fused silica capillary columns and methods for sample preparation. Newer separation and detection solutions, such as fast GC, multidimensional separation GC techniques (GC $\times$ GC), and hyphenation of GC and GC $\times$ GC with mass spectrometry (MS), with triple quadrupole mass spectrometry and with time-of-flight mass spectrometry (TOF-MS), have been developed and become industrial routine. Analytical pyrolysis (Py) technique hyphenated to GC and GC/MS has extended the range of possible tools for the characterization of synthetic polymers and copolymers. This technique has been

used extensively over the last 30 years as a complementary analytical tool used to characterize the structure of synthetic organic polymers and copolymers, polymer blends, biopolymers, and natural resins [3].

2. Multidimensional gas chromatography

Developed by Phillips and coworker at the Southern Illinois University (USA) in the early 1990s, comprehensive multidimensional GC (GC $\times$ GC or 2D GC) is a powerful technique for samples containing very large numbers of compounds of interest and also for samples which exhibit high chemical complexity. This technique can be used to separate very complex mixtures, such as those found in the petrochemical, environmental, and food and fragrance industries [4–6]. The method uses two capillary columns, typically of very different polarities, installed in series with a modulator in between. The first column is in principle nonpolar or low polar, and the second column is polar. The length of the first column might typically be 20–30 m, the inner diameter 0.25 mm, and the film thickness 0.25 μm . The second column is typically shorter (1–2 m), the inner diameter is narrower (0.1 mm), and the stationary phase is thinner (0.1 μm), to allow for faster separations. The entire assembly is located inside the GC oven [6]. The modulator collects effluent from the first column for a fraction of the time equal to peak width. The modulator focuses the material collected from each cut into a very narrow band through flow compression. It introduces the bands sequentially onto the second column, resulting in additional separation for each band injected onto the second column [4–9]. The most common data transformation is the construction of a 2D representation, in which one axis represents the separation on the first column (first dimension), and the other axis represents the secondary column separation (second dimension). Therefore, the look of GC $\times$ GC chromatograms appears completely different from conventional GC chromatogram showing a two-dimensional plane where analyte spots are scattered about [7, 8]. A contour plot, using elevation lines or color coding, represents the signal intensity. 2D GC data are primarily used for

Figure 1. 2D GC plot of a refinery stream boiling at diesel temperature range. The scale indicates the relative signal intensity. Figure reprinted from Ref. [10] with permission from ACS.

qualitative analysis; however, quantitative multidimensional GC analysis is also possible [9]. **Figure 1** shows an exemplary 2D GC plot of a refinery stream boiling at diesel temperature range [10].

In this book, state of the art of gas chromatography and new developments and applications are presented. New sample preparation techniques, derivatization methods, and hyphenation with mass spectrometry are described.

IntechOpen

IntechOpen

Author details

Peter Kusch

Department of Applied Natural Sciences, Bonn-Rhein-Sieg University of Applied Sciences, Rheinbach, Germany

*Address all correspondence to: ptrkusch@arcor.de

IntechOpen

© 2018 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/3.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

References

- [1] IUPAC Compendium of Chemical Terminology, International Union of Pure and Applied Chemistry, Version 2.3.3. 2014
- [2] James AT, Martin AJP. Gas-liquid partition chromatography: The separation and micro-estimation of volatile fatty acids from formic acid to dodecanoic acid. *The Biochemical Journal*. 1952;**50**(5):679-690
- [3] Kusch P. Pyrolysis—Gas Chromatography/Mass Spectrometry of Polymeric Materials. London, UK: World Scientific Europe; 2018
- [4] Liu Z, Phillips JB. Comprehensive two-dimensional gas chromatography using an on column thermal modulator interface. *Journal of Chromatographic Science*. 1991;**29**(6):227-231
- [5] Phillips JB, Xu J. Comprehensive multi-dimensional gas chromatography. *Journal of Chromatography A*. 1995;**703**:327-334
- [6] Comprehensive GC System Based on Flow Modulation for the 7890A GC, Application Brief. Wilmington, DE, USA: Agilent Technologies; 2008
- [7] Hinshaw JV. Comprehensive two-dimensional gas chromatography. *LCGC Europe*. 2004;**17**(2):86-95
- [8] Mondello L. GC × GC Handbook, Fundamental Principles of Comprehensive 2D GC. Japan: Shimadzu; 2012
- [9] Taylor T. A short introduction to multidimensional GC. *LCGC North America*. 2012;**30**(9):870
- [10] Wang FC, Qiang K, Green LA. GCxMS of diesel: A two-dimensional separation approach. *Analytical Chemistry*. 2005;**77**(9):2777-2785