We are IntechOpen, the world's leading publisher of Open Access books Built by scientists, for scientists

6,300

Open access books available

172,000

International authors and editors

190M

Downloads

Our authors are among the

154

Countries delivered to

TOP 1%

most cited scientists

12.2%

Contributors from top 500 universities

WEB OF SCIENCE™

Selection of our books indexed in the Book Citation Index
in Web of Science™ Core Collection (BKCI)

Interested in publishing with us?
Contact book.department@intechopen.com

Numbers displayed above are based on latest data collected.
For more information visit www.intechopen.com

Introductory Chapter: Technology and Orthotics and Prosthetics

Mokhtar Arazpour

1. Introduction

In recent years, the use of technology has seen significant growth in the design and manufacture of advanced and intelligent orthoses and prostheses. Progress has been made in 4 areas in orthosis and prosthetics.

- 1) Material
- 2) Fabrication approach
- 3) Intelligent joints for orthoses
- 4) Intelligent joint in prostheses.

This has been accompanied by increasing growth in the materials used to make orthoses and prostheses. Activities of subject, the weight of the user are two important issues in used materials in fabrication of orthoses and prostheses for each subject. Flexible polymers, Carbon fiber, Kevlar and titanium are materials that were used to provide comfort and the strength and durability of the fabricated devices [1].

Computer-aided design/computer-aided manufacturing (CAD/CAM) technology is new manner to design and fabricate each type of orthoses and prostheses [2]. Scanning done by laser or hand held wand to provide measurement of limb. Based on this information the size and shape of limbs is given to prosthetist-orthotist to design and fabricate the mentioned orthoses and prostheses for the subject by computer [3].

Subjects with poliomyelitis, incomplete spinal cord injury and quadriceps weakness who have not knee stabilization during walking used devices with locked knee joints [4]. Walking with this condition provided high energy consumption and fatigue. Stance control knee joints are new generation of orthoses that developed in recent years in the world by many different companies [5]. Some of this type of joints is intelligent and smart that provided stability in stance phase and free swing during walking [6].

Intelligent prostheses can now be fabricated for the subjects with amputation [7]. This type of prosthesis in the lower limb permits the knee joint to sense alternation in position, speed and force, enabling subjects with amputations to walk down stairs and hills with confidence [8]. There were many developments in the prosthetic feet for running and other sport conditions. In the upper limbs prostheses, electrically powered units were used. This type of prosthesis can detect myoelectric signals provided by muscles to control the prosthesis [9].

Development in provided different orthoses and prostheses with advanced technology cause the prosthetist/orthotist is well trained and educated to construct devices based on the subject needs of each patient and to develop appropriate suggestions.

IntechOpen

IntechOpen

Author details

Mokhtar Arazpour

University of Social Welfare and Rehabilitation Sciences, Tehran, Iran

*Address all correspondence to: m.arazpour@yahoo.com

IntechOpen

© 2021 The Author(s). Licensee IntechOpen. This chapter is distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/3.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

References

- [1] Wong MS, Beygi BH, Zheng Y. Materials for exoskeletal orthotic and prosthetic systems. *Encyclopedia of Biomedical Engineering*; Elsevier; 2018. p. 352-67.
- [2] Supan TJ. Principles of fabrication. *Atlas of Orthoses and Assistive Devices*; Elsevier; 2019. p. 42-8. e1.
- [3] Banga HK, Kalra P, Belokar RM, Kumar R. Design and Fabrication of Prosthetic and Orthotic Product by 3D Printing. *Orthotics and Prosthetics*; IntechOpen; 2020.
- [4] Moreno JC, Brunetti F, Rocon E, Pons JL. Immediate effects of a controllable knee ankle foot orthosis for functional compensation of gait in patients with proximal leg weakness. *Medical & biological engineering & computing*. 2008;46(1):43-53.
- [5] Del-Ama AJ, Koutsou AD, Moreno JC, De-Los-Reyes A, Gil-Agudo Á, Pons JL. Review of hybrid exoskeletons to restore gait following spinal cord injury. *Journal of Rehabilitation Research & Development*. 2012;49(4).
- [6] Fatone S. A review of the literature pertaining to KAFOs and HKAFOs for ambulation. *JPO: Journal of Prosthetics and Orthotics*. 2006;18(7):P137-P68.
- [7] Cochrane H, Orsi K, Reilly P. Lower limb amputation Part 3: Prosthetics-a 10 year literature review. *Prosthetics and orthotics international*. 2001;25(1):21-8.
- [8] Laferrier JZ, Gailey R. Advances in lower-limb prosthetic technology. *Physical Medicine and Rehabilitation Clinics*. 2010;21(1):87-110.
- [9] Herrmann S, Attenberger A, Buchenrieder K, editors. *Prostheses control with combined near-infrared and myoelectric signals*. *International Conference on Computer Aided Systems Theory*; 2011: Springer.